


Data Path : Z:\HPCHEM1\BNA_N\Data\BN030218\
 Data File : BN000226.D
 Acq On : 02 Mar 2018 09:19
 Operator : JU/SJ
 Sample : DFTPP51
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP51

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_N\Methods\SOM-EPA-SIM-BN030218.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Mar 05 10:04:58 2018



AutoFind: Scans 2581, 2582, 2583; Background Corrected with Scan 2574

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	39.0	138714	PASS
68	69	0.00	2	1.7	2877	PASS
69	198	0.00	100	47.0	167061	PASS
70	69	0.00	2	0.5	906	PASS
127	198	10	80	54.3	193088	PASS
197	198	0.00	2	0.7	2533	PASS
198	198	100	100	100.0	355669	PASS
199	198	5	9	6.7	23928	PASS
275	198	10	60	22.2	78962	PASS
365	198	1	100	2.7	9710	PASS
441	443	0.01	100	81.4	35960	PASS
442	198	50	100	63.9	227189	PASS
443	442	15	24	19.5	44192	PASS

m/z	Abundance
37.00	630.0
38.00	1774.0
39.10	9689.0
40.00	873.0
41.00	385.0
44.00	748.0
45.00	366.0
49.00	917.0
50.10	34464.0
51.10	134080.0
52.10	7210.0
53.00	353.0
55.10	786.0
56.00	3849.0
57.00	8831.0
58.00	520.0
61.00	2003.0
62.00	2598.0
63.00	5715.0
64.00	985.0
65.10	2760.0
68.00	2947.0
69.00	158016.0
70.00	775.0
73.10	1721.0
74.10	17040.0
75.00	27184.0
76.10	9094.0
77.10	189696.0
78.10	14615.0
79.00	12381.0
80.00	9706.0
81.00	13163.0
82.00	3474.0
83.00	2837.0
84.10	476.0
85.00	2296.0
86.00	3512.0
87.00	1849.0
88.00	630.0
89.00	380.0
91.00	2596.0
92.00	3832.0
93.00	21768.0
94.00	1584.0
95.10	537.0
96.10	1141.0
97.10	503.0
98.00	16150.0
99.00	17520.0
100.00	1271.0
101.00	6841.0
101.90	379.0
103.00	2282.0
104.00	4246.0
105.00	4148.0
106.00	1440.0
107.00	53456.0
108.00	7799.0
109.10	1558.0
110.00	95112.0
111.00	15476.0
112.00	2296.0
112.90	607.0
116.00	2828.0
117.00	46144.0
118.00	4079.0
119.00	476.0
120.00	686.0
121.00	326.0
122.00	2966.0
123.00	4712.0
124.00	2437.0
125.00	2392.0
127.00	169280.0
128.00	12800.0
129.00	63320.0
130.00	5621.0
131.00	1051.0
132.10	566.0
133.00	359.0
134.00	1908.0
135.00	5014.0
136.00	2363.0
137.00	4052.0
138.00	820.0
140.10	740.0
141.00	7737.0
142.00	2550.0
143.00	1957.0
144.10	512.0

Instrument :
BNA_N
ClientSampleId :
DFTPP51

145.00	527.0
146.00	1317.0
147.00	4087.0
148.00	10776.0
149.00	2851.0
150.00	871.0
151.00	1019.0
152.00	682.0
153.00	2240.0
154.00	2067.0
155.00	4871.0
156.00	6101.0
157.10	1327.0
158.00	1299.0
159.00	1195.0
160.00	2475.0
161.00	3781.0
162.00	891.0
163.10	346.0
164.00	608.0
165.00	2944.0
166.00	2069.0
167.00	16872.0
168.00	16744.0
169.10	2159.0
170.10	686.0
170.90	646.0
172.00	1368.0
173.00	1931.0
174.00	3204.0
175.00	6714.0
176.00	1866.0
177.00	3129.0
178.00	833.0
179.00	10790.0
180.00	9829.0
181.00	3486.0
182.00	594.0
184.00	766.0
185.00	5496.0
186.10	40352.0
187.10	11340.0
188.10	1607.0
189.00	2460.0
190.00	359.0
191.00	1837.0
192.10	4089.0
193.00	4428.0
194.00	912.0
195.10	556.0
196.00	7468.0
197.10	2031.0
198.00	285888.0
199.00	19552.0
200.00	1187.0
201.70	1123.0
203.10	1877.0
204.00	9786.0
205.10	17248.0
206.10	70424.0
207.10	12403.0
208.00	2834.0
209.00	1215.0
211.00	3023.0
212.00	581.0
215.00	829.0
216.10	1737.0
217.00	18360.0
218.00	2561.0
221.00	10040.0
221.80	2332.0
223.10	4271.0
224.10	39128.0
225.10	9963.0
226.00	1602.0
227.00	15942.0
228.00	1909.0
229.00	3655.0
230.00	438.0
231.10	1128.0
234.00	1168.0
235.00	1276.0
236.00	687.0
237.10	2736.0
238.00	439.0
239.00	592.0
240.00	649.0
241.00	972.0
242.00	2067.0
243.10	2187.0
244.10	29320.0
245.10	4403.0
246.00	5640.0
247.00	1218.0
249.00	983.0

Instrument :
BNA_N
ClientSampleId :
DFTPP51

253.00	762.0
255.00	136896.0
256.00	21296.0
257.00	2439.0
258.00	7856.0
259.00	1471.0
265.00	3746.0
265.90	439.0
271.90	389.0
273.00	3917.0
274.00	11061.0
275.10	56296.0
276.10	8436.0
277.00	4767.0
278.00	1409.0
283.10	680.0
284.00	500.0
285.00	780.0
291.90	354.0
292.90	1026.0
294.10	466.0
295.00	381.0
296.00	14527.0
297.00	2326.0
303.10	1936.0
304.00	512.0
314.10	644.0
315.00	1667.0
316.00	1119.0
321.00	632.0
323.10	4333.0
324.00	1073.0
326.90	905.0
328.00	512.0
333.10	568.0
334.10	2850.0
335.00	799.0
341.00	582.0
346.00	887.0
352.10	1512.0
353.00	861.0
354.00	1407.0
355.00	369.0
365.00	6152.0
366.00	834.0
371.00	356.0
372.10	2169.0
373.10	851.0
383.00	543.0
402.00	920.0
403.00	1334.0
404.00	676.0
421.00	1137.0
422.10	1137.0
423.10	7027.0
424.00	4208.0
425.00	914.0
441.10	20224.0
442.10	126112.0
443.10	24264.0
444.10	2858.0

Instrument :
BNA_N
ClientSampleId :
DFTPP51

m/z	Abundance
37.00	669.0
38.10	1687.0
39.10	10928.0
40.00	1070.0
41.10	444.0
44.00	801.0
49.10	1181.0
50.10	43096.0
51.10	165952.0
52.10	8364.0
53.00	357.0
55.00	1420.0
56.00	5456.0
57.00	11414.0
57.90	588.0
61.00	2703.0
62.00	2952.0
63.00	7514.0
64.00	1067.0
65.10	3810.0
66.10	331.0
67.00	378.0
68.00	3233.0
69.00	198784.0
70.10	1243.0
73.00	2318.0
74.00	21928.0
75.00	33448.0
76.10	11598.0
77.10	243776.0
78.10	18352.0
79.00	15410.0
80.00	11085.0
81.00	16163.0
82.00	4411.0
83.00	3679.0
84.00	770.0
85.00	2934.0
86.00	5055.0
87.00	2330.0
88.00	886.0
88.90	371.0
91.00	3435.0
92.00	4702.0
93.00	28328.0
94.00	1833.0
94.90	404.0
95.10	423.0
96.10	1788.0
97.00	560.0
98.00	20312.0
99.00	20888.0
100.00	1605.0
101.00	10032.0
102.00	537.0
103.00	2999.0
104.00	5744.0
105.00	5344.0
106.10	2035.0
107.00	72128.0
108.00	10572.0
109.10	2029.0
110.00	128712.0
111.00	20056.0
112.00	2351.0
113.00	864.0
116.00	3852.0
117.00	61368.0
118.00	4889.0
119.00	747.0
120.00	951.0
121.00	346.0
122.00	4664.0
123.00	7293.0
124.00	2762.0
125.00	2976.0
126.20	1313.0
127.00	236544.0
128.00	17600.0
129.00	89584.0
130.00	7535.0
131.00	1681.0
132.00	882.0
133.00	533.0
134.00	2417.0
135.00	7711.0
136.00	2608.0
137.00	4926.0
137.90	1002.0
138.90	468.0
140.00	1084.0

Instrument :
BNA_N
ClientSampleId :
DFTPP51

141.00	10807.0
142.00	3923.0
143.00	2779.0
144.00	583.0
144.90	790.0
146.00	2170.0
147.00	6222.0
148.00	14441.0
149.00	4034.0
150.10	978.0
151.10	1715.0
152.00	850.0
153.00	3485.0
154.00	2964.0
155.00	6424.0
156.10	9269.0
157.00	1943.0
157.90	1724.0
159.10	1794.0
160.00	3770.0
161.00	5294.0
162.00	1559.0
162.90	392.0
164.10	450.0
165.10	3869.0
166.00	3390.0
167.00	24688.0
168.00	18760.0
169.00	2645.0
170.00	629.0
170.80	893.0
172.10	2045.0
173.00	2884.0
174.00	4674.0
175.10	9444.0
176.10	2719.0
177.00	4378.0
178.00	1607.0
179.00	16392.0
180.00	13354.0
181.00	5208.0
182.00	975.0
182.90	475.0
184.10	1293.0
185.00	7496.0
186.10	59992.0
187.00	17976.0
188.00	2395.0
189.00	4171.0
190.00	796.0
191.00	1795.0
192.00	5143.0
193.00	6091.0
194.10	1362.0
195.00	715.0
196.00	11310.0
197.10	3045.0
198.00	434688.0
199.00	29976.0
200.00	2262.0
201.50	1660.0
203.00	2716.0
204.00	15564.0
205.00	25368.0
206.10	108272.0
207.10	16672.0
208.00	4232.0
209.00	1766.0
210.10	2124.0
211.10	4129.0
215.00	1157.0
216.00	2483.0
217.00	30056.0
218.00	4239.0
219.00	378.0
220.00	516.0
221.10	15852.0
221.90	3672.0
223.00	6811.0
224.10	58208.0
225.10	16464.0
226.00	2032.0
227.00	24736.0
228.00	3620.0
229.00	5299.0
229.90	767.0
231.00	2383.0
232.00	357.0
233.00	385.0
234.00	1692.0
235.10	1415.0
236.10	1441.0
237.10	3261.0
238.00	535.0
239.00	1056.0

Instrument :
BNA_N
ClientSampleId :
DFTPP51

240.00	1042.0
241.10	1533.0
242.10	3617.0
243.10	4181.0
244.10	49000.0
245.10	6451.0
246.00	8522.0
247.00	1901.0
248.00	357.0
249.00	1529.0
250.10	415.0
250.90	328.0
252.00	534.0
253.10	1142.0
255.10	223808.0
256.10	33648.0
257.00	2749.0
258.00	12651.0
259.00	2121.0
260.00	318.0
261.00	502.0
264.20	302.0
265.00	5136.0
266.00	821.0
271.10	501.0
272.00	574.0
273.00	6196.0
274.10	17432.0
275.10	95824.0
276.10	12688.0
277.10	7604.0
278.10	1824.0
279.20	352.0
283.00	909.0
284.00	678.0
285.10	1722.0
286.10	463.0
289.00	348.0
290.00	305.0
292.10	396.0
293.00	2108.0
294.00	680.0
295.10	635.0
296.00	26480.0
297.00	3776.0
301.10	442.0
302.00	476.0
303.00	2761.0
304.10	873.0
307.90	412.0
309.10	347.0
310.00	381.0
314.00	1244.0
315.00	3000.0
316.00	2007.0
317.10	388.0
321.00	700.0
322.00	376.0
323.10	8024.0
324.00	1633.0
327.00	1856.0
328.10	842.0
332.10	635.0
333.00	728.0
334.10	5124.0
335.00	1379.0
341.00	994.0
342.10	370.0
346.00	1836.0
347.10	461.0
352.00	2010.0
353.00	1665.0
354.10	2715.0
355.20	612.0
365.00	11758.0
366.00	1873.0
371.00	841.0
372.10	4245.0
373.10	1033.0
383.00	1189.0
384.00	444.0
389.90	512.0
391.00	344.0
402.00	2022.0
403.10	2747.0
404.00	1092.0
405.10	365.0
421.10	2130.0
422.10	2128.0
423.10	15904.0
424.10	6808.0
425.10	906.0
441.10	42072.0
442.10	273216.0
443.10	53424.0

Instrument :
BNA_N
ClientSampleId :
DFTPP51

444.10

5263.0

Instrument :
BNA_N
ClientSampleId :
DFTPP51

m/z	Abundance
37.00	626.0
38.10	1500.0
39.10	8115.0
40.00	880.0
41.00	414.0
43.90	726.0
44.90	355.0
49.00	1092.0
50.10	29176.0
51.10	116112.0
52.10	5630.0
53.00	310.0
55.00	781.0
56.00	3182.0
57.00	8297.0
58.00	415.0
61.00	1638.0
62.00	1854.0
63.00	5279.0
64.10	777.0
65.10	2738.0
67.00	361.0
68.00	2452.0
69.00	144384.0
70.00	702.0
73.00	1753.0
74.10	14831.0
75.00	24056.0
76.00	8010.0
77.10	175680.0
78.10	12213.0
79.00	11479.0
80.00	8191.0
81.00	12176.0
82.00	2849.0
83.00	2808.0
84.10	422.0
85.00	2610.0
86.00	3679.0
87.00	1633.0
87.90	689.0
89.10	412.0
91.00	2910.0
92.00	3609.0
93.00	20024.0
94.00	1331.0
95.10	562.0
96.00	1014.0
97.10	418.0
98.00	15393.0
99.00	14684.0
100.00	1199.0
101.00	6350.0
101.90	310.0
103.00	2029.0
104.00	4043.0
105.00	4001.0
106.00	1334.0
107.00	51704.0
108.00	8566.0
109.00	1396.0
110.00	93832.0
111.00	14379.0
112.00	2009.0
113.00	522.0
116.00	2562.0
117.00	46512.0
118.00	3809.0
119.00	614.0
120.00	788.0
121.00	321.0
122.00	2907.0
123.00	5034.0
124.00	2404.0
125.00	1931.0
127.00	173440.0
128.00	13556.0
129.00	65160.0
130.00	5626.0
131.00	1184.0
132.10	602.0
133.00	532.0
134.00	2096.0
135.00	5680.0
136.00	2267.0
137.10	3825.0
138.10	545.0
139.00	339.0
140.00	885.0
141.00	8507.0
142.00	3018.0

Instrument :
BNA_N
ClientSampleId :
DFTPP51

143.00	2042.0
144.00	511.0
145.00	543.0
146.00	1665.0
147.00	4810.0
148.00	10980.0
149.00	2442.0
150.00	858.0
151.00	1090.0
151.90	726.0
153.00	2883.0
154.00	2227.0
155.00	4639.0
156.10	7348.0
157.10	1599.0
158.00	1490.0
159.00	1249.0
160.00	2883.0
161.00	3850.0
162.10	1382.0
163.00	327.0
164.00	652.0
165.00	3197.0
166.00	2510.0
167.00	18400.0
168.00	14157.0
169.00	1810.0
170.00	731.0
170.90	599.0
172.00	1646.0
173.00	1897.0
174.00	3250.0
175.00	7122.0
176.00	2297.0
177.00	3070.0
178.00	1060.0
179.00	12643.0
180.00	9522.0
181.10	4056.0
182.00	863.0
183.10	340.0
184.00	1038.0
185.00	6498.0
186.10	48344.0
187.00	14435.0
188.00	1608.0
189.00	3529.0
190.00	613.0
191.00	1782.0
192.00	4486.0
193.00	5095.0
194.00	996.0
194.90	445.0
196.00	9826.0
197.00	2525.0
198.00	346432.0
199.00	22256.0
200.00	1856.0
200.80	390.0
201.60	1199.0
203.00	2409.0
204.00	12935.0
205.10	21952.0
206.10	88544.0
207.10	13652.0
208.10	3175.0
209.00	1567.0
210.20	1283.0
211.10	3609.0
211.90	709.0
213.10	304.0
215.00	1095.0
216.00	2074.0
217.00	23288.0
218.00	3546.0
221.00	13356.0
223.00	5066.0
224.10	49008.0
225.10	11779.0
226.10	1872.0
227.00	19672.0
228.00	2750.0
229.00	3816.0
230.00	511.0
231.10	1946.0
233.00	614.0
234.10	1438.0
235.00	1188.0
235.90	716.0
237.00	2657.0
238.00	306.0
239.00	921.0
240.00	701.0
241.00	1071.0
242.00	2817.0

Instrument :
BNA_N
ClientSampleId :
DFTPP51

243.00	2926.0
244.10	40760.0
245.00	5481.0
246.00	7364.0
247.00	1484.0
248.00	352.0
249.00	1293.0
251.00	376.0
251.90	489.0
253.00	1022.0
254.10	1537.0
255.00	190016.0
256.10	26896.0
257.00	2543.0
258.00	11305.0
259.00	1893.0
260.90	330.0
263.90	352.0
265.00	4107.0
266.00	552.0
271.00	474.0
271.90	504.0
273.10	5760.0
274.10	14877.0
275.10	84768.0
276.10	10984.0
277.00	6125.0
278.00	1542.0
279.10	308.0
281.00	322.0
283.00	725.0
284.00	496.0
285.10	1173.0
286.00	329.0
293.00	1799.0
293.90	489.0
295.10	633.0
296.00	24416.0
297.00	3298.0
301.00	302.0
302.10	471.0
303.10	2576.0
304.10	829.0
308.10	490.0
309.00	323.0
314.00	995.0
315.10	2749.0
316.10	1520.0
321.00	810.0
322.00	376.0
323.10	7792.0
324.00	1331.0
327.00	1732.0
328.00	839.0
332.00	598.0
333.00	717.0
334.10	4546.0
335.10	1344.0
341.10	932.0
346.00	1875.0
347.10	500.0
352.00	2157.0
353.10	1734.0
354.10	2288.0
355.10	373.0
365.00	11221.0
366.10	1701.0
369.90	316.0
371.00	605.0
372.10	4293.0
373.00	890.0
383.00	1104.0
384.10	352.0
390.00	598.0
391.00	371.0
392.10	467.0
402.10	1534.0
403.10	2453.0
404.00	1080.0
421.00	2301.0
422.00	2531.0
423.10	15562.0
424.10	5593.0
425.10	882.0
441.10	45584.0
442.10	282240.0
443.10	54888.0
444.10	4683.0
445.00	324.0

Instrument :
BNA_N
ClientSampleId :
DFTPP51

Scan 2574 (18.128 min): BN000226.D
DFTPP51

m/z	Abundance
40.00	672.0
44.00	817.0
73.00	543.0
207.00	1987.0
208.10	570.0

Instrument :
BNA_N
ClientSampleId :
DFTPP51