

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030224\
 Data File : BN030186.D
 Acq On : 02 Mar 2024 18:07
 Operator : MA/JU
 Sample : P1501-22
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 E10C1

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 03/04/2024
 Supervised By :mohammad ahmed 03/05/2024

Quant Time: Mar 03 22:48:24 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN030224.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Mar 02 16:09:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.834	152	1104	0.400	ng/ul	0.00
4) Naphthalene-d8	10.628	136	2912	0.400	ng/ul	-0.01
9) Acenaphthene-d10	14.479	164	1546	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.234	188	3145	0.400	ng/ul	-0.01
17) Chrysene-d12	21.456	240	2464	0.400	ng/ul	0.00
23) Perylene-d12	23.794	264	2324m	0.400	ng/ul	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.307	96	7630	4.992	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.229	152	1783	0.474	ng/ul	-0.01
18) Fluoranthene-d10	19.272	212	4078	0.508	ng/ul	-0.01
Target Compounds						
5) Naphthalene	10.678	128	263	0.031	ng/ul#	Qvalue 54

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030224\
Data File : BN030186.D
Acq On : 02 Mar 2024 18:07
Operator : MA/JU
Sample : P1501-22
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
E10C1

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 03/04/2024
Supervised By :mohammad ahmed 03/05/2024

Quant Time: Mar 03 22:48:24 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN030224.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Mar 02 16:09:30 2024
Response via : Initial Calibration

