

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030224\
 Data File : BN030205.D
 Acq On : 03 Mar 2024 06:09
 Operator : MA/JU
 Sample : PB159249BS
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS249

Quant Time: Mar 03 22:50:53 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN030224.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Mar 02 16:09:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.838	152	1596	0.400	ng/ul	0.00
4) Naphthalene-d8	10.634	136	4261	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.483	164	1939	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.247	188	3337	0.400	ng/ul	0.00
17) Chrysene-d12	21.450	240	2538	0.400	ng/ul	0.00
23) Perylene-d12	23.806	264	2679	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.307	96	1492	0.675	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.240	152	1824	0.331	ng/ul	0.00
18) Fluoranthene-d10	19.281	212	3042	0.368	ng/ul	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.340	88	4254	1.774	ng/ul#	86
5) Naphthalene	10.689	128	4079	0.324	ng/ul	99
7) 2-Methylnaphthalene	12.311	142	2117	0.311	ng/ul	100
8) 1-Methylnaphthalene	12.526	142	2286	0.316	ng/ul	98
10) Acenaphthylene	14.206	152	3463	0.319	ng/ul	96
11) Acenaphthene	14.548	153	2559	0.322	ng/ul	96
12) Fluorene	15.547	166	2580	0.318	ng/ul	95
14) Pentachlorophenol	16.897	266	725	0.761	ng/ul	99
15) Phenanthrene	17.289	178	3603	0.322	ng/ul	98
16) Anthracene	17.387	178	3183	0.323	ng/ul	98
19) Fluoranthene	19.314	202	3906	0.346	ng/ul	99
20) Pyrene	19.672	202	4245	0.346	ng/ul	99
21) Benzo(a)anthracene	21.438	228	2974	0.332	ng/ul	98
22) Chrysene	21.488	228	3579	0.332	ng/ul	97
24) Benzo(b)fluoranthene	23.090	252	3531	0.342	ng/ul	95
25) Benzo(k)fluoranthene	23.139	252	4033	0.331	ng/ul	96
26) Benzo(a)pyrene	23.704	252	3514	0.334	ng/ul	94
27) Indeno(1,2,3-cd)pyrene	26.246	276	4553	0.339	ng/ul#	97
28) Dibenzo(a,h)anthracene	26.282	278	3546	0.341	ng/ul	97
29) Benzo(g,h,i)perylene	26.980	276	4157	0.334	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030224\
Data File : BN030205.D
Acq On : 03 Mar 2024 06:09
Operator : MA/JU
Sample : PB159249BS
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SLCS249

Quant Time: Mar 03 22:50:53 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN030224.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Mar 02 16:09:30 2024
Response via : Initial Calibration

