

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030322\
 Data File : BN018793.D
 Acq On : 02 Mar 2022 16:03
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Quant Time: Mar 03 00:11:46 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN030322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 03 00:10:39 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.912	152	5931	0.400	ng	0.00
7) Naphthalene-d8	10.722	136	15717	0.400	ng	0.00
13) Acenaphthene-d10	14.538	164	9547	0.400	ng	0.00
19) Phenanthrene-d10	17.287	188	20822	0.400	ng	# 0.01
29) Chrysene-d12	21.458	240	17100	0.400	ng	# 0.00
35) Perylene-d12	23.861	264	12873	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.464	112	5302	0.391	ng	0.00
5) Phenol-d6	7.067	99	6183	0.396	ng	0.00
8) Nitrobenzene-d5	9.068	82	4277	0.352	ng	0.00
11) 2-Methylnaphthalene-d10	12.306	152	9969	0.421	ng	0.00
14) 2,4,6-Tribromophenol	16.026	330	1553	0.417	ng	0.00
15) 2-Fluorobiphenyl	13.170	172	16175	0.357	ng	0.00
27) Fluoranthene-d10	19.312	212	23618	0.418	ng	0.00
31) Terphenyl-d14	19.906	244	14760	0.398	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.319	88	2738	0.398	ng	98
3) n-Nitrosodimethylamine	3.637	42	2833	0.415	ng	# 98
6) bis(2-Chloroethyl)ether	7.327	93	6721	0.419	ng	99
9) Naphthalene	10.765	128	17815	0.393	ng	99
10) Hexachlorobutadiene	11.064	225	4086	0.375	ng	# 100
12) 2-Methylnaphthalene	12.378	142	11890	0.392	ng	99
16) Acenaphthylene	14.260	152	16437	0.363	ng	99
17) Acenaphthene	14.602	154	12076	0.386	ng	100
18) Fluorene	15.586	166	15557	0.413	ng	100
20) 4,6-Dinitro-2-methylph...	15.661	198	968	0.317	ng	99
21) 4-Bromophenyl-phenylether	16.477	248	5355	0.362	ng	99
22) Hexachlorobenzene	16.600	284	6796	0.367	ng	99
23) Atrazine	16.733	200	2903	0.355	ng	98
24) Pentachlorophenol	16.938	266	1476	0.377	ng	99
25) Phenanthrene	17.318	178	26977	0.387	ng	100
26) Anthracene	17.410	178	21843	0.370	ng	99
28) Fluoranthene	19.341	202	30284	0.416	ng	100
30) Pyrene	19.704	202	30269	0.396	ng	100
32) Benzo(a)anthracene	21.449	228	24648	0.380	ng	100
33) Chrysene	21.494	228	28165	0.389	ng	97
34) Bis(2-ethylhexyl)phtha...	21.378	149	9927	0.377	ng	99
36) Indeno(1,2,3-cd)pyrene	26.349	276	21393	0.315	ng	99
37) Benzo(b)fluoranthene	23.127	252	23015	0.394	ng	98
38) Benzo(k)fluoranthene	23.177	252	22528	0.389	ng	100
39) Benzo(a)pyrene	23.755	252	17152	0.368	ng	# 96
40) Dibenzo(a,h)anthracene	26.366	278	16389	0.309	ng	100
41) Benzo(g,h,i)perylene	27.109	276	18860	0.324	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030322\
Data File : BN018793.D
Acq On : 02 Mar 2022 16:03
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Quant Time: Mar 03 00:11:46 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN030322.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 03 00:10:39 2022
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.912 min Scan# 669
 Delta R.T. 0.000 min
 Lab File: BN018793.D
 Acq: 02 Mar 2022 16:03




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4


Tgt Ion:152 Resp: 5931
 Ion Ratio Lower Upper
 152 100
 150 155.0 123.9 185.9
 115 60.6 48.2 72.4



#2
 1,4-Dioxane
 Concen: 0.398 ng
 RT: 3.319 min Scan# 33
 Delta R.T. 0.000 min
 Lab File: BN018793.D
 Acq: 02 Mar 2022 16:03

Tgt Ion: 88 Resp: 2738
 Ion Ratio Lower Upper
 88 100
 43 31.4 24.6 37.0
 58 80.3 65.4 98.2









#3
n-Nitrosodimethylamine
Concen: 0.415 ng
RT: 3.637 min Scan# 71
Delta R.T. -0.007 min
Lab File: BN018793.D
Acq: 02 Mar 2022 16:03




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion: 42 Resp: 2833
Ion Ratio Lower Upper
42 100
74 119.0 96.9 145.3
44 6.7 7.3 10.9#



#4
2-Fluorophenol
Concen: 0.391 ng
RT: 5.464 min Scan# 330
Delta R.T. 0.000 min
Lab File: BN018793.D
Acq: 02 Mar 2022 16:03

Tgt Ion: 112 Resp: 5302
Ion Ratio Lower Upper
112 100
64 60.4 47.8 71.8
63 31.1 25.7 38.5









#5
Phenol-d6
Concen: 0.396 ng
RT: 7.067 min Scan# 511
Delta R.T. 0.000 min
Lab File: BN018793.D
Acq: 02 Mar 2022 16:03




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion: 99 Resp: 6183
Ion Ratio Lower Upper
99 100
42 20.3 16.2 24.4
71 34.0 27.7 41.5



#6
bis(2-Chloroethyl)ether
Concen: 0.419 ng
RT: 7.327 min Scan# 588
Delta R.T. 0.000 min
Lab File: BN018793.D
Acq: 02 Mar 2022 16:03

Tgt Ion: 93 Resp: 6721
Ion Ratio Lower Upper
93 100
63 79.2 62.8 94.2
95 32.9 26.2 39.4









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.722 min Scan# 950
Delta R.T. 0.000 min
Lab File: BN018793.D
Acq: 02 Mar 2022 16:03




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	136	Resp:	15717
Ion Ratio	Lower	Upper	
136	100		
137	11.4	9.0	13.4
54	7.4	5.8	8.6
68	5.8	4.6	7.0



#8
Nitrobenzene-d5
Concen: 0.352 ng
RT: 9.068 min Scan# 795
Delta R.T. 0.000 min
Lab File: BN018793.D
Acq: 02 Mar 2022 16:03

Tgt Ion:	82	Resp:	4277
Ion Ratio	Lower	Upper	
82	100		
128	40.1	33.0	49.6
54	51.1	42.5	63.7









#9
Naphthalene
Concen: 0.393 ng
RT: 10.765 min Scan# 954
Delta R.T. 0.000 min
Lab File: BN018793.D
Acq: 02 Mar 2022 16:03




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	128	Resp:	17815
Ion	Ratio	Lower	Upper
128	100		
129	11.3	9.0	13.4
127	13.8	10.7	16.1



#10
Hexachlorobutadiene
Concen: 0.375 ng
RT: 11.064 min Scan# 982
Delta R.T. 0.000 min
Lab File: BN018793.D
Acq: 02 Mar 2022 16:03

Tgt Ion:	225	Resp:	4086
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	63.4	51.0	76.6









#11
2-Methylnaphthalene-d10
Concen: 0.421 ng
RT: 12.306 min Scan# 11
Delta R.T. 0.000 min
Lab File: BN018793.D
Acq: 02 Mar 2022 16:03




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:152 Resp: 9969
Ion Ratio Lower Upper
152 100
151 20.8 16.5 24.7



#12
2-Methylnaphthalene
Concen: 0.392 ng
RT: 12.378 min Scan# 1253
Delta R.T. 0.000 min
Lab File: BN018793.D
Acq: 02 Mar 2022 16:03

Tgt Ion:142 Resp: 11890
Ion Ratio Lower Upper
142 100
141 88.1 70.0 105.0
115 37.5 29.8 44.8









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.538 min Scan# 1477
Delta R.T. 0.000 min
Lab File: BN018793.D
Acq: 02 Mar 2022 16:03




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:164 Resp: 9547
Ion Ratio Lower Upper
164 100
162 109.0 85.2 127.8
160 51.8 41.5 62.3



#14
2,4,6-Tribromophenol
Concen: 0.417 ng
RT: 16.026 min Scan# 1617
Delta R.T. 0.000 min
Lab File: BN018793.D
Acq: 02 Mar 2022 16:03

Tgt Ion:330 Resp: 1553
Ion Ratio Lower Upper
330 100
332 95.8 76.8 115.2
141 39.3 32.2 48.2









#15
2-Fluorobiphenyl
Concen: 0.357 ng
RT: 13.170 min Scan# 1349
Delta R.T. 0.000 min
Lab File: BN018793.D
Acq: 02 Mar 2022 16:03




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:172 Resp: 16175
Ion Ratio Lower Upper
172 100
171 35.4 28.2 42.2
170 23.7 18.7 28.1



#16
Acenaphthylene
Concen: 0.363 ng
RT: 14.260 min Scan# 1451
Delta R.T. 0.000 min
Lab File: BN018793.D
Acq: 02 Mar 2022 16:03

Tgt Ion:152 Resp: 16437
Ion Ratio Lower Upper
152 100
151 19.3 15.9 23.9
153 13.2 10.6 15.8









#17
Acenaphthene
Concen: 0.386 ng
RT: 14.602 min Scan# 1483
Delta R.T. 0.000 min
Lab File: BN018793.D
Acq: 02 Mar 2022 16:03




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:154 Resp: 12076
Ion Ratio Lower Upper
154 100
153 111.8 89.7 134.5
152 55.5 44.7 67.1



#18
Fluorene
Concen: 0.413 ng
RT: 15.586 min Scan# 1575
Delta R.T. 0.000 min
Lab File: BN018793.D
Acq: 02 Mar 2022 16:03

Tgt Ion:166 Resp: 15557
Ion Ratio Lower Upper
166 100
165 98.6 79.2 118.8
167 13.6 10.7 16.1









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.287 min Scan# 11
Delta R.T. 0.010 min
Lab File: BN018793.D
Acq: 02 Mar 2022 16:03




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:188 Resp: 20822
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 9.2 11.8 17.8#



#20
4,6-Dinitro-2-methylphenol
Concen: 0.317 ng
RT: 15.661 min Scan# 1582
Delta R.T. 0.000 min
Lab File: BN018793.D
Acq: 02 Mar 2022 16:03

Tgt Ion:198 Resp: 968
Ion Ratio Lower Upper
198 100
182 16.8 13.0 19.4
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 0.362 ng
RT: 16.477 min Scan# 1661
Delta R.T. 0.000 min
Lab File: BN018793.D
Acq: 02 Mar 2022 16:03




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:248 Resp: 5355
Ion Ratio Lower Upper
248 100
250 100.2 80.7 121.1
141 51.2 39.6 59.4



#22
Hexachlorobenzene
Concen: 0.367 ng
RT: 16.600 min Scan# 1673
Delta R.T. 0.000 min
Lab File: BN018793.D
Acq: 02 Mar 2022 16:03

Tgt Ion:284 Resp: 6796
Ion Ratio Lower Upper
284 100
142 38.4 29.8 44.8
249 33.1 26.2 39.2









#23
Atrazine
Concen: 0.355 ng
RT: 16.733 min Scan# 1686
Delta R.T. 0.000 min
Lab File: BN018793.D
Acq: 02 Mar 2022 16:03




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:200 Resp: 2903
Ion Ratio Lower Upper
200 100
173 29.4 23.9 35.9
215 47.3 36.6 54.8



#24
Pentachlorophenol
Concen: 0.377 ng
RT: 16.938 min Scan# 1706
Delta R.T. 0.000 min
Lab File: BN018793.D
Acq: 02 Mar 2022 16:03

Tgt Ion:266 Resp: 1476
Ion Ratio Lower Upper
266 100
264 61.7 50.2 75.4
268 65.5 51.9 77.9









#25
Phenanthrene
Concen: 0.387 ng
RT: 17.318 min Scan# 1743
Delta R.T. -0.010 min
Lab File: BN018793.D
Acq: 02 Mar 2022 16:03




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:178 Resp: 26977
Ion Ratio Lower Upper
178 100
176 19.3 15.4 23.0
179 15.2 12.2 18.2



#26
Anthracene
Concen: 0.370 ng
RT: 17.410 min Scan# 1752
Delta R.T. 0.000 min
Lab File: BN018793.D
Acq: 02 Mar 2022 16:03

Tgt Ion:178 Resp: 21843
Ion Ratio Lower Upper
178 100
176 18.8 14.8 22.2
179 15.0 12.2 18.4









#27
Fluoranthene-d10
Concen: 0.418 ng
RT: 19.312 min Scan# 2120
Delta R.T. 0.000 min
Lab File: BN018793.D
Acq: 02 Mar 2022 16:03




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:212 Resp: 23618
Ion Ratio Lower Upper
212 100
106 16.0 12.7 19.1
104 9.0 7.1 10.7



#28
Fluoranthene
Concen: 0.416 ng
RT: 19.341 min Scan# 2127
Delta R.T. 0.000 min
Lab File: BN018793.D
Acq: 02 Mar 2022 16:03

Tgt Ion:202 Resp: 30284
Ion Ratio Lower Upper
202 100
101 12.2 9.8 14.6
203 16.9 13.7 20.5









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.458 min Scan# 2447
Delta R.T. -0.009 min
Lab File: BN018793.D
Acq: 02 Mar 2022 16:03




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:240 Resp: 17100
Ion Ratio Lower Upper
240 100
120 18.9 8.5 12.7#
236 29.2 22.4 33.6



#30
Pyrene
Concen: 0.396 ng
RT: 19.704 min Scan# 2213
Delta R.T. 0.000 min
Lab File: BN018793.D
Acq: 02 Mar 2022 16:03

Tgt Ion:202 Resp: 30269
Ion Ratio Lower Upper
202 100
200 20.9 16.9 25.3
203 17.6 14.2 21.2









#31
 Terphenyl-d14
 Concen: 0.398 ng
 RT: 19.906 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: BN018793.D
 Acq: 02 Mar 2022 16:03




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4


Tgt Ion:244 Resp: 14760
 Ion Ratio Lower Upper
 244 100
 212 10.2 8.2 12.4
 122 0.0 0.0 0.0



#32
 Benzo(a)anthracene
 Concen: 0.380 ng
 RT: 21.449 min Scan# 2446
 Delta R.T. 0.000 min
 Lab File: BN018793.D
 Acq: 02 Mar 2022 16:03

Tgt Ion:228 Resp: 24648
 Ion Ratio Lower Upper
 228 100
 226 27.2 21.9 32.9
 229 20.0 16.0 24.0





#33

Chrysene

Concen: 0.389 ng

RT: 21.494 min Scan# 2452

Delta R.T. -0.009 min

Lab File: BN018793.D

Acq: 02 Mar 2022 16:03

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4





Tgt Ion:228 Resp: 28165

Ion Ratio Lower Upper

228 100

226 31.6 24.0 36.0

229 19.0 15.9 23.9



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.377 ng

RT: 21.378 min Scan# 2438

Delta R.T. 0.000 min

Lab File: BN018793.D

Acq: 02 Mar 2022 16:03




Tgt Ion:149 Resp: 9927


Ion Ratio Lower Upper

149 100

167 27.0 21.8 32.8

279 2.5 2.0 3.0





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.861 min Scan# 3051

Delta R.T. -0.003 min

Lab File: BN018793.D

Acq: 02 Mar 2022 16:03

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4





Tgt Ion:264 Resp: 12873

Ion Ratio Lower Upper

264 100

260 27.7 22.4 33.6

265 31.9 25.3 37.9



#36

Indeno(1,2,3-cd)pyrene

Concen: 0.315 ng

RT: 26.349 min Scan# 3901

Delta R.T. 0.000 min

Lab File: BN018793.D

Acq: 02 Mar 2022 16:03




Tgt Ion:276 Resp: 21393


Ion Ratio Lower Upper

276 100

138 30.0 24.5 36.7

277 24.9 19.8 29.8









#37
Benzo(b)fluoranthene
Concen: 0.394 ng
RT: 23.127 min Scan# 21
Delta R.T. -0.003 min
Lab File: BN018793.D
Acq: 02 Mar 2022 16:03




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:252 Resp: 23015
Ion Ratio Lower Upper
252 100
253 23.7 18.9 28.3
125 12.6 12.1 18.1



#38
Benzo(k)fluoranthene
Concen: 0.389 ng
RT: 23.177 min Scan# 2816
Delta R.T. 0.000 min
Lab File: BN018793.D
Acq: 02 Mar 2022 16:03

Tgt Ion:252 Resp: 22528
Ion Ratio Lower Upper
252 100
253 23.9 18.9 28.3
125 12.9 10.3 15.5









#39
Benzo(a)pyrene
Concen: 0.368 ng
RT: 23.755 min Scan# 3014
Delta R.T. 0.000 min
Lab File: BN018793.D
Acq: 02 Mar 2022 16:03




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:252 Resp: 17152
Ion Ratio Lower Upper
252 100
253 25.3 20.1 30.1
125 15.2 15.2 22.8#



#40
Dibenzo(a,h)anthracene
Concen: 0.309 ng
RT: 26.366 min Scan# 3907
Delta R.T. 0.000 min
Lab File: BN018793.D
Acq: 02 Mar 2022 16:03

Tgt Ion:278 Resp: 16389
Ion Ratio Lower Upper
278 100
139 21.5 17.2 25.8
279 25.0 19.9 29.9





#41

Benzo(g,h,i)perylene

Concen: 0.324 ng

RT: 27.109 min Scan# 41

Delta R.T. -0.003 min

Lab File: BN018793.D


Acq: 02 Mar 2022 16:03

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4



Tgt Ion:276 Resp: 18860

Ion Ratio Lower Upper

276 100

277 24.4 19.4 29.0

138 26.9 21.0 31.6

