

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030322\
 Data File : BN018809.D
 Acq On : 03 Mar 2022 02:53
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 03 04:35:15 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN030322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 03 00:44:34 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	110	0.00
2	1,4-Dioxane	0.400	0.411	-2.7	112	0.00
3	n-Nitrosodimethylamine	0.400	0.395	1.3	114	0.00
4 S	2-Fluorophenol	0.400	0.387	3.3	111	0.00
5 S	Phenol-d6	0.400	0.375	6.3	112	0.00
6	bis(2-Chloroethyl)ether	0.400	0.393	1.8	110	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	111	-0.01
8 S	Nitrobenzene-d5	0.400	0.356	11.0	111	0.00
9	Naphthalene	0.400	0.385	3.8	111	0.00
10	Hexachlorobutadiene	0.400	0.386	3.5	108	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.376	6.0	109	0.00
12	2-Methylnaphthalene	0.400	0.366	8.5	109	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	106	0.00
14 S	2,4,6-Tribromophenol	0.400	0.324	19.0	103	0.00
15 S	2-Fluorobiphenyl	0.400	0.393	1.8	107	0.00
16	Acenaphthylene	0.400	0.358	10.5	105	0.00
17	Acenaphthene	0.400	0.386	3.5	108	0.00
18	Fluorene	0.400	0.380	5.0	106	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	104	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.337	15.8	95	0.00
21	4-Bromophenyl-phenylether	0.400	0.381	4.8	105	-0.01
22	Hexachlorobenzene	0.400	0.393	1.8	106	0.00
23	Atrazine	0.400	0.333	16.8	101	0.00
24	Pentachlorophenol	0.400	0.407	-1.7	96	0.00
25	Phenanthrene	0.400	0.384	4.0	104	-0.01
26	Anthracene	0.400	0.362	9.5	104	0.00
27 SURR	Fluoranthene-d10	0.400	0.356	11.0	100	0.00
28	Fluoranthene	0.400	0.359	10.3	100	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	91	0.00
30	Pyrene	0.400	0.424	-6.0	100	0.00
31 S	Terphenyl-d14	0.400	0.422	-5.5	101	0.00
32	Benzo(a)anthracene	0.400	0.357	10.8	87	0.00
33	Chrysene	0.400	0.398	0.5	94	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.235	41.3#	20	0.00
35 I	Perylene-d12	0.400	0.400	0.0	90	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.376	6.0	92	0.00
37	Benzo(b)fluoranthene	0.400	0.379	5.3	90	0.00
38	Benzo(k)fluoranthene	0.400	0.375	6.3	88	0.00
39 C	Benzo(a)pyrene	0.400	0.373	6.8	89	0.00
40	Dibenzo(a,h)anthracene	0.400	0.373	6.8	91	0.00
41	Benzo(g,h,i)perylene	0.400	0.381	4.8	92	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4