

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030422\
 Data File : BN018839.D
 Acq On : 03 Mar 2022 23:38
 Operator : CG/JU
 Sample : PB143055BS
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS055

Quant Time: Mar 04 01:48:12 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN021822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 18 13:39:55 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.925	152	84	0.400	ng/ul	# 0.00
4) Naphthalene-d8	10.721	136	244	0.400	ng/ul	# 0.00
9) Acenaphthene-d10	14.547	164	89	0.400	ng/ul	# 0.00
13) Phenanthrene-d10	17.296	188	106	0.400	ng/ul	# 0.00
17) Chrysene-d12	21.473	240	117	0.400	ng/ul	# 0.00
23) Perylene-d12	23.867	264	29	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.297	96	44	0.461	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.310	152	118	0.354	ng/ul	0.00
18) Fluoranthene-d10	19.312	212	129	0.369	ng/ul	0.00
Target Compounds						
					Qvalue	
5) Naphthalene	10.776	128	291	0.421	ng/ul#	35
7) 2-Methylnaphthalene	12.381	142	132	0.296	ng/ul	98
8) 1-Methylnaphthalene	12.601	142	160	0.360	ng/ul	96
10) Acenaphthylene	14.269	152	147	0.366	ng/ul#	15
11) Acenaphthene	14.607	153	121	0.376	ng/ul#	85
12) Fluorene	15.597	166	85	0.246	ng/ul#	56
14) Pentachlorophenol	16.954	266	8	0.370	ng/ul#	42
15) Phenanthrene	17.338	178	160	0.459	ng/ul#	1
16) Anthracene	17.427	178	93	0.299	ng/ul#	1
19) Fluoranthene	19.345	202	179	0.364	ng/ul#	1
20) Pyrene	19.703	202	190	0.375	ng/ul#	9
21) Benzo(a)anthracene	21.453	228	121	0.267	ng/ul#	31
22) Chrysene	21.508	228	184	0.390	ng/ul#	49
24) Benzo(b)fluoranthene	23.133	252	213	1.651	ng/ul#	1
25) Benzo(k)fluoranthene	23.183	252	106	0.823	ng/ul#	1
26) Benzo(a)pyrene	23.750	252	128	1.125	ng/ul#	1
27) Indeno(1,2,3-cd)pyrene	26.360	276	146	1.008	ng/ul#	56
28) Dibenzo(a,h)anthracene	26.373	278	44	0.387	ng/ul#	1
29) Benzo(g,h,i)perylene	27.104	276	145	1.168	ng/ul#	1

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030422\
Data File : BN018839.D
Acq On : 03 Mar 2022 23:38
Operator : CG/JU
Sample : PB143055BS
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SLCS055

Quant Time: Mar 04 01:48:12 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN021822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 18 13:39:55 2022
Response via : Initial Calibration

