

Data Path : Z:\HPCHEM1\BNA_N\Data\BN030518\
 Data File : BN000235.D
 Acq On : 05 Mar 2018 09:23
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTD0.459

Quant Time: Mar 05 14:06:45 2018
 Quant Method : Z:\HPCHEM1\BNA_N\Methods\SOM-EPA-SIM-BN030218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 05 10:04:58 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	95	0.00
2 I	Naphthalene-d8	0.400	0.400	0.0	94	0.00
3	Naphthalene	0.400	0.373	6.8	95	0.00
4 SURR	2-Methylnaphthalene-d10	0.400	0.370	7.5	94	0.00
5	2-Methylnaphthalene	0.400	0.369	7.8	94	0.00
6 I	Acenaphthene-d10	0.400	0.400	0.0	93	0.00
7	Acenaphthylene	0.400	0.368	8.0	93	0.00
8 C	Acenaphthene	0.400	0.367	8.3	93	0.00
9	Fluorene	0.400	0.360	10.0	91	0.00
10 I	Phenanthrene-d10	0.400	0.400	0.0	90	0.00
11	Pentachlorophenol	0.400	0.309	22.8#	89	0.00
12	Phenanthrene	0.400	0.370	7.5	90	0.00
13	Anthracene	0.400	0.355	11.3	87	0.00
14 SURR	Fluoranthene-d10	0.400	0.353	11.8	88	0.00
15 C	Fluoranthene	0.400	0.354	11.5	88	0.00
16 I	Chrysene-d12	0.400	0.400	0.0	80	0.00
17	Pyrene	0.400	0.412	-3.0	89	0.00
18	Benzo(a)anthracene	0.400	0.366	8.5	81	0.00
19	Chrysene	0.400	0.372	7.0	82	0.00
20 I	Perylene-d12	0.400	0.400	0.0	79	0.00
21	Benzo(b)fluoranthene	0.400	0.373	6.8	82	0.00
22	Benzo(k)fluoranthene	0.400	0.369	7.8	80	0.00
23 C	Benzo(a)pyrene	0.400	0.361	9.8	78	0.00
24	Indeno(1,2,3-cd)pyrene	0.400	0.337	15.8	74	0.00
25	Dibenzo(a,h)anthracene	0.400	0.316	21.0#	70	-0.01
26	Benzo(g,h,i)perylene	0.400	0.363	9.3	80	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0