

Data Path : Z:\HPCHEM1\BNA_N\Data\BN030518\
 Data File : BN000253.D
 Acq On : 05 Mar 2018 19:44
 Operator : JU/SJ
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTD0.461

Quant Time: Mar 06 04:39:17 2018
 Quant Method : Z:\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN030218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 06 04:37:49 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	86	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	85	0.00
3	Naphthalene	1.281	1.184	7.6	85	0.00
4 SURR	2-Methylnaphthalene-d10	0.526	0.477	9.3	83	0.00
5	2-Methylnaphthalene	0.812	0.735	9.5	83	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	83	0.00
7	Acenaphthylene	1.990	1.783	10.4	81	0.00
8 C	Acenaphthene	1.921	1.773	7.7	83	0.00
9	Fluorene	2.086	1.931	7.4	84	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	83	0.00
11	Pentachlorophenol	0.080	0.064	20.0	85	0.00
12	Phenanthrene	1.632	1.492	8.6	82	0.00
13	Anthracene	1.213	1.076	11.3	81	0.00
14 SURR	Fluoranthene-d10	0.927	0.826	10.9	82	0.00
15 C	Fluoranthene	1.584	1.410	11.0	82	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	79	0.00
17	Pyrene	2.354	2.273	3.4	82	0.00
18	Benzo(a)anthracene	1.570	1.424	9.3	80	0.00
19	Chrysene	2.137	1.996	6.6	81	0.00
20 I	Perylene-d12	1.000	1.000	0.0	81	0.00
21	Benzo(b)fluoranthene	2.407	2.255	6.3	85	0.00
22	Benzo(k)fluoranthene	2.268	2.068	8.8	81	0.00
23 C	Benzo(a)pyrene	1.883	1.755	6.8	83	0.00
24	Indeno(1,2,3-cd)pyrene	2.126	1.942	8.7	83	0.00
25	Dibenzo(a,h)anthracene	1.534	1.396	9.0	83	0.00
26	Benzo(g,h,i)perylene	2.094	1.887	9.9	82	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0