


Data Path : Z:\HPCHEM1\BNA_N\Data\BN030518\
 Data File : BN000234.D
 Acq On : 05 Mar 2018 08:11
 Operator : JU/SJ
 Sample : DFTPP52
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP52

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_N\Methods\SOM-EPA-SIM-BN030218.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Mar 05 14:32:42 2018



AutoFind: Scans 2581, 2582, 2583; Background Corrected with Scan 2574

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	37.8	134176	PASS
68	69	0.00	2	1.8	2883	PASS
69	198	0.00	100	45.8	162624	PASS
70	69	0.00	2	0.6	905	PASS
127	198	10	80	53.9	191466	PASS
197	198	0.00	2	0.7	2333	PASS
198	198	100	100	100.0	354922	PASS
199	198	5	9	6.7	23898	PASS
275	198	10	60	22.5	80000	PASS
365	198	1	100	2.7	9582	PASS
441	443	0.01	100	82.6	35552	PASS
442	198	50	100	64.4	228522	PASS
443	442	15	24	18.8	43034	PASS

m/z	Abundance
37.00	651.0
38.00	1556.0
39.10	9758.0
40.00	1140.0
41.00	471.0
43.00	316.0
44.00	715.0
49.00	996.0
50.10	35528.0
51.10	137344.0
52.10	7168.0
53.10	521.0
55.00	843.0
56.00	4099.0
57.00	9522.0
58.00	446.0
61.00	1774.0
62.00	2285.0
63.10	5511.0
64.00	650.0
65.10	3303.0
68.10	2772.0
69.00	160896.0
70.00	896.0
73.10	1864.0
74.10	17672.0
75.00	26504.0
76.10	9352.0
77.10	191296.0
78.10	14338.0
79.00	12274.0
80.00	9383.0
81.00	13440.0
82.00	3062.0
83.00	2768.0
84.10	678.0
85.00	2115.0
86.00	3829.0
87.00	1935.0
88.00	692.0
89.00	535.0
91.00	3080.0
92.00	3939.0
93.00	21776.0
94.00	1341.0
95.00	521.0
96.00	1135.0
97.20	449.0
98.00	16174.0
99.00	16720.0
100.00	1363.0
101.00	7342.0
101.90	509.0
103.00	2581.0
104.00	4205.0
105.00	4326.0
106.00	1603.0
107.00	54728.0
108.00	8457.0
109.10	1554.0
110.00	99232.0
111.00	15725.0
112.00	2037.0
113.10	752.0
114.80	398.0
116.00	2892.0
117.00	45472.0
118.00	4078.0
119.00	619.0
120.00	726.0
121.20	415.0
122.00	3479.0
123.00	5717.0
124.00	2546.0
125.00	2233.0
127.00	178496.0
128.00	13140.0
129.00	67416.0
130.00	5825.0
131.00	974.0
132.10	723.0
133.00	620.0
134.00	1963.0
135.00	5236.0
136.00	2079.0
137.00	3981.0
138.00	790.0
139.00	411.0
140.00	987.0
141.00	7722.0
142.00	3216.0

Instrument :
BNA_N
ClientSampleId :
DFTPP52

143.00	2155.0
144.00	490.0
144.90	354.0
146.00	1457.0
147.00	4712.0
148.00	11019.0
149.00	3038.0
150.10	825.0
151.00	1195.0
151.90	921.0
153.00	2615.0
154.00	1894.0
155.00	4596.0
156.00	6335.0
157.00	1513.0
158.10	1712.0
159.00	1190.0
160.00	2526.0
161.00	3775.0
162.00	1209.0
163.10	445.0
164.00	435.0
165.00	2849.0
166.00	2113.0
167.00	19296.0
168.00	16029.0
169.00	1907.0
170.00	511.0
171.00	815.0
172.00	1364.0
173.00	2119.0
174.00	3579.0
175.00	6232.0
176.10	1799.0
177.00	3038.0
178.00	1023.0
179.00	11591.0
180.00	10201.0
181.00	3953.0
182.10	749.0
182.90	411.0
184.00	1026.0
185.00	6169.0
186.10	42480.0
187.00	11978.0
188.00	1749.0
189.00	2775.0
189.90	400.0
191.00	1406.0
192.00	3811.0
193.10	4264.0
194.10	1162.0
195.00	516.0
196.00	8578.0
197.00	1928.0
198.00	300864.0
199.00	21848.0
200.00	1363.0
201.50	1477.0
203.00	1897.0
204.10	11243.0
205.10	18760.0
206.10	77360.0
207.10	12159.0
208.00	3031.0
209.00	1257.0
210.20	1440.0
211.00	2767.0
212.00	634.0
213.00	352.0
214.90	928.0
216.00	1560.0
217.00	19712.0
218.00	2938.0
221.10	10711.0
221.80	2941.0
223.00	4545.0
224.10	41272.0
225.10	10740.0
226.00	1793.0
227.00	16704.0
228.00	2684.0
229.00	3682.0
230.00	505.0
231.10	1613.0
233.00	317.0
234.00	1134.0
235.00	961.0
236.00	873.0
237.00	2870.0
238.00	494.0
239.00	706.0
240.00	540.0
241.00	907.0
242.10	2759.0

Instrument :
BNA_N
ClientSampleId :
DFTPP52

243.10	2100.0
244.10	32616.0
245.10	4211.0
246.00	6018.0
247.00	1329.0
248.00	347.0
249.00	984.0
251.10	327.0
252.00	404.0
253.10	538.0
255.00	147968.0
256.00	22256.0
257.00	2437.0
258.00	9140.0
259.00	1385.0
260.10	356.0
265.10	3676.0
266.00	659.0
271.00	375.0
272.10	389.0
273.10	4524.0
274.10	11031.0
275.10	60936.0
276.10	8648.0
277.00	5822.0
278.00	1386.0
279.00	306.0
283.00	652.0
284.00	321.0
285.10	1006.0
292.00	331.0
293.00	1384.0
294.10	413.0
295.00	336.0
296.00	17296.0
297.00	2295.0
302.00	312.0
303.00	1882.0
304.00	591.0
314.00	700.0
315.00	1715.0
316.00	1029.0
321.00	437.0
323.10	5166.0
324.10	847.0
327.00	1213.0
328.10	724.0
331.90	315.0
333.20	500.0
334.00	3037.0
335.10	822.0
341.00	817.0
346.00	1171.0
346.90	396.0
352.10	1608.0
353.10	1004.0
354.00	1852.0
365.00	6614.0
366.00	1103.0
371.10	338.0
372.00	2361.0
373.10	733.0
383.00	729.0
390.10	306.0
402.00	885.0
403.10	1475.0
404.00	701.0
421.10	1263.0
422.10	1146.0
423.00	8573.0
424.10	3860.0
425.10	644.0
441.10	20792.0
442.10	139776.0
443.10	25896.0
444.10	2657.0

Instrument :
BNA_N
ClientSampleId :
DFTPP52

m/z	Abundance
37.00	618.0
38.10	1839.0
39.10	10552.0
40.10	1093.0
41.00	556.0
44.00	858.0
49.00	1129.0
50.10	40792.0
51.10	157888.0
52.10	8524.0
53.00	495.0
55.00	1060.0
56.00	4815.0
57.00	11059.0
58.00	523.0
61.00	2283.0
62.00	2503.0
63.00	7066.0
64.00	1157.0
65.10	3987.0
66.00	352.0
67.10	463.0
68.00	3402.0
69.00	193792.0
70.00	1208.0
73.00	2431.0
74.10	21144.0
75.00	32360.0
76.10	11910.0
77.10	240576.0
78.10	17512.0
79.00	15537.0
80.00	11506.0
81.00	16512.0
82.00	4155.0
83.00	3853.0
83.90	598.0
85.00	3160.0
86.00	4900.0
87.00	2460.0
88.00	699.0
89.10	397.0
91.00	3957.0
92.00	4272.0
93.00	28536.0
94.00	1730.0
95.00	497.0
96.00	1511.0
97.00	537.0
98.00	20312.0
99.00	19016.0
100.00	1709.0
101.00	9206.0
102.10	541.0
103.00	3465.0
104.00	5607.0
105.00	5292.0
106.00	2037.0
107.00	69768.0
108.00	10906.0
109.00	2159.0
110.00	123336.0
111.00	19400.0
112.00	2505.0
113.00	707.0
115.10	340.0
116.00	3597.0
117.00	61264.0
118.00	4458.0
119.10	976.0
120.00	1052.0
120.90	420.0
122.00	4836.0
123.00	6736.0
124.00	3238.0
125.00	2700.0
127.00	230464.0
128.00	16784.0
129.00	90096.0
130.00	7393.0
131.10	1406.0
132.00	735.0
133.00	588.0
134.00	2646.0
135.00	6704.0
136.00	2613.0
137.00	4363.0
137.90	914.0
139.00	477.0
140.00	1155.0
141.00	10913.0

Instrument :
BNA_N
ClientSampleId :
DFTPP52

142.00	3681.0
143.00	2212.0
144.00	641.0
144.90	704.0
146.00	2099.0
147.00	5854.0
148.00	14589.0
149.00	3462.0
150.00	879.0
151.10	1742.0
151.90	1020.0
153.00	3214.0
154.00	2579.0
155.10	6349.0
156.00	9511.0
157.00	1865.0
158.00	2103.0
159.00	1805.0
160.00	3736.0
161.00	5112.0
162.00	1486.0
163.10	574.0
163.90	483.0
165.00	4167.0
166.00	3306.0
167.00	23120.0
168.00	17360.0
169.00	2748.0
170.10	876.0
171.00	1082.0
172.00	1932.0
173.10	2715.0
174.00	4874.0
175.10	8763.0
176.10	2518.0
177.00	3735.0
178.00	1747.0
179.00	15747.0
180.00	12476.0
181.00	5530.0
182.00	769.0
183.00	543.0
184.00	1555.0
185.00	8307.0
186.10	57336.0
187.10	17608.0
188.00	2506.0
189.00	4152.0
190.00	667.0
191.00	2129.0
192.00	5345.0
193.10	6123.0
194.00	1546.0
195.10	803.0
196.00	11892.0
197.00	2772.0
198.00	434560.0
199.00	28304.0
200.00	2159.0
201.50	1645.0
203.00	3090.0
204.10	14828.0
205.10	27280.0
206.10	108264.0
207.10	15792.0
208.10	3608.0
209.00	1489.0
210.00	1789.0
211.10	4239.0
213.00	547.0
215.10	1137.0
216.00	2620.0
217.00	28176.0
218.10	3837.0
221.10	15741.0
221.90	3691.0
223.00	6241.0
224.10	59256.0
225.10	15355.0
226.10	1900.0
227.00	23920.0
228.10	3602.0
229.00	5054.0
230.00	669.0
231.10	2065.0
232.10	328.0
233.00	524.0
234.10	1604.0
235.00	1813.0
236.00	1325.0
237.10	3393.0
238.00	519.0
239.10	913.0
240.00	1043.0
241.00	1559.0

Instrument :
BNA_N
ClientSampleId :
DFTPP52

242.10	3124.0
243.10	3911.0
244.10	46232.0
245.10	6270.0
246.10	8674.0
247.00	1627.0
248.10	438.0
249.00	1842.0
249.90	329.0
250.90	381.0
251.90	363.0
253.00	974.0
255.00	218496.0
256.00	32632.0
257.00	3079.0
258.00	13047.0
259.10	2099.0
261.10	490.0
264.10	331.0
265.00	4736.0
266.00	637.0
269.90	329.0
271.10	552.0
272.00	608.0
273.10	7059.0
274.10	16332.0
275.10	98984.0
276.10	12214.0
277.10	8168.0
278.00	1659.0
281.10	311.0
283.00	823.0
284.10	642.0
285.10	1444.0
290.00	357.0
292.00	308.0
293.00	2037.0
293.90	533.0
295.20	501.0
296.00	27480.0
297.00	3224.0
301.00	426.0
302.00	591.0
303.10	2368.0
304.00	824.0
308.00	449.0
314.00	942.0
315.00	3288.0
316.10	1722.0
321.00	857.0
322.00	332.0
323.10	7590.0
324.10	1290.0
326.90	1691.0
328.00	897.0
332.00	535.0
333.10	898.0
334.10	5617.0
335.10	1513.0
341.10	1133.0
342.10	390.0
346.00	2264.0
347.00	573.0
352.10	2571.0
353.10	1752.0
354.10	2610.0
355.10	552.0
365.00	11279.0
366.10	1936.0
371.00	594.0
372.10	4264.0
373.00	1075.0
383.00	1163.0
390.10	469.0
391.00	380.0
402.00	1730.0
403.00	2307.0
404.00	1215.0
405.00	350.0
421.00	2093.0
422.00	2091.0
423.10	15154.0
424.10	7024.0
425.10	1040.0
441.10	42888.0
442.10	272640.0
443.10	51640.0
444.10	4897.0

Instrument :
BNA_N
ClientSampleId :
DFTPP52

m/z	Abundance
37.10	499.0
38.00	1243.0
39.10	7908.0
40.00	1009.0
41.10	373.0
44.00	624.0
49.00	771.0
50.00	28208.0
51.10	107296.0
52.10	5441.0
55.00	1016.0
56.00	3357.0
57.00	7758.0
58.00	490.0
61.00	1740.0
62.10	1669.0
63.00	4875.0
64.00	971.0
65.00	2339.0
67.00	334.0
68.00	2475.0
69.00	133184.0
70.00	612.0
73.00	2000.0
74.10	14571.0
75.00	22944.0
76.10	7757.0
77.10	166656.0
78.10	11576.0
79.00	10938.0
80.00	8148.0
81.00	10976.0
82.00	2473.0
83.00	2546.0
84.00	445.0
85.00	2569.0
86.00	3256.0
87.00	1604.0
88.00	638.0
88.80	311.0
91.00	2574.0
92.00	2623.0
93.00	18984.0
94.00	1275.0
95.00	418.0
96.10	1189.0
97.00	599.0
98.00	14611.0
99.00	13628.0
100.00	951.0
101.00	6063.0
102.00	481.0
103.00	1998.0
104.00	4252.0
105.00	4008.0
106.00	1444.0
107.00	49144.0
108.00	7249.0
109.00	1611.0
110.00	87744.0
111.00	13413.0
112.00	2007.0
113.00	596.0
115.00	359.0
116.00	2478.0
117.00	43776.0
118.00	3806.0
119.00	619.0
120.10	594.0
121.00	377.0
122.00	3178.0
123.00	4193.0
124.00	2587.0
125.00	2217.0
126.10	938.0
127.00	165440.0
128.00	12290.0
129.00	61256.0
130.00	4811.0
131.00	1199.0
132.00	583.0
133.00	666.0
134.00	1895.0
135.00	5286.0
136.00	2166.0
137.00	3459.0
137.90	711.0
140.00	729.0
141.00	7982.0
142.00	2732.0
143.00	2018.0

Instrument :
BNA_N
ClientSampleId :
DFTPP52

143.90	520.0
145.00	445.0
146.00	1566.0
147.00	4231.0
148.00	10294.0
149.00	2255.0
150.00	1108.0
150.90	675.0
151.20	684.0
151.80	743.0
153.00	2439.0
154.00	2401.0
155.00	4592.0
156.00	6475.0
157.10	1333.0
158.00	1986.0
159.00	1090.0
160.00	2526.0
161.10	3744.0
162.00	1147.0
163.10	458.0
164.10	459.0
165.00	3376.0
166.00	2291.0
167.00	17816.0
168.00	13671.0
169.00	1705.0
170.00	577.0
171.00	716.0
172.00	1548.0
173.00	1886.0
174.10	3745.0
175.10	6757.0
176.00	1730.0
177.00	2692.0
178.00	1118.0
179.00	12514.0
180.00	10262.0
181.00	4055.0
182.00	783.0
184.00	1236.0
185.00	6552.0
186.00	44424.0
187.00	12830.0
188.10	1524.0
189.00	2753.0
190.00	531.0
191.00	2011.0
192.00	4555.0
193.10	4467.0
194.00	1194.0
195.00	623.0
196.00	9198.0
197.10	2300.0
198.00	329344.0
199.00	21544.0
200.00	1820.0
201.50	1347.0
203.00	2469.0
204.00	11889.0
205.10	21224.0
206.10	84056.0
207.10	12433.0
208.10	3845.0
209.00	1423.0
210.00	1268.0
211.00	3411.0
211.90	674.0
212.90	357.0
215.00	984.0
216.10	2105.0
217.00	22720.0
218.00	3180.0
219.10	400.0
221.10	12629.0
222.00	2481.0
223.00	5434.0
224.10	48160.0
225.10	12252.0
226.10	1755.0
227.00	19696.0
228.00	2785.0
229.00	3780.0
230.00	460.0
231.10	1422.0
232.00	307.0
233.10	579.0
234.00	954.0
235.00	1328.0
236.00	883.0
237.00	2660.0
238.00	304.0
239.00	780.0
240.00	693.0
241.00	1037.0

Instrument :
BNA_N
ClientSampleId :
DFTPP52

242.00	2529.0
243.10	3016.0
244.10	40824.0
245.10	5690.0
246.00	6624.0
247.00	1455.0
249.00	1172.0
252.10	361.0
253.00	789.0
254.20	1941.0
255.10	181376.0
256.10	26400.0
257.00	2118.0
258.00	10257.0
259.00	1584.0
260.00	347.0
261.10	371.0
265.00	4041.0
266.00	651.0
271.10	656.0
272.00	539.0
273.00	5123.0
274.10	14417.0
275.10	80080.0
276.10	10158.0
277.00	7132.0
278.00	1430.0
283.10	779.0
284.00	510.0
285.00	1471.0
292.10	303.0
293.10	1668.0
294.10	574.0
296.10	22624.0
297.00	3416.0
302.10	364.0
303.10	2674.0
304.00	678.0
310.00	345.0
314.10	1025.0
315.10	2602.0
316.10	1697.0
321.10	665.0
322.00	417.0
323.10	7355.0
324.10	1382.0
327.00	1801.0
328.10	784.0
332.10	541.0
333.10	734.0
334.00	4606.0
335.00	1175.0
341.10	775.0
346.00	1601.0
347.00	386.0
352.00	2350.0
353.10	1658.0
354.10	2136.0
355.00	594.0
365.00	10853.0
366.00	1766.0
371.00	582.0
372.10	3670.0
373.00	1141.0
383.00	958.0
384.00	308.0
390.00	688.0
391.00	321.0
402.10	1556.0
403.00	2646.0
404.00	1134.0
405.10	313.0
421.00	1762.0
422.10	2458.0
423.00	15768.0
424.10	5817.0
425.00	1182.0
441.10	42976.0
442.10	273152.0
443.10	51568.0
444.10	4700.0
445.00	336.0

Instrument :
BNA_N
ClientSampleId :
DFTPP52

m/z	Abundance
39.90	545.0
44.00	585.0
73.00	751.0
132.90	369.0
147.00	326.0
185.10	361.0
186.00	461.0
207.00	2269.0
208.00	707.0

Instrument :
BNA_N
ClientSampleId :
DFTPP52