

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN030521\
 Data File : BN013854.D
 Acq On : 05 Mar 2021 10:43
 Operator : CG/JU
 Sample : PB134916BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK16

Quant Time: Mar 05 11:15:21 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN030421MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Mar 04 14:09:31 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.72	152	139351	20.00	ng/ul	0.00
18) Naphthalene-d8	10.50	136	573166	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.36	164	338654	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.10	188	734450	20.00	ng/ul	0.00
78) Chrysene-d12	21.27	240	716553	20.00	ng/ul	0.00
86) Perylene-d12	23.50	264	754595	20.00	ng/ul	0.00

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.22	96	22779	6.49	ng/uL	0.00
5) Phenol-d5	6.88	99	381494	33.85	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.06	67	248395	35.70	ng/ul	0.00
9) 2-Chlorophenol-d4	7.25	132	311766	36.17	ng/ul	0.00
13) 4-Methylphenol-d8	8.42	113	293669	34.41	ng/ul	0.00
19) Nitrobenzene-d5	8.87	128	142967	37.27	ng/ul	0.00
22) 2-Nitrophenol-d4	9.59	143	137410	37.50	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.12	165	277644	34.69	ng/ul	0.00
29) 4-Chloroaniline-d4	10.63	131	425895	42.97	ng/ul	0.00
44) Dimethylphthalate-d6	13.77	166	975685	42.91	ng/ul	0.00
47) Acenaphthylene-d8	14.05	160	1167678	38.83	ng/ul	0.00
52) 4-Nitrophenol-d4	14.55	143	148093	34.72	ng/ul	0.00
58) Fluorene-d10	15.35	176	810641	40.37	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.46	200	99451	29.13	ng/ul	0.00
71) Anthracene-d10	17.20	188	1317607	41.05	ng/ul	0.00
79) Pyrene-d10	19.49	212	1526710	43.80	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.37	264	1568811	41.76	ng/ul	0.00

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN030521\
Data File : BN013854.D
Acq On : 05 Mar 2021 10:43
Operator : CG/JU
Sample : PB134916BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK16

Quant Time: Mar 05 11:15:21 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN030421MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Mar 04 14:09:31 2021
Response via : Initial Calibration

