

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030524\
 Data File : BN030275.D
 Acq On : 05 Mar 2024 16:59
 Operator : MA/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN030524

Quant Time: Mar 05 18:03:17 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN030524.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 05 17:44:16 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	135	0.00
2	1,4-Dioxane	0.400	0.371	7.3	118	0.00
3	n-Nitrosodimethylamine	0.400	0.349	12.8	119	0.00
4 S	2-Fluorophenol	0.400	0.370	7.5	125	0.00
5 S	Phenol-d6	0.400	0.345	13.8	118	0.00
6	bis(2-Chloroethyl)ether	0.400	0.372	7.0	127	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	135	0.00
8 S	Nitrobenzene-d5	0.400	0.384	4.0	128	0.00
9	Naphthalene	0.400	0.369	7.8	119	-0.01
10	Hexachlorobutadiene	0.400	0.352	12.0	112	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.376	6.0	121	0.00
12	2-Methylnaphthalene	0.400	0.365	8.8	118	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	131	0.00
14 S	2,4,6-Tribromophenol	0.400	0.329	17.8	102	0.00
15 S	2-Fluorobiphenyl	0.400	0.392	2.0	124	0.00
16	Acenaphthylene	0.400	0.387	3.3	125	0.00
17	Acenaphthene	0.400	0.364	9.0	115	0.00
18	Fluorene	0.400	0.346	13.5	110	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	123	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.330	17.5	109	0.00
21	4-Bromophenyl-phenylether	0.400	0.357	10.8	105	0.00
22	Hexachlorobenzene	0.400	0.373	6.8	111	0.00
23	Atrazine	0.400	0.365	8.8	106	0.00
24	Pentachlorophenol	0.400	0.317	20.8	93	0.00
25	Phenanthrene	0.400	0.376	6.0	113	0.00
26	Anthracene	0.400	0.372	7.0	112	0.00
27 SURR	Fluoranthene-d10	0.400	0.375	6.3	112	0.00
28	Fluoranthene	0.400	0.356	11.0	105	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	121	0.00
30	Pyrene	0.400	0.364	9.0	105	0.00
31 S	Terphenyl-d14	0.400	0.407	-1.7	119	0.00
32	Benzo(a)anthracene	0.400	0.373	6.8	112	0.00
33	Chrysene	0.400	0.377	5.8	110	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.338	15.5	97	0.00
35 I	Perylene-d12	0.400	0.400	0.0	122	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.367	8.3	111	0.00
37	Benzo(b)fluoranthene	0.400	0.348	13.0	104	0.00
38	Benzo(k)fluoranthene	0.400	0.358	10.5	106	0.00
39 C	Benzo(a)pyrene	0.400	0.371	7.3	112	0.00
40	Dibenzo(a,h)anthracene	0.400	0.344	14.0	109	0.00
41	Benzo(g,h,i)perylene	0.400	0.338	15.5	102	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN030524