

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030524\
 Data File : BN030280.D
 Acq On : 06 Mar 2024 01:29
 Operator : MA/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 06 02:19:55 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN030524.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 05 17:44:16 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	127	0.00
2	1,4-Dioxane	0.519	0.526	-1.3	120	0.00
3	n-Nitrosodimethylamine	0.548	0.564	-2.9	131	0.00
4 S	2-Fluorophenol	1.050	1.126	-7.2	135	0.00
5 S	Phenol-d6	1.105	1.165	-5.4	135	0.00
6	bis(2-Chloroethyl)ether	0.876	0.943	-7.6	137	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	134	0.00
8 S	Nitrobenzene-d5	0.365	0.359	1.6	130	0.00
9	Naphthalene	1.117	1.168	-4.6	134	-0.01
10	Hexachlorobutadiene	0.255	0.270	-5.9	134	0.00
11 SURR	2-Methylnaphthalene-d10	0.580	0.613	-5.7	135	0.00
12	2-Methylnaphthalene	0.689	0.703	-2.0	131	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	129	0.00
14 S	2,4,6-Tribromophenol	0.206	0.202	1.9	120	0.00
15 S	2-Fluorobiphenyl	1.661	1.735	-4.5	131	0.00
16	Acenaphthylene	1.930	1.977	-2.4	130	0.00
17	Acenaphthene	1.246	1.313	-5.4	131	0.00
18	Fluorene	1.499	1.573	-4.9	131	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	124	0.00
20	4,6-Dinitro-2-methylphenol	0.083	0.081	2.4	129	0.00
21	4-Bromophenyl-phenylether	0.244	0.241	1.2	117	0.01
22	Hexachlorobenzene	0.283	0.297	-4.9	127	0.00
23	Atrazine	0.217	0.213	1.8	116	0.00
24	Pentachlorophenol	0.147	0.147	0.0	118	0.00
25	Phenanthrene	1.144	1.174	-2.6	125	0.00
26	Anthracene	1.051	1.026	2.4	118	0.00
27 SURR	Fluoranthene-d10	1.098	1.123	-2.3	123	0.00
28	Fluoranthene	1.393	1.400	-0.5	120	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	125	0.00
30	Pyrene	1.679	1.678	0.1	119	0.00
31 S	Terphenyl-d14	1.037	1.019	1.7	119	0.00
32	Benzo(a)anthracene	1.402	1.351	3.6	120	0.00
33	Chrysene	1.518	1.615	-6.4	128	0.00
34	Bis(2-ethylhexyl)phthalate	1.118	1.022	8.6	109	0.00
35 I	Perylene-d12	1.000	1.000	0.0	127	0.00
36	Indeno(1,2,3-cd)pyrene	1.676	1.687	-0.7	127	0.00
37	Benzo(b)fluoranthene	1.398	1.387	0.8	124	0.00
38	Benzo(k)fluoranthene	1.420	1.451	-2.2	126	0.00
39 C	Benzo(a)pyrene	1.235	1.259	-1.9	128	0.00
40	Dibenzo(a,h)anthracene	1.317	1.365	-3.6	137	0.00
41	Benzo(g,h,i)perylene	1.519	1.550	-2.0	128	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4