

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030621\
 Data File : BN013862.D
 Acq On : 05 Mar 2021 16:39
 Operator : CG/JU
 Sample : M1574-03
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 GWBR3-160-170-030421

Quant Time: Mar 05 17:11:02 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN022421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 05 13:34:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.724	152	5521	0.40	ng	0.00
7) Naphthalene-d8	10.505	136	21253	0.40	ng	# 0.00
13) Acenaphthene-d10	14.346	164	12125	0.40	ng	0.00
19) Phenanthrene-d10	17.092	188	26881	0.40	ng	0.00
29) Chrysene-d12	21.271	240	28832	0.40	ng	# 0.00
36) Perylene-d12	23.499	264	23653	0.40	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.316	112	1712	0.12	ng	0.00
5) Phenol-d6	6.887	99	1272	0.06	ng	0.00
8) Nitrobenzene-d5	8.870	82	4954	0.36	ng	0.00
11) 2-Methylnaphthalene-d10	12.093	152	10713	0.31	ng	0.00
14) 2,4,6-Tribromophenol	15.843	330	1817	0.37	ng	0.00
15) 2-Fluorobiphenyl	12.976	172	16392	0.38	ng	0.00
27) Fluoranthene-d10	19.124	212	30457	0.38	ng	0.00
31) Terphenyl-d14	19.724	244	35812	0.56	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.255	88	36440	5.46	ng	96
9) Naphthalene	10.543	128	1405	0.03	ng	# 81
24) Pentachlorophenol	16.739	266	103	0.02	ng	92
25) Phenanthrene	17.128	178	8247	0.12	ng	98
28) Fluoranthene	19.154	202	8217	0.10	ng	96
30) Pyrene	19.516	202	6385	0.07	ng	97
32) Benzo(a)anthracene	21.261	228	3151	0.04	ng	# 91
33) Chrysene	21.303	228	4147	0.05	ng	# 94
34) Bis(2-ethylhexyl)phtha...	21.197	149	6775	0.16	ng	98
37) Benzo(b)fluoranthene	22.832	252	3714	0.05	ng	# 64
38) Benzo(k)fluoranthene	22.873	252	2532	0.03	ng	# 56
39) Benzo(a)pyrene	23.400	252	1858	0.03	ng	# 31
41) Benzo(g,h,i)perylene	26.419	276	1641	0.02	ng	# 74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030621\
Data File : BN013862.D
Acq On : 05 Mar 2021 16:39
Operator : CG/JU
Sample : M1574-03
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
GWBR3-160-170-030421

Quant Time: Mar 05 17:11:02 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN022421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 05 13:34:35 2021
Response via : Initial Calibration

