




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030621\
Data File : BN013856.D
Acq On : 05 Mar 2021 12:11
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Mar 05 13:12:19 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 05 13:12:06 2021
Response via : Initial Calibration






TIC: BN013856.D\data.ms

(70) Pentachlorophenol (C)**16.751min (0.000) 18722.30 ng****response 173228**

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	66.55
264.00	61.60	62.74
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030621\
Data File : BN013856.D
Acq On : 05 Mar 2021 12:11
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Mar 05 13:12:19 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 05 13:12:06 2021
Response via : Initial Calibration



TIC: BN013856.D\data.ms

(77) Benzidine

19.345min (0.000) 26890.89 ng

response 1292289

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	15.63
183.00	13.20	11.27
0.00	0.00	0.00