

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030621\
 Data File : BN013857.D
 Acq On : 05 Mar 2021 12:47
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Manual Integrations
 APPROVED

mohammad

Quant Time: Mar 05 13:34:52 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN022421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 05 13:34:35 2021
 Response via : Initial Calibration

3/8/2021 9:35:47 AM

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.724	152	5505	0.40	ng	0.00
7) Naphthalene-d8	10.505	136	20767	0.40	ng	0.00
13) Acenaphthene-d10	14.346	164	11484	0.40	ng	0.00
19) Phenanthrene-d10	17.092	188	25469	0.40	ng	0.00
29) Chrysene-d12	21.271	240	27839	0.40	ng	0.00
36) Perylene-d12	23.503	264	24591	0.40	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.316	112	5239	0.35	ng	0.00
5) Phenol-d6	6.887	99	6777	0.34	ng	0.00
8) Nitrobenzene-d5	8.870	82	5462	0.41	ng	0.00
11) 2-Methylnaphthalene-d10	12.093	152	13239	0.39	ng	0.00
14) 2,4,6-Tribromophenol	15.843	330	1416	0.30	ng	0.00
15) 2-Fluorobiphenyl	12.976	172	17830	0.44	ng	0.00
27) Fluoranthene-d10	19.129	212	28052	0.37	ng	0.00
31) Terphenyl-d14	19.729	244	22354	0.36	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.255	88	2888	0.43	ng	99
3) n-Nitrosodimethylamine	3.585	42	1749	0.34	ng	# 99
6) bis(2-Chloroethyl)ether	7.152	93	6092	0.44	ng	96
9) Naphthalene	10.555	128	21300	0.42	ng	100
10) Hexachlorobutadiene	10.847	225	4039	0.41	ng	# 99
12) 2-Methylnaphthalene	12.168	142	14320	0.40	ng	99
16) Acenaphthylene	14.067	152	17688	0.34	ng	99
17) Acenaphthene	14.409	154	13228	0.41	ng	99
18) Fluorene	15.399	166	16426	0.40	ng	99
20) 4,6-Dinitro-2-methylph...	15.475	198	1004	0.57	ng	# 85
21) 4-Bromophenyl-phenylether	16.301	248	5402	0.39	ng	# 74
22) Hexachlorobenzene	16.410	284	6355	0.44	ng	99
23) Atrazine	16.556	200	3306	0.26	ng	# 91
24) Pentachlorophenol	16.751	266	1546	0.33	ng	95
25) Phenanthrene	17.140	178	28130	0.42	ng	99
26) Anthracene	17.226	178	22415	0.35	ng	100
28) Fluoranthene	19.158	202	31523	0.40	ng	99
30) Pyrene	19.521	202	32202	0.38	ng	100
32) Benzo(a)anthracene	21.260	228	28848	0.34	ng	99
33) Chrysene	21.314	228	34300	0.41	ng	97
34) Bis(2-ethylhexyl)phtha...	21.197	149	8129	0.20	ng	100
35) Indeno(1,2,3-cd)pyrene	25.745	276	38670	0.38	ng	99
37) Benzo(b)fluoranthene	22.835	252	33815	0.43	ng	97
38) Benzo(k)fluoranthene	22.880	252	35984m	0.47	ng	
39) Benzo(a)pyrene	23.407	252	30851	0.44	ng	# 94
40) Dibenzo(a,h)anthracene	25.758	278	31918	0.43	ng	97
41) Benzo(g,h,i)perylene	26.426	276	32510	0.43	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030621\
Data File : BN013857.D
Acq On : 05 Mar 2021 12:47
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED


mohammad

Quant Time: Mar 05 13:34:52 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN022421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 05 13:34:35 2021
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 0.40 ng
RT: 7.724 min Scan# 614
Delta R.T. 0.000 min
Lab File: BN013857.D
Acq: 05 Mar 2021 12:47



Tgt Ion: 152 Resp: 5505
Ion Ratio Lower Upper
152 100
150 154.1 121.8 182.8
115 58.6 46.6 69.8




Instrument :
BNA_N
ClientSampled :
SSTDCCC0.4


Manual Integrations
APPROVED

mohammad



3/8/2021 9:35:47 AM




#2
1,4-Dioxane
Concen: 0.43 ng
RT: 3.255 min Scan# 59
Delta R.T. 0.000 min
Lab File: BN013857.D
Acq: 05 Mar 2021 12:47



Tgt Ion: 88 Resp: 2888
Ion Ratio Lower Upper
88 100
58 79.6 64.6 96.8
43 32.8 26.6 40.0









#3
n-Nitrosodimethylamine
Concen: 0.34 ng
RT: 3.585 min Scan# 100
Delta R.T. 0.000 min
Lab File: BN013857.D
Acq: 05 Mar 2021 12:47

Tgt Ion: 42 Resp: 1749
Ion Ratio Lower Upper
42 100
74 119.0 95.0 142.6
44 13.3 5.8 8.8#

Instrument :
BNA_N
ClientSampled :
SSTDCCC0.4




Manual Integrations
APPROVED


mohammad



#4
2-Fluorophenol
Concen: 0.35 ng
RT: 5.316 min Scan# 315
Delta R.T. 0.000 min
Lab File: BN013857.D
Acq: 05 Mar 2021 12:47

Tgt Ion: 112 Resp: 5239
Ion Ratio Lower Upper
112 100
64 59.8 44.3 66.5
63 31.8 23.0 34.6









#5
Phenol-d6
Concen: 0.34 ng
RT: 6.887 min Scan# 510
Delta R.T. 0.000 min
Lab File: BN013857.D
Acq: 05 Mar 2021 12:47

Tgt Ion: 99 Resp: 6777
Ion Ratio Lower Upper
99 100
42 22.9 15.2 22.8#
71 33.1 26.1 39.1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Manual Integrations
APPROVED


mohammad



#6
bis(2-Chloroethyl)ether
Concen: 0.44 ng
RT: 7.152 min Scan# 543
Delta R.T. 0.000 min
Lab File: BN013857.D
Acq: 05 Mar 2021 12:47

Tgt Ion: 93 Resp: 6092
Ion Ratio Lower Upper
93 100
63 80.9 60.8 91.2
95 32.8 26.0 39.0







#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.505 min Scan# 860
Delta R.T. 0.000 min
Lab File: BN013857.D
Acq: 05 Mar 2021 12:47

Tgt Ion:	136	Resp:	20767
Ion Ratio	Lower	Upper	
136	100		
137	11.2	9.0	13.6
54	6.4	6.4	9.6
68	5.0	4.5	6.7



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED

mohammad







3/8/2021 9:35:47 AM



#8
Nitrobenzene-d5
Concen: 0.41 ng
RT: 8.870 min Scan# 731
Delta R.T. 0.000 min
Lab File: BN013857.D
Acq: 05 Mar 2021 12:47

Tgt Ion:	82	Resp:	5462
Ion Ratio	Lower	Upper	
82	100		
128	44.6	39.2	58.8
54	48.1	40.1	60.1







#9
Naphthalene
Concen: 0.42 ng
RT: 10.555 min Scan# 864
Delta R.T. 0.000 min
Lab File: BN013857.D
Acq: 05 Mar 2021 12:47

Tgt Ion:	128	Resp:	21300
Ion Ratio	Lower	Upper	
128	100		
129	11.3	9.0	13.4
127	12.9	10.4	15.6



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED

mohammad







3/8/2021 9:35:47 AM



#10
Hexachlorobutadiene
Concen: 0.41 ng
RT: 10.847 min Scan# 887
Delta R.T. 0.000 min
Lab File: BN013857.D
Acq: 05 Mar 2021 12:47

Tgt Ion:	225	Resp:	4039
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	64.4	51.2	76.8







#11
2-Methylnaphthalene-d10
Concen: 0.39 ng
RT: 12.093 min Scan# 1071
Delta R.T. 0.000 min
Lab File: BN013857.D
Acq: 05 Mar 2021 12:47

Tgt Ion:152 Resp: 13239
Ion Ratio Lower Upper
152 100
151 20.4 16.2 24.2



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED

mohammad







3/8/2021 9:35:47 AM



#12
2-Methylnaphthalene
Concen: 0.40 ng
RT: 12.168 min Scan# 1088
Delta R.T. 0.000 min
Lab File: BN013857.D
Acq: 05 Mar 2021 12:47

Tgt Ion:142 Resp: 14320
Ion Ratio Lower Upper
142 100
141 87.7 70.5 105.7
115 35.2 27.4 41.0









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.346 min Scan# 1309
Delta R.T. 0.000 min
Lab File: BN013857.D
Acq: 05 Mar 2021 12:47

Tgt Ion:	164	Resp:	11484
Ion Ratio	Lower	Upper	
164	100		
162	104.6	79.6	119.4
160	49.8	36.0	54.0

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Manual Integrations
APPROVED


mohammad



#14
2,4,6-Tribromophenol
Concen: 0.30 ng
RT: 15.843 min Scan# 1427
Delta R.T. 0.000 min
Lab File: BN013857.D
Acq: 05 Mar 2021 12:47

Tgt Ion:	330	Resp:	1416
Ion Ratio	Lower	Upper	
330	100		
332	0.0	0.0	0.0
141	36.5	26.8	40.2







#15
2-Fluorobiphenyl
Concen: 0.44 ng
RT: 12.976 min Scan# 1201
Delta R.T. 0.000 min
Lab File: BN013857.D
Acq: 05 Mar 2021 12:47

Tgt Ion:	172	Resp:	17830
Ion Ratio	Lower	Upper	
172	100		
171	34.7	28.6	42.8
170	22.8	18.9	28.3



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED

mohammad







3/8/2021 9:35:47 AM



#16
Acenaphthylene
Concen: 0.34 ng
RT: 14.067 min Scan# 1287
Delta R.T. 0.000 min
Lab File: BN013857.D
Acq: 05 Mar 2021 12:47


Tgt Ion:	152	Resp:	17688
Ion Ratio	Lower	Upper	
152	100		
151	19.1	15.7	23.5
153	13.0	10.5	15.7





#17
Acenaphthene
Concen: 0.41 ng
RT: 14.409 min Scan# 1314
Delta R.T. 0.000 min
Lab File: BN013857.D
Acq: 05 Mar 2021 12:47



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Tgt Ion	Ratio	Lower	Upper
154	100		
153	109.6	88.4	132.6
152	55.2	43.0	64.4

Manual Integrations
APPROVED


mohammad





3/8/2021 9:35:47 AM




#18
Fluorene
Concen: 0.40 ng
RT: 15.399 min Scan# 1392
Delta R.T. 0.000 min
Lab File: BN013857.D
Acq: 05 Mar 2021 12:47



Tgt Ion	Ratio	Lower	Upper
166	100		
165	98.0	77.8	116.8
167	13.6	10.6	16.0







#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.092 min Scan# 1529
Delta R.T. 0.000 min
Lab File: BN013857.D
Acq: 05 Mar 2021 12:47

Tgt Ion:188	Resp:	25469
Ion Ratio	Lower	Upper
188	100	
94	0.0	0.0
80	12.0	9.0
		13.6



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED

mohammad







3/8/2021 9:35:47 AM



#20
4,6-Dinitro-2-methylphenol
Concen: 0.57 ng
RT: 15.475 min Scan# 1398
Delta R.T. 0.000 min
Lab File: BN013857.D
Acq: 05 Mar 2021 12:47

Tgt Ion:198	Resp:	1004
Ion Ratio	Lower	Upper
198	100	
51	94.4	61.7
105	45.5	33.0
		49.6







#21
4-Bromophenyl-phenylether
Concen: 0.39 ng
RT: 16.301 min Scan# 1464
Delta R.T. 0.000 min
Lab File: BN013857.D
Acq: 05 Mar 2021 12:47

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion	Ratio	Lower	Upper
248	100		
250	106.3	75.5	113.3
141	31.3	53.8	80.8#

Manual Integrations
APPROVED

mohammad







3/8/2021 9:35:47 AM



#22
Hexachlorobenzene
Concen: 0.44 ng
RT: 16.410 min Scan# 1473
Delta R.T. 0.000 min
Lab File: BN013857.D
Acq: 05 Mar 2021 12:47


Tgt Ion	Ratio	Lower	Upper
284	100		
142	35.0	27.9	41.9
249	30.7	24.2	36.2





#23
Atrazine
Concen: 0.26 ng
RT: 16.556 min Scan# 1485
Delta R.T. 0.000 min
Lab File: BN013857.D
Acq: 05 Mar 2021 12:47



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Tgt Ion:200 Resp: 3306
Ion Ratio Lower Upper
200 100
173 31.4 18.6 28.0#
215 47.9 41.0 61.6

Manual Integrations
APPROVED


mohammad





3/8/2021 9:35:47 AM




#24
Pentachlorophenol
Concen: 0.33 ng
RT: 16.751 min Scan# 1501
Delta R.T. 0.000 min
Lab File: BN013857.D
Acq: 05 Mar 2021 12:47



Tgt Ion:266 Resp: 1546
Ion Ratio Lower Upper
266 100
264 62.2 51.6 77.4
268 62.9 54.6 82.0







#25
Phenanthrene
Concen: 0.42 ng
RT: 17.140 min Scan# 1533
Delta R.T. 0.000 min
Lab File: BN013857.D
Acq: 05 Mar 2021 12:47

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion:178 Resp: 28130
Ion Ratio Lower Upper
178 100
176 19.0 15.0 22.4
179 15.5 12.3 18.5

Manual Integrations
APPROVED

mohammad







3/8/2021 9:35:47 AM



#26
Anthracene
Concen: 0.35 ng
RT: 17.226 min Scan# 1540
Delta R.T. 0.000 min
Lab File: BN013857.D
Acq: 05 Mar 2021 12:47

Tgt Ion:178 Resp: 22415
Ion Ratio Lower Upper
178 100
176 18.0 14.6 21.8
179 15.3 12.2 18.2







#27
Fluoranthene-d10
Concen: 0.37 ng
RT: 19.129 min Scan# 1829
Delta R.T. 0.000 min
Lab File: BN013857.D
Acq: 05 Mar 2021 12:47

Tgt Ion	Ratio	Lower	Upper
212	100		
106	12.6	9.7	14.5
104	7.1	5.5	8.3



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED

mohammad







3/8/2021 9:35:47 AM



#28
Fluoranthene
Concen: 0.40 ng
RT: 19.158 min Scan# 1835
Delta R.T. 0.000 min
Lab File: BN013857.D
Acq: 05 Mar 2021 12:47

Tgt Ion	Ratio	Lower	Upper
202	100		
101	10.9	8.1	12.1
203	17.1	13.9	20.9







#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.271 min Scan# 2125
Delta R.T. 0.000 min
Lab File: BN013857.D
Acq: 05 Mar 2021 12:47

Tgt Ion:240	Resp:	27839
Ion Ratio	Lower	Upper
240	100	
120	13.6	10.2 15.4
236	26.3	21.3 31.9



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED

mohammad







3/8/2021 9:35:47 AM




#30
Pyrene
Concen: 0.38 ng
RT: 19.521 min Scan# 1908
Delta R.T. 0.000 min
Lab File: BN013857.D
Acq: 05 Mar 2021 12:47

Tgt Ion:202	Resp:	32202
Ion Ratio	Lower	Upper
202	100	
200	20.3	16.3 24.5
203	17.6	14.0 21.0






#31
Terphenyl-d14
Concen: 0.36 ng
RT: 19.729 min Scan# 1950
Delta R.T. 0.000 min
Lab File: BN013857.D
Acq: 05 Mar 2021 12:47



Tgt Ion:244 Resp: 22354
Ion Ratio Lower Upper
244 100
212 7.6 6.1 9.1
122 10.7 9.8 14.8




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Manual Integrations
APPROVED

mohammad



3/8/2021 9:35:47 AM




#32
Benzo(a)anthracene
Concen: 0.34 ng
RT: 21.260 min Scan# 2124
Delta R.T. 0.000 min
Lab File: BN013857.D
Acq: 05 Mar 2021 12:47



Tgt Ion:228 Resp: 28848
Ion Ratio Lower Upper
228 100
226 26.3 20.5 30.7
229 19.7 16.2 24.4







#33
Chrysene
Concen: 0.41 ng
RT: 21.314 min Scan# 2129
Delta R.T. 0.000 min
Lab File: BN013857.D
Acq: 05 Mar 2021 12:47

Tgt Ion:	228	Resp:	34300
Ion Ratio	Lower	Upper	
228	100		
226	28.8	24.3	36.5
229	19.8	15.1	22.7



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED

mohammad







3/8/2021 9:35:47 AM




#34
Bis(2-ethylhexyl)phthalate
Concen: 0.20 ng
RT: 21.197 min Scan# 2118
Delta R.T. 0.000 min
Lab File: BN013857.D
Acq: 05 Mar 2021 12:47

Tgt Ion:	149	Resp:	8129
Ion Ratio	Lower	Upper	
149	100		
167	27.6	22.1	33.1
279	4.0	3.5	5.3









#35
Indeno(1,2,3-cd)pyrene
Concen: 0.38 ng
RT: 25.745 min Scan# 3207
Delta R.T. 0.000 min
Lab File: BN013857.D
Acq: 05 Mar 2021 12:47





Tgt Ion: 276 Resp: 38670
Ion Ratio Lower Upper
276 100
138 23.9 19.8 29.6
277 24.8 19.9 29.9



#36
Perylene-d12
Concen: 0.40 ng
RT: 23.503 min Scan# 2552
Delta R.T. 0.000 min
Lab File: BN013857.D
Acq: 05 Mar 2021 12:47



Tgt Ion: 264 Resp: 24591
Ion Ratio Lower Upper
264 100
260 24.7 19.3 28.9
265 34.6 66.2 99.4#




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED

mohammad

3/8/2021 9:35:47 AM



#37
Benzo(b)fluoranthene
Concen: 0.43 ng
RT: 22.835 min Scan# 2357
Delta R.T. 0.000 min
Lab File: BN013857.D
Acq: 05 Mar 2021 12:47





Tgt Ion:252 Resp: 33815
Ion Ratio Lower Upper
252 100
253 23.5 20.4 30.6
125 13.5 11.6 17.4

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED




mohammad


3/8/2021 9:35:47 AM




#38
Benzo(k)fluoranthene
Concen: 0.47 ng m
RT: 22.880 min Scan# 2370
Delta R.T. 0.000 min
Lab File: BN013857.D
Acq: 05 Mar 2021 12:47

Tgt Ion:252 Resp: 35984
Ion Ratio Lower Upper
252 100
253 23.4 20.5 30.7
125 11.2 12.3 18.5#







#39
Benzo(a)pyrene
Concen: 0.44 ng
RT: 23.407 min Scan# 2524
Delta R.T. 0.000 min
Lab File: BN013857.D
Acq: 05 Mar 2021 12:47



Tgt Ion:252 Resp: 30851
Ion Ratio Lower Upper
252 100
253 25.1 21.8 32.6
125 22.4 14.6 21.8#




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Manual Integrations
APPROVED

mohammad



3/8/2021 9:35:47 AM




#40
Dibenzo(a,h)anthracene
Concen: 0.43 ng
RT: 25.758 min Scan# 3211
Delta R.T. 0.000 min
Lab File: BN013857.D
Acq: 05 Mar 2021 12:47




Tgt Ion:278 Resp: 31918
Ion Ratio Lower Upper
278 100
139 16.4 13.1 19.7
279 24.5 21.4 32.0







#41
Benzo(g,h,i)perylene
Concen: 0.43 ng
RT: 26.426 min Scan# 3406
Delta R.T. 0.000 min
Lab File: BN013857.D
Acq: 05 Mar 2021 12:47



Tgt Ion:	276	Resp:	32510
Ion Ratio	Lower	Upper	
276	100		
277	24.5	20.0	30.0
138	21.6	17.8	26.8



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED

mohammad

3/8/2021 9:35:47 AM