

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030623\
 Data File : BN024127.D
 Acq On : 06 Mar 2023 09:27
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020257

Quant Time: Mar 07 04:28:19 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN030323.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Mar 07 04:27:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.058	152	273689	20.000	ng/u1	0.00
20) Naphthalene-d8	10.881	136	1160264	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.698	164	717953	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.439	188	1502550	20.000	ng/u1	0.00
79) Chrysene-d12	21.633	240	1412203	20.000	ng/u1	0.00
88) Perylene-d12	24.151	264	1618435	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.411	96	56931	8.115	ng/uL	0.00
4) Pyridine-d5	3.834	84	396780	20.438	ng/u1	0.00
7) Phenol-d5	7.199	99	497637	20.271	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.381	67	331951	21.115	ng/u1	0.00
11) 2-Chlorophenol-d4	7.581	132	393281	20.860	ng/u1	0.00
15) 4-Methylphenol-d8	8.763	113	396361	20.310	ng/u1	0.00
21) Nitrobenzene-d5	9.228	128	164171	21.800	ng/u1	0.00
24) 2-Nitrophenol-d4	9.952	143	178669	21.757	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.493	165	378410	21.354	ng/u1	0.00
31) 4-Chloroaniline-d4	11.010	131	566472	21.426	ng/u1	0.00
46) Dimethylphthalate-d6	14.104	166	1155614	20.856	ng/u1	0.00
49) Acenaphthylene-d8	14.387	160	1377846	21.363	ng/u1	0.00
54) 4-Nitrophenol-d4	14.869	143	217188	20.514	ng/u1	0.00
60) Fluorene-d10	15.687	176	983850	21.127	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.792	200	150943	18.646	ng/u1	0.00
73) Anthracene-d10	17.539	188	1499942	21.167	ng/u1	0.00
81) Pyrene-d10	19.828	212	1720176	20.842	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.986	264	1771608	22.020	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.446	88	58468	7.696	ng/uL	96
5) Pyridine	3.852	79	408072	20.305	ng/u1	94
6) Benzaldehyde	7.187	77	301561	25.283	ng/u1	99
8) Phenol	7.228	94	519970	20.432	ng/u1	100
10) Bis(2-Chloroethyl)ether	7.475	93	434277	21.059	ng/u1	100
12) 2-Chlorophenol	7.616	128	404705	20.676	ng/u1	99
13) 2-Methylphenol	8.493	108	391358	20.450	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.593	45	678597	20.741	ng/u1	99
16) Acetophenone	8.887	105	644350	21.268	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.875	70	310717	20.887	ng/u1	100
18) 4-Methylphenol	8.828	108	431588	20.490	ng/u1	98
19) Hexachloroethane	9.152	117	161640	21.029	ng/u1	97
22) Nitrobenzene	9.269	77	448429	21.996	ng/u1	98
23) Isophorone	9.799	82	946279	21.237	ng/u1	99
25) 2-Nitrophenol	9.987	139	210966	22.363	ng/u1	96
26) 2,4-Dimethylphenol	10.040	107	459189	21.357	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.281	93	585767	21.172	ng/u1	98
29) 2,4-Dichlorophenol	10.516	162	375541	20.997	ng/u1	99
30) Naphthalene	10.928	128	1341160	21.005	ng/u1	99
32) 4-Chloroaniline	11.034	127	577179	21.649	ng/u1	99
33) Hexachlorobutadiene	11.216	225	224030	20.553	ng/u1	96
34) Caprolactam	11.793	113	65060	10.667	ng/u1	96
35) 4-Chloro-3-methylphenol	12.146	107	421122	21.373	ng/u1	97
36) 2-Methylnaphthalene	12.534	142	892696	21.235	ng/u1	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030623\
 Data File : BN024127.D
 Acq On : 06 Mar 2023 09:27
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1


Instrument :
 BNA_N
 ClientSampleId :
 SSTD020257


Quant Time: Mar 07 04:28:19 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN030323.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Mar 07 04:27:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.751	142	895139	21.247	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.893	216	437099	20.232	ng/ul	95
40) Hexachlorocyclopentadiene	12.881	237	220216	18.913	ng/ul	96
41) 2,4,6-Trichlorophenol	13.128	196	290336	20.514	ng/ul	99
42) 2,4,5-Trichlorophenol	13.198	196	316504	20.764	ng/ul	98
43) 1,1'-Biphenyl	13.534	154	1192010	20.843	ng/ul	98
44) 2-Chloronaphthalene	13.575	162	922010	20.692	ng/ul	98
45) 2-Nitroaniline	13.775	65	208985	21.527	ng/ul	99
47) Dimethylphthalate	14.151	163	1157897	20.733	ng/ul	100
48) 2,6-Dinitrotoluene	14.269	165	178567	21.722	ng/ul	94
50) Acenaphthylene	14.416	152	1475528	21.230	ng/ul	99
51) 3-Nitroaniline	14.593	138	211206	21.137	ng/ul	99
52) Acenaphthene	14.757	153	997069	21.031	ng/ul	98
53) 2,4-Dinitrophenol	14.793	184	95351	16.754	ng/ul	99
55) 4-Nitrophenol	14.887	109	166076	20.591	ng/ul	96
56) Dibenzofuran	15.092	168	1363053	20.465	ng/ul	99
57) 2,4-Dinitrotoluene	15.045	165	282794	21.824	ng/ul	99
58) 2,3,4,6-Tetrachlorophenol	15.310	232	269410	20.561	ng/ul	97
59) Diethylphthalate	15.510	149	1149829	20.655	ng/ul	99
61) Fluorene	15.740	166	1106992	21.215	ng/ul	96
62) 4-Chlorophenyl-phenyle...	15.734	204	518499	20.782	ng/ul	97
63) 4-Nitroaniline	15.751	138	211718	21.910	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.810	198	159397	19.465	ng/ul	99
67) N-Nitrosodiphenylamine	15.945	169	943948	21.215	ng/ul	99
68) 4-Bromophenyl-phenylether	16.628	248	307583	20.590	ng/ul	99
69) Hexachlorobenzene	16.739	284	367813	20.239	ng/ul	96
70) Atrazine	16.892	200	316640	21.023	ng/ul	96
71) Pentachlorophenol	17.081	266	234426	20.852	ng/ul	94
72) Phenanthrene	17.481	178	1763262	21.184	ng/ul	99
74) Anthracene	17.575	178	1814409	21.620	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.498	216	460453	21.026	ng/ul	99
76) Pentachlorobenzene	15.010	250	439763	20.609	ng/ul	96
77) Carbazole	17.839	167	1648229	21.930	ng/ul	98
78) Di-n-butylphthalate	18.410	149	1962593	21.938	ng/ul	99
80) Fluoranthene	19.492	202	2027671	21.133	ng/ul	100
82) Pyrene	19.857	202	2111166	20.855	ng/ul	98
83) Butylbenzylphthalate	20.751	149	554740	21.519	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.539	252	628133	23.078	ng/ul	88
85) Benzo(a)anthracene	21.610	228	2073933	21.069	ng/ul	97
86) Bis(2-ethylhexyl)phtha...	21.533	149	950881	22.226	ng/ul#	98
87) Chrysene	21.669	228	1985601	21.074	ng/ul	96
89) Di-n-octyl phthalate	22.504	149	1676093	19.364	ng/ul	100
90) Benzo(b)fluoranthene	23.380	252	2171466	21.311	ng/ul	97
91) Benzo(k)fluoranthene	23.427	252	2115980	20.996	ng/ul	98
93) Benzo(a)pyrene	24.045	252	1925471	21.338	ng/ul	97
94) Indeno(1,2,3-cd)pyrene	26.810	276	2550160	22.441	ng/ul	98
95) Dibenzo(a,h)anthracene	26.827	278	2129185	23.010	ng/ul	98
96) Benzo(g,h,i)perylene	27.621	276	2150169	22.933	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Quant Time: Mar 07 04:28:19 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN030323.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Mar 07 04:27:04 2023
Response via : Initial Calibration








#2
1,4-Dioxane
Concen: 7.696 ng/uL
RT: 3.446 min Scan# 61
Delta R.T. 0.000 min
Lab File: BN024127.D
Acq: 06 Mar 2023 09:27

Instrument :
BNA_N
ClientSampleId :
SSTD020257



Tgt Ion: 88 Resp: 58468




Ion	Ratio	Lower	Upper
88	100		
43	32.6	26.2	39.4
58	76.4	64.9	97.3




#5
Pyridine
Concen: 20.305 ng/uL
RT: 3.852 min Scan# 130
Delta R.T. 0.000 min
Lab File: BN024127.D
Acq: 06 Mar 2023 09:27

Tgt Ion: 79 Resp: 408072

Ion	Ratio	Lower	Upper
79	100		
52	88.5	65.1	97.7
51	39.8	31.4	47.0









#6
Benzaldehyde
Concen: 25.283 ng/ul
RT: 7.187 min Scan# 697
Delta R.T. 0.000 min
Lab File: BN024127.D
Acq: 06 Mar 2023 09:27




Instrument :
BNA_N
ClientSampleId :
SSTD020257


Tgt Ion: 77 Resp: 301561
Ion Ratio Lower Upper
77 100
105 97.2 79.2 118.8
106 94.4 75.9 113.9



#8
Phenol
Concen: 20.432 ng/ul
RT: 7.228 min Scan# 704
Delta R.T. 0.000 min
Lab File: BN024127.D
Acq: 06 Mar 2023 09:27

Tgt Ion: 94 Resp: 519970
Ion Ratio Lower Upper
94 100
65 28.9 22.9 34.3
66 36.1 29.0 43.6









#10
 Bis(2-Chloroethyl)ether
 Concen: 21.059 ng/ul
 RT: 7.475 min Scan# 746
 Delta R.T. 0.000 min
 Lab File: BN024127.D
 Acq: 06 Mar 2023 09:27

Instrument :
 BNA_N
 ClientSampleId :
 SST020257

Tgt Ion: 93 Resp: 434277




Ion	Ratio	Lower	Upper
93	100		
63	78.1	62.6	93.8
95	32.3	25.8	38.6




#12
 2-Chlorophenol
 Concen: 20.676 ng/ul
 RT: 7.616 min Scan# 770
 Delta R.T. 0.000 min
 Lab File: BN024127.D
 Acq: 06 Mar 2023 09:27

Tgt Ion: 128 Resp: 404705

Ion	Ratio	Lower	Upper
128	100		
64	48.8	39.3	58.9
130	32.2	25.1	37.7





#13

2-Methylphenol

Concen: 20.450 ng/ul

RT: 8.493 min Scan# 91

Delta R.T. 0.000 min

Lab File: BN024127.D

Acq: 06 Mar 2023 09:27

Instrument :

BNA_N

ClientSampleId :





SSTD020257

Tgt Ion:108 Resp: 391358

Ion Ratio Lower Upper

108 100

107 89.8 72.3 108.5



#14

2,2'-oxybis(1-Chloropropane)

Concen: 20.741 ng/ul

RT: 8.593 min Scan# 936

Delta R.T. 0.000 min

Lab File: BN024127.D

Acq: 06 Mar 2023 09:27




Tgt Ion: 45 Resp: 678597


Ion Ratio Lower Upper

45 100

77 12.5 9.8 14.6

79 9.8 7.4 11.0









#16
Acetophenone
Concen: 21.268 ng/ul
RT: 8.887 min Scan# 986
Delta R.T. 0.000 min
Lab File: BN024127.D
Acq: 06 Mar 2023 09:27




Instrument :
BNA_N
ClientSampleId :
SSTD020257


Tgt Ion:105 Resp: 644350
Ion Ratio Lower Upper
105 100
77 77.7 61.4 92.0
51 30.7 23.8 35.8



#17
N-Nitroso-di-n-propylamine
Concen: 20.887 ng/ul
RT: 8.875 min Scan# 984
Delta R.T. 0.000 min
Lab File: BN024127.D
Acq: 06 Mar 2023 09:27

Tgt Ion: 70 Resp: 310717
Ion Ratio Lower Upper
70 100
42 63.5 50.8 76.2
101 9.1 7.2 10.8
130 19.3 15.8 23.8









#18
4-Methylphenol
Concen: 20.490 ng/ul
RT: 8.828 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN024127.D
Acq: 06 Mar 2023 09:27




Instrument :
BNA_N
ClientSampleId :
SSTD020257


Tgt Ion:108 Resp: 431588
Ion Ratio Lower Upper
108 100
107 107.2 87.8 131.6



#19
Hexachloroethane
Concen: 21.029 ng/ul
RT: 9.152 min Scan# 1031
Delta R.T. 0.000 min
Lab File: BN024127.D
Acq: 06 Mar 2023 09:27

Tgt Ion:117 Resp: 161640
Ion Ratio Lower Upper
117 100
201 90.0 73.8 110.6
199 61.9 47.2 70.8









#22
Nitrobenzene
Concen: 21.996 ng/ul
RT: 9.269 min Scan# 1051
Delta R.T. 0.000 min
Lab File: BN024127.D
Acq: 06 Mar 2023 09:27




Instrument :
BNA_N
ClientSampleId :
SSTD020257


Tgt Ion: 77 Resp: 448429
Ion Ratio Lower Upper
77 100
123 45.4 34.8 52.2
65 13.4 11.0 16.4



#23
Isophorone
Concen: 21.237 ng/ul
RT: 9.799 min Scan# 1141
Delta R.T. 0.000 min
Lab File: BN024127.D
Acq: 06 Mar 2023 09:27

Tgt Ion: 82 Resp: 946279
Ion Ratio Lower Upper
82 100
95 6.4 5.0 7.6
138 16.6 13.8 20.6





#25

2-Nitrophenol

Concen: 22.363 ng/ul

RT: 9.987 min Scan# 1173

Delta R.T. 0.000 min

Lab File: BN024127.D

Acq: 06 Mar 2023 09:27

Instrument :

BNA_N

ClientSampleId :

SSTD020257





Tgt Ion:139 Resp: 210966

Ion Ratio Lower Upper

139 100

65 51.5 44.5 66.7

109 28.3 23.4 35.2



#26

2,4-Dimethylphenol

Concen: 21.357 ng/ul

RT: 10.040 min Scan# 1182

Delta R.T. 0.000 min

Lab File: BN024127.D

Acq: 06 Mar 2023 09:27




Tgt Ion:107 Resp: 459189


Ion Ratio Lower Upper

107 100

121 52.0 40.2 60.2

122 88.2 71.8 107.8









#27
 Bis(2-Chloroethoxy)methane
 Concen: 21.172 ng/ul
 RT: 10.281 min Scan# 1223
 Delta R.T. 0.000 min
 Lab File: BN024127.D
 Acq: 06 Mar 2023 09:27




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020257


Tgt Ion: 93 Resp: 585767
 Ion Ratio Lower Upper
 93 100
 95 32.9 25.4 38.2
 123 11.5 8.6 12.8



#29
 2,4-Dichlorophenol
 Concen: 20.997 ng/ul
 RT: 10.516 min Scan# 1263
 Delta R.T. 0.000 min
 Lab File: BN024127.D
 Acq: 06 Mar 2023 09:27

Tgt Ion: 162 Resp: 375541
 Ion Ratio Lower Upper
 162 100
 164 64.8 52.6 78.8
 98 37.8 29.8 44.8









#30
Naphthalene
Concen: 21.005 ng/ul
RT: 10.928 min Scan# 1333
Delta R.T. 0.000 min
Lab File: BN024127.D
Acq: 06 Mar 2023 09:27




Instrument :
BNA_N
ClientSampleId :
SSTD020257


Tgt Ion:128 Resp: 1341160
Ion Ratio Lower Upper
128 100
129 11.6 8.6 12.8
127 12.8 10.1 15.1



#32
4-Chloroaniline
Concen: 21.649 ng/ul
RT: 11.034 min Scan# 1351
Delta R.T. 0.000 min
Lab File: BN024127.D
Acq: 06 Mar 2023 09:27

Tgt Ion:127 Resp: 577179
Ion Ratio Lower Upper
127 100
129 32.0 25.1 37.7





#33

Hexachlorobutadiene

Concen: 20.553 ng/ul

RT: 11.216 min Scan# 1382

Delta R.T. 0.000 min

Lab File: BN024127.D

Acq: 06 Mar 2023 09:27

Instrument :

BNA_N

ClientSampleId :

SSTD020257





Tgt Ion:225 Resp: 224030

Ion Ratio Lower Upper

225 100

223 61.3 53.6 80.4

227 65.7 53.0 79.4



#34

Caprolactam

Concen: 10.667 ng/ul

RT: 11.793 min Scan# 1480

Delta R.T. 0.000 min

Lab File: BN024127.D

Acq: 06 Mar 2023 09:27




Tgt Ion:113 Resp: 65060


Ion Ratio Lower Upper

113 100

55 188.8 155.4 233.2

56 145.0 111.0 166.6









#35
4-Chloro-3-methylphenol
Concen: 21.373 ng/ul
RT: 12.146 min Scan# 11
Delta R.T. 0.000 min
Lab File: BN024127.D
Acq: 06 Mar 2023 09:27




Instrument :
BNA_N
ClientSampleId :
SSTD020257


Tgt Ion:107 Resp: 421122
Ion Ratio Lower Upper
107 100
144 25.8 21.0 31.4
142 79.1 65.7 98.5



#36
2-Methylnaphthalene
Concen: 21.235 ng/ul
RT: 12.534 min Scan# 1606
Delta R.T. 0.000 min
Lab File: BN024127.D
Acq: 06 Mar 2023 09:27

Tgt Ion:142 Resp: 892696
Ion Ratio Lower Upper
142 100
141 87.0 68.2 102.4





#37

1-Methylnaphthalene

Concen: 21.247 ng/ul

RT: 12.751 min Scan# 1643

Delta R.T. 0.000 min

Lab File: BN024127.D

Acq: 06 Mar 2023 09:27

Instrument :

BNA_N

ClientSampleId :

SSTD020257





Tgt Ion:142 Resp: 895139

Ion Ratio Lower Upper

142 100

141 93.0 75.5 113.3

116 3.7 3.1 4.7



#39

1,2,4,5-Tetrachlorobenzene

Concen: 20.232 ng/ul

RT: 12.893 min Scan# 1667

Delta R.T. 0.000 min

Lab File: BN024127.D

Acq: 06 Mar 2023 09:27

Tgt Ion:216 Resp: 437099




Ion Ratio Lower Upper


216 100

214 80.6 60.0 90.0

179 20.9 17.0 25.6

108 19.0 14.3 21.5





#40

Hexachlorocyclopentadiene

Concen: 18.913 ng/ul

RT: 12.881 min Scan# 1665

Delta R.T. 0.000 min

Lab File: BN024127.D

Acq: 06 Mar 2023 09:27

Instrument :

BNA_N

ClientSampleId :

SSTD020257





Tgt Ion:237 Resp: 220216

Ion Ratio Lower Upper

237 100

235 66.0 50.1 75.1

272 12.9 11.1 16.7



#41

2,4,6-Trichlorophenol

Concen: 20.514 ng/ul

RT: 13.128 min Scan# 1707

Delta R.T. 0.000 min

Lab File: BN024127.D

Acq: 06 Mar 2023 09:27




Tgt Ion:196 Resp: 290336


Ion Ratio Lower Upper

196 100

198 100.1 80.6 121.0

200 33.8 26.1 39.1





#42

2,4,5-Trichlorophenol

Concen: 20.764 ng/ul

RT: 13.198 min Scan# 1719

Delta R.T. 0.000 min

Lab File: BN024127.D

Acq: 06 Mar 2023 09:27

Instrument :

BNA_N

ClientSampleId :

SSTD020257





Tgt Ion:196 Resp: 316504

Ion Ratio Lower Upper

196 100

198 97.3 79.7 119.5

200 31.9 26.1 39.1



#43

1,1'-Biphenyl

Concen: 20.843 ng/ul

RT: 13.534 min Scan# 1776

Delta R.T. 0.000 min

Lab File: BN024127.D

Acq: 06 Mar 2023 09:27




Tgt Ion:154 Resp: 1192010


Ion Ratio Lower Upper

154 100

153 40.3 30.8 46.2

76 13.8 11.1 16.7









#44
2-Chloronaphthalene
Concen: 20.692 ng/ul
RT: 13.575 min Scan# 11
Delta R.T. 0.000 min
Lab File: BN024127.D
Acq: 06 Mar 2023 09:27




Instrument :
BNA_N
ClientSampleId :
SSTD020257


Tgt Ion:162 Resp: 922010
Ion Ratio Lower Upper
162 100
164 31.2 25.8 38.6
127 39.7 30.8 46.2



#45
2-Nitroaniline
Concen: 21.527 ng/ul
RT: 13.775 min Scan# 1817
Delta R.T. 0.000 min
Lab File: BN024127.D
Acq: 06 Mar 2023 09:27

Tgt Ion: 65 Resp: 208985
Ion Ratio Lower Upper
65 100
92 67.8 53.8 80.8
138 105.8 84.1 126.1









#47
Dimethylphthalate
Concen: 20.733 ng/ul
RT: 14.151 min Scan# 1881
Delta R.T. 0.000 min
Lab File: BN024127.D
Acq: 06 Mar 2023 09:27




Instrument :
BNA_N
ClientSampleId :
SSTD020257


Tgt Ion:163 Resp: 1157897
Ion Ratio Lower Upper
163 100
194 3.7 3.1 4.7
164 9.9 8.1 12.1



#48
2,6-Dinitrotoluene
Concen: 21.722 ng/ul
RT: 14.269 min Scan# 1901
Delta R.T. 0.000 min
Lab File: BN024127.D
Acq: 06 Mar 2023 09:27

Tgt Ion:165 Resp: 178567
Ion Ratio Lower Upper
165 100
89 60.1 52.6 78.8
121 22.2 19.4 29.2





#50

Acenaphthylene

Concen: 21.230 ng/ul

RT: 14.416 min Scan# 1927

Delta R.T. 0.000 min

Lab File: BN024127.D

Acq: 06 Mar 2023 09:27

Instrument :

BNA_N

ClientSampleId :

SSTD020257





Tgt Ion:152 Resp: 1475528

Ion Ratio Lower Upper

152 100

151 19.3 15.6 23.4

153 13.2 10.1 15.1



#51

3-Nitroaniline

Concen: 21.137 ng/ul

RT: 14.593 min Scan# 1956

Delta R.T. 0.000 min

Lab File: BN024127.D

Acq: 06 Mar 2023 09:27




Tgt Ion:138 Resp: 211206


Ion Ratio Lower Upper

138 100

108 11.8 9.6 14.4

92 117.9 95.3 142.9









#52
Acenaphthene
Concen: 21.031 ng/ul
RT: 14.757 min Scan# 1984
Delta R.T. 0.000 min
Lab File: BN024127.D
Acq: 06 Mar 2023 09:27




Instrument :
BNA_N
ClientSampleId :
SSTD020257


Tgt Ion:153 Resp: 997069
Ion Ratio Lower Upper
153 100
152 46.7 39.7 59.5
154 89.1 70.6 106.0



#53
2,4-Dinitrophenol
Concen: 16.754 ng/ul
RT: 14.793 min Scan# 1990
Delta R.T. 0.000 min
Lab File: BN024127.D
Acq: 06 Mar 2023 09:27

Tgt Ion:184 Resp: 95351
Ion Ratio Lower Upper
184 100
63 76.4 61.6 92.4
154 64.3 50.2 75.4









#55
4-Nitrophenol
Concen: 20.591 ng/ul
RT: 14.887 min Scan# 2006
Delta R.T. 0.000 min
Lab File: BN024127.D
Acq: 06 Mar 2023 09:27




Instrument :
BNA_N
ClientSampleId :
SSTD020257


Tgt Ion:109 Resp: 166076
Ion Ratio Lower Upper
109 100
139 138.5 115.0 172.4
65 139.5 115.7 173.5



#56
Dibenzofuran
Concen: 20.465 ng/ul
RT: 15.092 min Scan# 2041
Delta R.T. 0.000 min
Lab File: BN024127.D
Acq: 06 Mar 2023 09:27

Tgt Ion:168 Resp: 1363053
Ion Ratio Lower Upper
168 100
139 38.8 30.7 46.1









#57
2,4-Dinitrotoluene
Concen: 21.824 ng/ul
RT: 15.045 min Scan# 2034
Delta R.T. 0.000 min
Lab File: BN024127.D
Acq: 06 Mar 2023 09:27




Instrument :
BNA_N
ClientSampleId :
SSTD020257


Tgt Ion:165 Resp: 282794
Ion Ratio Lower Upper
165 100
63 48.5 38.5 57.7
182 2.5 2.2 3.4



#58
2,3,4,6-Tetrachlorophenol
Concen: 20.561 ng/ul
RT: 15.310 min Scan# 2078
Delta R.T. 0.000 min
Lab File: BN024127.D
Acq: 06 Mar 2023 09:27

Tgt Ion:232 Resp: 269410
Ion Ratio Lower Upper
232 100
131 50.6 38.5 57.7
130 2.6 2.0 3.0
166 34.0 26.0 39.0









#59
Diethylphthalate
Concen: 20.655 ng/ul
RT: 15.510 min Scan# 2112
Delta R.T. 0.000 min
Lab File: BN024127.D
Acq: 06 Mar 2023 09:27




Instrument :
BNA_N
ClientSampleId :
SSTD020257


Tgt Ion:149 Resp: 1149829
Ion Ratio Lower Upper
149 100
177 21.9 16.9 25.3
150 12.1 9.7 14.5



#61
Fluorene
Concen: 21.215 ng/ul
RT: 15.740 min Scan# 2151
Delta R.T. 0.000 min
Lab File: BN024127.D
Acq: 06 Mar 2023 09:27

Tgt Ion:166 Resp: 1106992
Ion Ratio Lower Upper
166 100
165 90.6 75.6 113.4
167 12.9 11.2 16.8









#62
4-Chlorophenyl-phenylether
Concen: 20.782 ng/ul
RT: 15.734 min Scan# 2150
Delta R.T. 0.000 min
Lab File: BN024127.D
Acq: 06 Mar 2023 09:27




Instrument :
BNA_N
ClientSampleId :
SSTD020257


Tgt Ion:204 Resp: 518499
Ion Ratio Lower Upper
204 100
206 33.2 27.1 40.7
141 62.3 47.4 71.0



#63
4-Nitroaniline
Concen: 21.910 ng/ul
RT: 15.751 min Scan# 2153
Delta R.T. 0.000 min
Lab File: BN024127.D
Acq: 06 Mar 2023 09:27

Tgt Ion:138 Resp: 211718
Ion Ratio Lower Upper
138 100
92 52.6 40.0 60.0
108 75.9 60.4 90.6









#66
 4,6-Dinitro-2-methylphenol
 Concen: 19.465 ng/ul
 RT: 15.810 min Scan# 2163
 Delta R.T. 0.000 min
 Lab File: BN024127.D
 Acq: 06 Mar 2023 09:27




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020257


Tgt Ion:198 Resp: 159397
 Ion Ratio Lower Upper
 198 100
 182 4.3 3.4 5.0
 77 26.3 20.4 30.6



#67
 N-Nitrosodiphenylamine
 Concen: 21.215 ng/ul
 RT: 15.945 min Scan# 2186
 Delta R.T. 0.000 min
 Lab File: BN024127.D
 Acq: 06 Mar 2023 09:27

Tgt Ion:169 Resp: 943948
 Ion Ratio Lower Upper
 169 100
 168 67.9 54.5 81.7
 167 36.4 28.3 42.5









#68
4-Bromophenyl-phenylether
Concen: 20.590 ng/ul
RT: 16.628 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN024127.D
Acq: 06 Mar 2023 09:27




Instrument :
BNA_N
ClientSampleId :
SSTD020257


Tgt Ion:248 Resp: 307583
Ion Ratio Lower Upper
248 100
250 96.2 76.6 115.0
141 77.4 61.4 92.0



#69
Hexachlorobenzene
Concen: 20.239 ng/ul
RT: 16.739 min Scan# 2321
Delta R.T. 0.000 min
Lab File: BN024127.D
Acq: 06 Mar 2023 09:27

Tgt Ion:284 Resp: 367813
Ion Ratio Lower Upper
284 100
142 38.7 29.0 43.6
249 28.9 25.0 37.6









#70
Atrazine
Concen: 21.023 ng/ul
RT: 16.892 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN024127.D
Acq: 06 Mar 2023 09:27




Instrument :
BNA_N
ClientSampleId :
SSTD020257


Tgt Ion:200 Resp: 316640
Ion Ratio Lower Upper
200 100
173 26.7 20.7 31.1
215 45.4 38.8 58.2



#71
Pentachlorophenol
Concen: 20.852 ng/ul
RT: 17.081 min Scan# 2379
Delta R.T. 0.000 min
Lab File: BN024127.D
Acq: 06 Mar 2023 09:27

Tgt Ion:266 Resp: 234426
Ion Ratio Lower Upper
266 100
264 60.4 49.5 74.3
268 55.7 51.3 76.9









#72
Phenanthrene
Concen: 21.184 ng/ul
RT: 17.481 min Scan# 2447
Delta R.T. 0.000 min
Lab File: BN024127.D
Acq: 06 Mar 2023 09:27




Instrument :
BNA_N
ClientSampleId :
SSTD020257


Tgt Ion:178 Resp: 1763262
Ion Ratio Lower Upper
178 100
179 14.9 11.6 17.4
176 19.2 14.9 22.3



#74
Anthracene
Concen: 21.620 ng/ul
RT: 17.575 min Scan# 2463
Delta R.T. 0.000 min
Lab File: BN024127.D
Acq: 06 Mar 2023 09:27

Tgt Ion:178 Resp: 1814409
Ion Ratio Lower Upper
178 100
179 15.0 11.9 17.9
176 18.9 14.6 21.8









#75
1,2,3,4-Tetrachlorobenzene
Concen: 21.026 ng/uL
RT: 13.498 min Scan# 1770
Delta R.T. 0.000 min
Lab File: BN024127.D
Acq: 06 Mar 2023 09:27




Instrument :
BNA_N
ClientSampleId :
SSTD020257


Tgt Ion:216 Resp: 460453
Ion Ratio Lower Upper
216 100
214 73.8 59.6 89.4
218 47.0 37.0 55.4



#76
Pentachlorobenzene
Concen: 20.609 ng/uL
RT: 15.010 min Scan# 2027
Delta R.T. 0.000 min
Lab File: BN024127.D
Acq: 06 Mar 2023 09:27

Tgt Ion:250 Resp: 439763
Ion Ratio Lower Upper
250 100
248 62.3 53.0 79.4
252 64.8 54.6 81.8









#77
 Carbazole
 Concen: 21.930 ng/ul
 RT: 17.839 min Scan# 2508
 Delta R.T. 0.000 min
 Lab File: BN024127.D
 Acq: 06 Mar 2023 09:27

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020257

Tgt Ion:167 Resp: 1648229




Ion	Ratio	Lower	Upper
167	100		
166	20.2	17.3	25.9
139	12.9	10.5	15.7




#78
 Di-n-butylphthalate
 Concen: 21.938 ng/ul
 RT: 18.410 min Scan# 2605
 Delta R.T. 0.000 min
 Lab File: BN024127.D
 Acq: 06 Mar 2023 09:27

Tgt Ion:149 Resp: 1962593

Ion	Ratio	Lower	Upper
149	100		
150	8.9	7.4	11.0
104	5.0	4.0	6.0









#80
 Fluoranthene
 Concen: 21.133 ng/ul
 RT: 19.492 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: BN024127.D
 Acq: 06 Mar 2023 09:27




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020257


Tgt Ion:202 Resp: 2027671
 Ion Ratio Lower Upper
 202 100
 101 12.3 10.0 15.0
 100 9.3 7.6 11.4



#82
 Pyrene
 Concen: 20.855 ng/ul
 RT: 19.857 min Scan# 2851
 Delta R.T. 0.000 min
 Lab File: BN024127.D
 Acq: 06 Mar 2023 09:27

Tgt Ion:202 Resp: 2111166
 Ion Ratio Lower Upper
 202 100
 101 14.3 11.9 17.9
 100 11.4 9.9 14.9









#83
Butylbenzylphthalate
Concen: 21.519 ng/ul
RT: 20.751 min Scan# 3003
Delta R.T. 0.000 min
Lab File: BN024127.D
Acq: 06 Mar 2023 09:27




Instrument :
BNA_N
ClientSampleId :
SSTD020257


Tgt Ion:149 Resp: 554740
Ion Ratio Lower Upper
149 100
91 74.2 62.2 93.2
206 20.2 16.4 24.6



#84
3,3'-Dichlorobenzidine
Concen: 23.078 ng/ul
RT: 21.539 min Scan# 3137
Delta R.T. 0.000 min
Lab File: BN024127.D
Acq: 06 Mar 2023 09:27

Tgt Ion:252 Resp: 628133
Ion Ratio Lower Upper
252 100
254 56.6 54.6 81.8
126 12.3 11.0 16.6





#85

Benzo(a)anthracene

Concen: 21.069 ng/ul

RT: 21.610 min Scan# 3150

Delta R.T. 0.000 min

Lab File: BN024127.D

Acq: 06 Mar 2023 09:27

Instrument :

BNA_N

ClientSampleId :

SSTD020257





Tgt Ion:228 Resp: 2073933

Ion Ratio Lower Upper

228 100

229 19.1 16.3 24.5

226 26.3 22.6 34.0



#86

Bis(2-ethylhexyl)phthalate

Concen: 22.226 ng/ul

RT: 21.533 min Scan# 3136

Delta R.T. 0.000 min

Lab File: BN024127.D

Acq: 06 Mar 2023 09:27




Tgt Ion:149 Resp: 950881


Ion Ratio Lower Upper

149 100

167 27.8 21.6 32.4

279 3.4 3.4 5.2#





#87

Chrysene

Concen: 21.074 ng/ul

RT: 21.669 min Scan# 3160

Delta R.T. 0.000 min

Lab File: BN024127.D

Acq: 06 Mar 2023 09:27

Instrument :

BNA_N

ClientSampleId :

SSTD020257





Tgt Ion:228 Resp: 1985601

Ion Ratio Lower Upper

228 100

226 29.0 25.8 38.6

229 20.8 17.5 26.3



#89

Di-n-octyl phthalate

Concen: 19.364 ng/ul




RT: 22.504 min Scan# 3301


Delta R.T. 0.000 min

Lab File: BN024127.D

Acq: 06 Mar 2023 09:27

Tgt Ion:149 Resp: 1676093









#90
Benzo(b)fluoranthene
Concen: 21.311 ng/ul
RT: 23.380 min Scan# 3451
Delta R.T. 0.000 min
Lab File: BN024127.D
Acq: 06 Mar 2023 09:27




Instrument :
BNA_N
ClientSampleId :
SSTD020257


Tgt Ion:252 Resp: 2171466
Ion Ratio Lower Upper
252 100
253 20.8 18.3 27.5
125 11.7 9.5 14.3



#91
Benzo(k)fluoranthene
Concen: 20.996 ng/ul
RT: 23.427 min Scan# 3458
Delta R.T. 0.000 min
Lab File: BN024127.D
Acq: 06 Mar 2023 09:27

Tgt Ion:252 Resp: 2115980
Ion Ratio Lower Upper
252 100
253 23.4 17.4 26.0
125 11.1 9.0 13.6





#93

Benzo(a)pyrene

Concen: 21.338 ng/ul

RT: 24.045 min Scan# 3563

Delta R.T. 0.000 min

Lab File: BN024127.D

Acq: 06 Mar 2023 09:27

Instrument :

BNA_N

ClientSampleId :

SSTD020257





Tgt Ion:252 Resp: 1925471

Ion Ratio Lower Upper

252 100

253 22.7 17.0 25.4

125 13.0 9.5 14.3



#94

Indeno(1,2,3-cd)pyrene

Concen: 22.441 ng/ul

RT: 26.810 min Scan# 4033

Delta R.T. 0.000 min

Lab File: BN024127.D

Acq: 06 Mar 2023 09:27




Tgt Ion:276 Resp: 2550160


Ion Ratio Lower Upper

276 100

138 25.9 19.6 29.4

277 23.9 19.7 29.5





#95

Dibenzo(a,h)anthracene

Concen: 23.010 ng/ul

RT: 26.827 min Scan# 4038

Delta R.T. 0.000 min

Lab File: BN024127.D

Acq: 06 Mar 2023 09:27

Instrument :

BNA_N

ClientSampleId :

SSTD020257





Tgt Ion:278 Resp: 2129185

Ion Ratio Lower Upper

278 100

139 17.6 14.4 21.6

279 24.0 18.2 27.2



#96

Benzo(g,h,i)perylene

Concen: 22.933 ng/ul

RT: 27.621 min Scan# 4171

Delta R.T. 0.000 min

Lab File: BN024127.D

Acq: 06 Mar 2023 09:27

Tgt Ion:276 Resp: 2150169

Ion Ratio Lower Upper

276 100

138 23.8 18.8 28.2

277 22.5 18.6 27.8

