

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030623\
 Data File : BN024133.D
 Acq On : 06 Mar 2023 13:37
 Operator : CG/JU
 Sample : PB150960BL
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK960

Quant Time: Mar 07 04:31:08 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN030323.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Mar 07 04:27:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc Units	Dev(Min)

Internal Standards					
1) 1,4-Dichlorobenzene-d4	8.058	152	223716	20.000 ng/u1	0.00
20) Naphthalene-d8	10.881	136	978380	20.000 ng/u1	0.00
38) Acenaphthene-d10	14.693	164	605123	20.000 ng/u1	0.00
64) Phenanthrene-d10	17.439	188	1262107	20.000 ng/u1	0.00
79) Chrysene-d12	21.633	240	1122282	20.000 ng/u1	0.00
88) Perylene-d12	24.163	264	1286642	20.000 ng/u1	0.01

System Monitoring Compounds					
3) 1,4-Dioxane-d8	3.405	96	35449	6.181 ng/uL	0.00
4) Pyridine-d5	3.828	84	493920	31.125 ng/u1	0.00
7) Phenol-d5	7.199	99	623020	31.047 ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.381	67	412580	32.106 ng/u1	0.00
11) 2-Chlorophenol-d4	7.581	132	487508	31.634 ng/u1	0.00
15) 4-Methylphenol-d8	8.758	113	482295	30.233 ng/u1	0.00
21) Nitrobenzene-d5	9.228	128	204005	32.126 ng/u1	0.00
24) 2-Nitrophenol-d4	9.952	143	228704	33.027 ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.487	165	445274	29.799 ng/u1	0.00
31) 4-Chloroaniline-d4	11.010	131	684948	30.724 ng/u1	0.00
46) Dimethylphthalate-d6	14.104	166	1523505	32.622 ng/u1	0.00
49) Acenaphthylene-d8	14.393	160	1704761	31.360 ng/u1	0.00
54) 4-Nitrophenol-d4	14.869	143	209244	23.448 ng/u1	0.00
60) Fluorene-d10	15.687	176	1258753	32.071 ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.793	200	145859	21.450 ng/u1	0.00
73) Anthracene-d10	17.540	188	1889171	31.738 ng/u1	0.00
81) Pyrene-d10	19.828	212	2268706	34.589 ng/u1	0.00
92) Benzo(a)pyrene-d12	23.998	264	2256724	35.283 ng/u1	0.01

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030623\
Data File : BN024133.D
Acq On : 06 Mar 2023 13:37
Operator : CG/JU
Sample : PB150960BL
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK960

Quant Time: Mar 07 04:31:08 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN030323.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Mar 07 04:27:04 2023
Response via : Initial Calibration

