

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030624\
 Data File : BN030308.D
 Acq On : 07 Mar 2024 02:12
 Operator : MA/JU
 Sample : P1590-11
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BPOW1-7-20240223

Quant Time: Mar 07 03:08:03 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN030524.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 05 17:44:16 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.811	152	1619	0.400	ng	0.00
7) Naphthalene-d8	10.605	136	4143	0.400	ng	0.00
13) Acenaphthene-d10	14.457	164	2232	0.400	ng	0.00
19) Phenanthrene-d10	17.230	188	4483	0.400	ng	0.01
29) Chrysene-d12	21.438	240	3909	0.400	ng	0.00
35) Perylene-d12	23.797	264	4162	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.384	112	430	0.101	ng	0.00
5) Phenol-d6	6.988	99	237	0.053	ng	0.00
8) Nitrobenzene-d5	8.982	82	1149	0.304	ng	0.01
11) 2-Methylnaphthalene-d10	12.208	152	2014	0.335	ng	0.00
14) 2,4,6-Tribromophenol	15.977	330	349	0.304	ng	0.01
15) 2-Fluorobiphenyl	13.089	172	3018	0.326	ng	0.00
27) Fluoranthene-d10	19.261	212	4902	0.398	ng	0.00
31) Terphenyl-d14	19.875	244	4005	0.395	ng	0.00
Target Compounds						
34) Bis(2-ethylhexyl)phtha...	21.366	149	384	0.035	ng	Qvalue # 85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030624\
Data File : BN030308.D
Acq On : 07 Mar 2024 02:12
Operator : MA/JU
Sample : P1590-11
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BPOW1-7-20240223

Quant Time: Mar 07 03:08:03 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN030524.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 05 17:44:16 2024
Response via : Initial Calibration

