




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030624\
Data File : BN030286.D
Acq On : 06 Mar 2024 09:56
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Mar 07 04:38:50 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 07 04:37:16 2024
Response via : Initial Calibration






TIC: BN030286.D\data.ms

(70) Pentachlorophenol (C)**16.863min (+ 0.006) 16107.74 ng****response 119457**

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	63.47
264.00	61.60	61.58
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030624\
Data File : BN030286.D
Acq On : 06 Mar 2024 09:56
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Mar 07 04:38:50 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 07 04:37:16 2024
Response via : Initial Calibration



TIC: BN030286.D\data.ms

(77) Benzidine

19.492min (-0.000) 0.00 ng

response 706591

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	13.15
183.00	13.20	10.50#
0.00	0.00	0.00