

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030624\
 Data File : BN030321.D
 Acq On : 07 Mar 2024 10:02
 Operator : MA/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 07 11:11:11 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN030524.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 05 17:44:16 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	139	0.00
2	1,4-Dioxane	0.400	0.359	10.3	117	0.00
3	n-Nitrosodimethylamine	0.400	0.327	18.3	115	0.00
4 S	2-Fluorophenol	0.400	0.378	5.5	131	0.00
5 S	Phenol-d6	0.400	0.382	4.5	134	0.00
6	bis(2-Chloroethyl)ether	0.400	0.379	5.3	133	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	147	-0.01
8 S	Nitrobenzene-d5	0.400	0.367	8.3	133	0.00
9	Naphthalene	0.400	0.414	-3.5	146	-0.01
10	Hexachlorobutadiene	0.400	0.442	-10.5	154	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.429	-7.2	151	0.00
12	2-Methylnaphthalene	0.400	0.417	-4.2	147	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	144	0.00
14 S	2,4,6-Tribromophenol	0.400	0.355	11.3	122	0.00
15 S	2-Fluorobiphenyl	0.400	0.426	-6.5	149	-0.01
16	Acenaphthylene	0.400	0.408	-2.0	145	0.00
17	Acenaphthene	0.400	0.418	-4.5	146	0.00
18	Fluorene	0.400	0.412	-3.0	144	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	137	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.308	23.0	114	0.00
21	4-Bromophenyl-phenylether	0.400	0.403	-0.8	133	0.00
22	Hexachlorobenzene	0.400	0.431	-7.7	144	0.00
23	Atrazine	0.400	0.375	6.3	122	0.00
24	Pentachlorophenol	0.400	0.340	15.0	111	0.00
25	Phenanthrene	0.400	0.409	-2.2	137	0.00
26	Anthracene	0.400	0.398	0.5	134	0.00
27 SURR	Fluoranthene-d10	0.400	0.393	1.8	131	0.00
28	Fluoranthene	0.400	0.389	2.8	129	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	127	0.00
30	Pyrene	0.400	0.424	-6.0	128	0.00
31 S	Terphenyl-d14	0.400	0.399	0.3	122	0.00
32	Benzo(a)anthracene	0.400	0.403	-0.8	127	0.00
33	Chrysene	0.400	0.413	-3.2	126	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.359	10.3	108	0.00
35 I	Perylene-d12	0.400	0.400	0.0	126	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.401	-0.3	126	0.00
37	Benzo(b)fluoranthene	0.400	0.404	-1.0	125	0.00
38	Benzo(k)fluoranthene	0.400	0.411	-2.7	126	0.00
39 C	Benzo(a)pyrene	0.400	0.404	-1.0	126	0.00
40	Dibenzo(a,h)anthracene	0.400	0.394	1.5	129	0.00
41	Benzo(g,h,i)perylene	0.400	0.419	-4.7	130	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4