

DDT Breakdown

Date	Instrument Name	DFTPP Data File
3/7/2024	BNA_N	<u>BN030322.D</u>
Compound Name	Response	Retention Time
DDT	379758	20.733
DDD	8531	20.292
DDE	294	19.774
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
8825	388583	2.27