



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030724\
Data File : BN030341.D
Acq On : 08 Mar 2024 01:44
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 60 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN030524.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Tue Mar 05 17:44:16 2024






AutoFind: Scans 2478, 2479, 2480; Background Corrected with Scan 2471

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	41.3	58845	PASS
68	69	0.00	2	1.9	1317	PASS
69	198	0.00	100	49.4	70379	PASS
70	69	0.00	2	0.4	306	PASS
127	198	10	80	56.9	81024	PASS
197	198	0.00	2	1.1	1590	PASS
198	198	100	100	100.0	142464	PASS
199	198	5	9	6.8	9633	PASS
275	198	10	60	26.1	37181	PASS
365	198	1	100	4.4	6311	PASS
441	198	0.01	100	10.0	14239	PASS
442	442	50	100	100.0	91136	PASS
443	442	15	24	19.5	17778	PASS

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Mar 08 04:34:43 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 08 04:34:38 2024
Response via : Initial Calibration



TIC: BN030341.D\data.ms

(70) Pentachlorophenol (C)

16.857min (0.000) 21020.12 ng




response **163311**

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	61.84
264.00	61.60	63.47
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030724\
Data File : BN030341.D
Acq On : 08 Mar 2024 01:44
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 60 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Mar 08 04:34:43 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 08 04:34:38 2024
Response via : Initial Calibration



TIC: BN030341.D\data.ms

(77) Benzidine

19.486min (0.000) 0.00 ng

response 830486

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	13.84
183.00	13.20	11.86
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
3/7/2024	BNA_N	<u>BN030341.D</u>
Compound Name	Response	Retention Time
DDT	484525	20.727
DDD	11856	20.286
DDE	398	19.775
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
12254	496779	2.47

Instrument :
BNA_N
ClientSampleId :
DFTPP