

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030725\
 Data File : BN036545.D
 Acq On : 07 Mar 2025 13:48
 Operator : RC/JU
 Sample : PB167026BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB167026BL

Quant Time: Mar 07 14:29:31 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 05 16:08:57 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.725	152	2367	0.400	ng	0.00
7) Naphthalene-d8	10.530	136	5026	0.400	ng	0.02
13) Acenaphthene-d10	14.377	164	2916	0.400	ng	0.01
19) Phenanthrene-d10	17.124	188	5638	0.400	ng	# 0.01
29) Chrysene-d12	21.304	240	4494	0.400	ng	# 0.00
35) Perylene-d12	23.563	264	4150	0.400	ng	# 0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.320	112	2250	0.402	ng	0.00
5) Phenol-d6	6.908	99	2206	0.336	ng	0.00
8) Nitrobenzene-d5	8.886	82	1937	0.391	ng	0.01
11) 2-Methylnaphthalene-d10	12.131	152	2570	0.333	ng	0.03
14) 2,4,6-Tribromophenol	15.883	330	303	0.210	ng	0.02
15) 2-Fluorobiphenyl	13.009	172	4001	0.365	ng	0.02
27) Fluoranthene-d10	19.150	212	6316	0.403	ng	0.00
31) Terphenyl-d14	19.754	244	4265	0.445	ng	0.00

Target Compounds	Qvalue
-----	-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030725\
Data File : BN036545.D
Acq On : 07 Mar 2025 13:48
Operator : RC/JU
Sample : PB167026BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB167026BL

Quant Time: Mar 07 14:29:31 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 05 16:08:57 2025
Response via : Initial Calibration

