

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN030920\
 Data File : BN010143.D
 Acq On : 09 Mar 2020 11:27
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 09 17:06:36 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN021920.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 19 15:57:43 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	-0.05
2	1,4-Dioxane	0.410	0.435	-6.1	103	-0.05
3	n-Nitrosodimethylamine	0.673	0.617	8.3	100	-0.02
4 S	2-Fluorophenol	0.889	0.887	0.2	100	-0.04
5 S	Phenol-d6	1.082	1.004	7.2	94	-0.03
6	bis(2-Chloroethyl)ether	1.052	1.070	-1.7	102	-0.05
7 I	Naphthalene-d8	1.000	1.000	0.0	107	-0.05
8 S	Nitrobenzene-d5	0.333	0.279	16.2	96	-0.03
9	Naphthalene	1.091	1.059	2.9	107	-0.05
10	Hexachlorobutadiene	0.240	0.241	-0.4	108	-0.06
11 SURR	2-Methylnaphthalene-d10	0.774	0.775	-0.1	111	-0.04
12	2-Methylnaphthalene	0.748	0.740	1.1	110	-0.05
13 I	Acenaphthene-d10	1.000	1.000	0.0	115	-0.05
14 S	2,4,6-Tribromophenol	0.154	0.154	0.0	128	-0.04
15 S	2-Fluorobiphenyl	1.510	1.390	7.9	110	-0.05
16	Acenaphthylene	1.705	1.628	4.5	122	-0.05
17	Acenaphthene	1.199	1.137	5.2	113	-0.05
18	Fluorene	1.582	1.470	7.1	110	-0.04
19 I	Phenanthrene-d10	1.000	1.000	0.0	110	-0.04
20	4-Bromophenyl-phenylether	0.279	0.098	64.9#	41#	0.00
21	Hexachlorobenzene	0.249	0.246	1.2	114	-0.05
22	Pentachlorophenol	0.064	0.050	21.9	115	-0.04
23	Phenanthrene	1.190	1.074	9.7	107	-0.04
24	Anthracene	1.006	0.903	10.2	110	-0.04
25 SURR	Fluoranthene-d10	1.384	1.249	9.8	107	-0.04
26	Fluoranthene	1.425	1.281	10.1	107	-0.04
27 I	Chrysene-d12	1.000	1.000	0.0	110	-0.04
28	Pvrene	1.519	1.435	5.5	107	-0.04
29 S	Terphenyl-d14	0.956	0.880	7.9	104	-0.04
30	Benzo(a)anthracene	1.327	1.226	7.6	115	-0.04
31	Chrysene	1.536	1.353	11.9	100	-0.03
32	Bis(2-ethylhexyl)phthalate	0.648	0.579	10.6	110	-0.04
33	Indeno(1,2,3-cd)pyrene	1.599	1.612	-0.8	111	-0.07
34 I	Perylene-d12	1.000	1.000	0.0	115	-0.05
35	Benzo(b)fluoranthene	1.462	1.294	11.5	111	-0.05
36	Benzo(k)fluoranthene	1.592	1.499	5.8	114	-0.05
37 C	Benzo(a)pyrene	1.334	1.255	5.9	111	-0.05
38	Dibenzo(a,h)anthracene	1.314	1.201	8.6	112	-0.08
39	Benzo(a,h,i)perylene	1.400	1.300	7.1	113	-0.08

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN030920\
Data File : BN010143.D
Acq On : 09 Mar 2020 11:27
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 09 17:06:36 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN021920.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 19 15:57:43 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			