

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN030920\
 Data File : BN010156.D
 Acq On : 09 Mar 2020 21:42
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 09 22:39:07 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN021920.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 19 15:57:43 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	86	-0.04
2	1,4-Dioxane	0.410	0.470	-14.6	99	-0.05
3	n-Nitrosodimethylamine	0.673	0.712	-5.8	104	-0.02
4 S	2-Fluorophenol	0.889	0.864	2.8	87	-0.03
5 S	Phenol-d6	1.082	0.953	11.9	80	0.00
6	bis(2-Chloroethyl)ether	1.052	0.870	17.3	74	-0.03
7 I	Naphthalene-d8	1.000	1.000	0.0	92	-0.05
8 S	Nitrobenzene-d5	0.333	0.263	21.0	77	0.00
9	Naphthalene	1.091	1.042	4.5	90	-0.05
10	Hexachlorobutadiene	0.240	0.243	-1.3	93	-0.06
11 SURR	2-Methylnaphthalene-d10	0.774	0.757	2.2	93	-0.04
12	2-Methylnaphthalene	0.748	0.713	4.7	91	-0.04
13 I	Acenaphthene-d10	1.000	1.000	0.0	95	-0.04
14 S	2,4,6-Tribromophenol	0.154	0.152	1.3	105	-0.03
15 S	2-Fluorobiphenyl	1.510	1.389	8.0	91	-0.04
16	Acenaphthylene	1.705	1.614	5.3	100	-0.04
17	Acenaphthene	1.199	1.148	4.3	95	-0.04
18	Fluorene	1.582	1.458	7.8	90	-0.03
19 I	Phenanthrene-d10	1.000	1.000	0.0	88	-0.03
20	4-Bromophenyl-phenylether	0.279	0.141	49.5#	47#	-0.01
21	Hexachlorobenzene	0.249	0.248	0.4	92	-0.04
22	Pentachlorophenol	0.064	0.053	17.2	99	-0.04
23	Phenanthrene	1.190	1.065	10.5	86	-0.03
24	Anthracene	1.006	0.901	10.4	88	-0.02
25 SURR	Fluoranthene-d10	1.384	1.255	9.3	86	-0.03
26	Fluoranthene	1.425	1.278	10.3	85	-0.03
27 I	Chrysene-d12	1.000	1.000	0.0	89	-0.03
28	Pvrene	1.519	1.443	5.0	87	-0.03
29 S	Terphenyl-d14	0.956	0.894	6.5	85	-0.03
30	Benzo(a)anthracene	1.327	1.115	16.0	84	-0.03
31	Chrysene	1.536	1.453	5.4	86	-0.03
32	Bis(2-ethylhexyl)phthalate	0.648	0.626	3.4	96	-0.04
33	Indeno(1,2,3-cd)pyrene	1.599	1.595	0.3	89	-0.06
34 I	Perylene-d12	1.000	1.000	0.0	93	-0.04
35	Benzo(b)fluoranthene	1.462	1.309	10.5	91	-0.04
36	Benzo(k)fluoranthene	1.592	1.498	5.9	92	-0.04
37 C	Benzo(a)pyrene	1.334	1.255	5.9	90	-0.04
38	Dibenzo(a,h)anthracene	1.314	1.169	11.0	88	-0.07
39	Benzo(a,h,i)perylene	1.400	1.332	4.9	94	-0.07

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN030920\
Data File : BN010156.D
Acq On : 09 Mar 2020 21:42
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 09 22:39:07 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN021920.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 19 15:57:43 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			