

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030922\
 Data File : BN018893.D
 Acq On : 09 Mar 2022 20:21
 Operator : CG/JU
 Sample : PB143227BS
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS227

Quant Time: Mar 10 01:38:25 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN030822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Mar 08 14:15:20 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 03/10/2022
 Supervised By :Yogesh Patel 03/11/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.905	152	140860	20.000	ng/u1	0.00
20) Naphthalene-d8	10.710	136	672606	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.540	164	465473	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.281	188	988446	20.000	ng/u1	0.00
79) Chrysene-d12	21.451	240	1051814	20.000	ng/u1	0.00
88) Perylene-d12	23.851	264	1152304	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.270	96	18246	5.181	ng/uL	0.00
4) Pyridine-d5	3.687	84	262206	24.331	ng/u1	0.00
7) Phenol-d5	7.058	99	389769	28.342	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.228	67	234058	28.257	ng/u1	0.00
11) 2-Chlorophenol-d4	7.434	132	281864	28.600	ng/u1	0.00
15) 4-Methylphenol-d8	8.611	113	321490	28.750	ng/u1	0.00
21) Nitrobenzene-d5	9.063	128	144535	27.975	ng/u1	0.00
24) 2-Nitrophenol-d4	9.787	143	166195	29.278	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.328	165	313290	28.853	ng/u1	0.00
31) 4-Chloroaniline-d4	10.840	131	385016	23.548	ng/u1	0.00
46) Dimethylphthalate-d6	13.946	166	1018786	28.908	ng/u1	0.00
49) Acenaphthylene-d8	14.234	160	1256520	28.644	ng/u1	0.00
54) 4-Nitrophenol-d4	14.722	143	209993	29.423	ng/u1	0.00
60) Fluorene-d10	15.528	176	886755	28.947	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.645	200	187418	30.595	ng/u1	0.00
73) Anthracene-d10	17.375	188	1432206	30.293	ng/u1	0.00
81) Pyrene-d10	19.669	212	1754173	30.818	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.698	264	1873144	31.190	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.305	88	36734	9.609	ng/uL	96
5) Pyridine	3.711	79	273867	24.641	ng/u1	100
6) Benzaldehyde	7.028	77	229734	32.803	ng/u1	97
8) Phenol	7.081	94	390151	27.116	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.316	93	296361	26.702	ng/u1	98
12) 2-Chlorophenol	7.464	128	286277	27.393	ng/u1	97
13) 2-Methylphenol	8.340	108	288911	26.896	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.434	45	430330	27.159	ng/u1	99
16) Acetophenone	8.722	105	469839	27.112	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.711	70	247130	27.982	ng/u1	100
18) 4-Methylphenol	8.669	108	325847	27.125	ng/u1	94
19) Hexachloroethane	8.993	117	108817	26.686	ng/u1	97
22) Nitrobenzene	9.105	77	356487	26.587	ng/u1	99
23) Isophorone	9.628	82	687388	26.825	ng/u1	99
25) 2-Nitrophenol	9.822	139	172486	27.341	ng/u1	99
26) 2,4-Dimethylphenol	9.875	107	368121	26.866	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.116	93	426697	26.379	ng/u1	99
29) 2,4-Dichlorophenol	10.357	162	296690	26.856	ng/u1	99
30) Naphthalene	10.763	128	975984	26.331	ng/u1	99
32) 4-Chloroaniline	10.863	127	371282	21.886	ng/u1	99
33) Hexachlorobutadiene	11.057	225	179670	25.887	ng/u1	97
34) Caprolactam	11.604	113	110848m	27.591	ng/u1	
35) 4-Chloro-3-methylphenol	11.987	107	359895	27.818	ng/u1	99
36) 2-Methylnaphthalene	12.369	142	693316	26.605	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030922\
 Data File : BN018893.D
 Acq On : 09 Mar 2022 20:21
 Operator : CG/JU
 Sample : PB143227BS
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :

BNA_N

ClientSampleId :

SLCS227

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 03/10/2022
 Supervised By :Yogesh Patel 03/11/2022

Quant Time: Mar 10 01:38:25 2022

Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN030822.M

Quant Title : SVOA CALIBRATION

QLast Update : Tue Mar 08 14:15:20 2022

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.587	142	713413	26.510	ng/ul	94
39) 1,2,4,5-Tetrachloroben...	12.740	216	367658	25.834	ng/ul	100
40) Hexachlorocyclopentadiene	12.728	237	211159	25.255	ng/ul	98
41) 2,4,6-Trichlorophenol	12.975	196	245881	26.727	ng/ul	99
42) 2,4,5-Trichlorophenol	13.046	196	275817	27.143	ng/ul	95
43) 1,1'-Biphenyl	13.375	154	965304	26.385	ng/ul	98
44) 2-Chloronaphthalene	13.416	162	733977	26.051	ng/ul	98
45) 2-Nitroaniline	13.610	65	241686	28.782	ng/ul	99
47) Dimethylphthalate	13.993	163	995088	27.539	ng/ul	99
48) 2,6-Dinitrotoluene	14.110	165	201257	29.067	ng/ul	93
50) Acenaphthylene	14.263	152	1232546	26.866	ng/ul	97
51) 3-Nitroaniline	14.434	138	193435	28.841	ng/ul	95
52) Acenaphthene	14.604	153	786583	26.192	ng/ul	98
53) 2,4-Dinitrophenol	14.640	184	111264	25.136	ng/ul	97
55) 4-Nitrophenol	14.734	109	170160	27.689	ng/ul	99
56) Dibenzofuran	14.934	168	1160144	26.743	ng/ul	98
57) 2,4-Dinitrotoluene	14.893	165	304766	29.244	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	15.163	232	236750	28.326	ng/ul	96
59) Diethylphthalate	15.351	149	996239	27.702	ng/ul	98
61) Fluorene	15.587	166	904180	26.473	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.575	204	455700	27.085	ng/ul	99
63) 4-Nitroaniline	15.592	138	217574	34.838	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.657	198	179981	28.835	ng/ul	97
67) N-Nitrosodiphenylamine	15.787	169	829192	28.133	ng/ul	99
68) 4-Bromophenyl-phenylether	16.469	248	282772	27.753	ng/ul	97
69) Hexachlorobenzene	16.592	284	331343	28.362	ng/ul	97
70) Atrazine	16.734	200	315277	28.549	ng/ul	98
71) Pentachlorophenol	16.928	266	199423	27.397	ng/ul	99
72) Phenanthrene	17.322	178	1571980	27.916	ng/ul	99
74) Anthracene	17.410	178	1595590	28.415	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.346	216	397160	25.984	ng/ul	99
76) Pentachlorobenzene	14.863	250	386215	27.926	ng/ul	97
77) Carbazole	17.675	167	1470761	29.007	ng/ul	98
78) Di-n-butylphthalate	18.245	149	1687120	29.608	ng/ul	98
80) Fluoranthene	19.333	202	1928088	28.615	ng/ul	100
82) Pyrene	19.698	202	2007416	29.164	ng/ul	100
83) Butylbenzylphthalate	20.586	149	812292	30.998	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.369	252	614614	26.965	ng/ul	97
85) Benzo(a)anthracene	21.439	228	2014979	28.353	ng/ul	97
86) Bis(2-ethylhexyl)phtha...	21.369	149	1195353	30.313	ng/ul	99
87) Chrysene	21.492	228	1981966	28.800	ng/ul	98
89) Di-n-octyl phthalate	22.286	149	2163884	29.965	ng/ul	100
90) Benzo(b)fluoranthene	23.122	252	2148153	27.558	ng/ul	98
91) Benzo(k)fluoranthene	23.169	252	2081720	28.778	ng/ul	97
93) Benzo(a)pyrene	23.745	252	2157482	28.592	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	26.333	276	2489866	29.789	ng/ul	99
95) Dibenzo(a,h)anthracene	26.345	278	2107303	28.438	ng/ul	97
96) Benzo(g,h,i)perylene	27.098	276	2113356	28.417	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030922\
 Data File : BN018893.D
 Acq On : 09 Mar 2022 20:21
 Operator : CG/JU
 Sample : PB143227BS
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :

BNA_N

ClientSampleId :


SLCS227


Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 03/10/2022
 Supervised By :Yogesh Patel 03/11/2022


Quant Time: Mar 10 01:38:25 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN030822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Mar 08 14:15:20 2022
 Response via : Initial Calibration





#2
1,4-Dioxane
Concen: 9.609 ng/uL
RT: 3.305 min Scan# 54
Delta R.T. -0.000 min
Lab File: BN018893.D
Acq: 09 Mar 2022 20:21




Instrument :
BNA_N
ClientSampleId :
SLCS227



Tgt Ion: 88 Resp: 36734
Ion Ratio Lower Upper
88 100
43 30.9 22.4 33.6
58 73.6 61.0 91.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 03/10/2022
Supervised By :Yogesh Patel 03/11/2022



#5
Pyridine
Concen: 24.641 ng/uL
RT: 3.711 min Scan# 123
Delta R.T. 0.000 min
Lab File: BN018893.D
Acq: 09 Mar 2022 20:21


Tgt Ion: 79 Resp: 273867
Ion Ratio Lower Upper
79 100
52 81.0 65.0 97.4
51 38.9 31.6 47.4





#6
Benzaldehyde
Concen: 32.803 ng/ul
RT: 7.028 min Scan# 687
Delta R.T. -0.006 min
Lab File: BN018893.D
Acq: 09 Mar 2022 20:21




Instrument :
BNA_N
ClientSampleId :
SLCS227



Tgt Ion: 77 Resp: 229734
Ion Ratio Lower Upper
77 100
105 96.9 74.7 112.1
106 90.0 70.2 105.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 03/10/2022
Supervised By :Yogesh Patel 03/11/2022



#8
Phenol
Concen: 27.116 ng/ul
RT: 7.081 min Scan# 696
Delta R.T. -0.006 min
Lab File: BN018893.D
Acq: 09 Mar 2022 20:21


Tgt Ion: 94 Resp: 390151
Ion Ratio Lower Upper
94 100
65 29.3 23.6 35.4
66 39.9 31.2 46.8





#10
Bis(2-Chloroethyl)ether
Concen: 26.702 ng/ul
RT: 7.316 min Scan# 736
Delta R.T. 0.000 min
Lab File: BN018893.D
Acq: 09 Mar 2022 20:21




Instrument :
BNA_N
ClientSampleId :
SLCS227




Tgt Ion: 93 Resp: 29636
Ion Ratio Lower Upper
93 100
63 81.2 63.4 95.0
95 31.8 26.2 39.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 03/10/2022
Supervised By :Yogesh Patel 03/11/2022




#12
2-Chlorophenol
Concen: 27.393 ng/ul
RT: 7.464 min Scan# 761
Delta R.T. 0.000 min
Lab File: BN018893.D
Acq: 09 Mar 2022 20:21



Tgt Ion:128 Resp: 286277
Ion Ratio Lower Upper
128 100
64 50.6 38.6 57.8
130 31.8 25.0 37.4





#13

2-Methylphenol

Concen: 26.896 ng/ul

RT: 8.340 min Scan# 910

Delta R.T. -0.006 min

Lab File: BN018893.D

Acq: 09 Mar 2022 20:21

Instrument :

BNA_N

ClientSampleId :

SLCS227

Tgt Ion:108 Resp: 28891

Ion Ratio Lower Upper

108 100





107 88.7 71.5 107.3

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 03/10/2022

Supervised By :Yogesh Patel 03/11/2022



#14

2,2'-oxybis(1-Chloropropane)

Concen: 27.159 ng/ul

RT: 8.434 min Scan# 926

Delta R.T. -0.006 min

Lab File: BN018893.D

Acq: 09 Mar 2022 20:21




Tgt Ion: 45 Resp: 430330


Ion Ratio Lower Upper

45 100

77 13.8 10.9 16.3


79 10.0 8.6 13.0





#16
Acetophenone
Concen: 27.112 ng/ul
RT: 8.722 min Scan# 975
Delta R.T. 0.000 min
Lab File: BN018893.D
Acq: 09 Mar 2022 20:21




Instrument :
BNA_N
ClientSampleId :
SLCS227



Tgt Ion:105 Resp: 469839
Ion Ratio Lower Upper
105 100
77 78.7 63.5 95.3
51 29.2 23.5 35.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 03/10/2022
Supervised By :Yogesh Patel 03/11/2022



#17
N-Nitroso-di-n-propylamine
Concen: 27.982 ng/ul
RT: 8.711 min Scan# 973
Delta R.T. 0.000 min
Lab File: BN018893.D
Acq: 09 Mar 2022 20:21


Tgt Ion: 70 Resp: 247130
Ion Ratio Lower Upper
70 100
42 59.8 47.8 71.6
101 9.3 7.8 11.6
130 20.0 15.8 23.8





#18
4-Methylphenol
Concen: 27.125 ng/ul
RT: 8.669 min Scan# 966
Delta R.T. 0.000 min
Lab File: BN018893.D
Acq: 09 Mar 2022 20:21




Instrument :
BNA_N
ClientSampleId :
SLCS227



Tgt Ion:108 Resp: 32584
Ion Ratio Lower Upper
108 100
107 116.0 88.0 132.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 03/10/2022
Supervised By :Yogesh Patel 03/11/2022



#19
Hexachloroethane
Concen: 26.686 ng/ul
RT: 8.993 min Scan# 1021
Delta R.T. 0.000 min
Lab File: BN018893.D
Acq: 09 Mar 2022 20:21

Tgt Ion:117 Resp: 108817
Ion Ratio Lower Upper
117 100
201 101.1 83.2 124.8
199 60.2 50.2 75.4





#22

Nitrobenzene

Concen: 26.587 ng/ul

RT: 9.105 min Scan# 1040

Delta R.T. 0.000 min

Lab File: BN018893.D

Acq: 09 Mar 2022 20:21

Instrument :

BNA_N

ClientSampleId :

SLCS227

Tgt Ion: 77 Resp: 35648

Ion Ratio Lower Upper

77 100

123 43.6 34.5 51.7





65 13.8 11.1 16.7

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 03/10/2022

Supervised By :Yogesh Patel 03/11/2022



#23

Isophorone

Concen: 26.825 ng/ul

RT: 9.628 min Scan# 1129

Delta R.T. 0.000 min

Lab File: BN018893.D

Acq: 09 Mar 2022 20:21




Tgt Ion: 82 Resp: 687388


Ion Ratio Lower Upper

82 100

95 7.5 5.8 8.8


138 16.9 13.8 20.8





#25
2-Nitrophenol
Concen: 27.341 ng/ul
RT: 9.822 min Scan# 1161
Delta R.T. 0.000 min
Lab File: BN018893.D
Acq: 09 Mar 2022 20:21




Instrument :
BNA_N
ClientSampleId :
SLCS227



Tgt Ion:139 Resp: 172480
Ion Ratio Lower Upper
139 100
65 54.3 43.7 65.5
109 31.3 25.8 38.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 03/10/2022
Supervised By :Yogesh Patel 03/11/2022



#26
2,4-Dimethylphenol
Concen: 26.866 ng/ul
RT: 9.875 min Scan# 1171
Delta R.T. -0.006 min
Lab File: BN018893.D
Acq: 09 Mar 2022 20:21

Tgt Ion:107 Resp: 368121
Ion Ratio Lower Upper
107 100
121 48.1 38.3 57.5
122 81.0 66.3 99.5





#27

Bis(2-Chloroethoxy)methane

Concen: 26.379 ng/ul

RT: 10.116 min Scan# 1212

Delta R.T. 0.000 min

Lab File: BN018893.D

Acq: 09 Mar 2022 20:21

Instrument :

BNA_N

ClientSampleId :

SLCS227

Tgt Ion: 93 Resp: 426690

Ion Ratio Lower Upper

93 100

95 32.1 25.4 38.0





123 11.5 8.3 12.5

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 03/10/2022

Supervised By :Yogesh Patel 03/11/2022



#29

2,4-Dichlorophenol

Concen: 26.856 ng/ul

RT: 10.357 min Scan# 1253

Delta R.T. 0.000 min

Lab File: BN018893.D

Acq: 09 Mar 2022 20:21




Tgt Ion:162 Resp: 296690


Ion Ratio Lower Upper

162 100

164 64.8 52.9 79.3


98 37.4 29.8 44.8





#30
Naphthalene
Concen: 26.331 ng/ul
RT: 10.763 min Scan# 1321
Delta R.T. 0.000 min
Lab File: BN018893.D
Acq: 09 Mar 2022 20:21




Instrument :
BNA_N
ClientSampleId :
SLCS227



Tgt Ion:128 Resp: 975984
Ion Ratio Lower Upper
128 100
129 10.9 8.7 13.1
127 12.9 10.6 16.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 03/10/2022
Supervised By :Yogesh Patel 03/11/2022



#32
4-Chloroaniline
Concen: 21.886 ng/ul
RT: 10.863 min Scan# 1339
Delta R.T. 0.000 min
Lab File: BN018893.D
Acq: 09 Mar 2022 20:21


Tgt Ion:127 Resp: 371282
Ion Ratio Lower Upper
127 100
129 33.4 26.1 39.1





#33
Hexachlorobutadiene
Concen: 25.887 ng/ul
RT: 11.057 min Scan# 1372
Delta R.T. 0.000 min
Lab File: BN018893.D
Acq: 09 Mar 2022 20:21




Instrument :
BNA_N
ClientSampleId :
SLCS227




Tgt Ion:225 Resp: 179670
Ion Ratio Lower Upper
225 100
223 65.4 51.8 77.6
227 67.3 50.6 76.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 03/10/2022
Supervised By :Yogesh Patel 03/11/2022




#34
Caprolactam
Concen: 27.591 ng/ul m
RT: 11.604 min Scan# 1465
Delta R.T. 0.000 min
Lab File: BN018893.D
Acq: 09 Mar 2022 20:21




Tgt Ion:113 Resp: 110848
Ion Ratio Lower Upper
113 100
55 180.1 145.5 218.3
56 135.4 106.5 159.7





#35
4-Chloro-3-methylphenol
Concen: 27.818 ng/ul
RT: 11.987 min Scan# 1530
Delta R.T. 0.000 min
Lab File: BN018893.D
Acq: 09 Mar 2022 20:21




Instrument :
BNA_N
ClientSampleId :
SLCS227



Tgt Ion:107 Resp: 35989
Ion Ratio Lower Upper
107 100
144 24.9 20.6 31.0
142 77.5 62.2 93.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 03/10/2022
Supervised By :Yogesh Patel 03/11/2022



#36
2-Methylnaphthalene
Concen: 26.605 ng/ul
RT: 12.369 min Scan# 1595
Delta R.T. 0.000 min
Lab File: BN018893.D
Acq: 09 Mar 2022 20:21

Tgt Ion:142 Resp: 693316
Ion Ratio Lower Upper
142 100
141 87.0 70.6 105.8





#37

1-Methylnaphthalene

Concen: 26.510 ng/ul

RT: 12.587 min Scan# 1632

Delta R.T. -0.006 min

Lab File: BN018893.D

Acq: 09 Mar 2022 20:21

Instrument :

BNA_N

ClientSampleId :





SLCS227

Tgt Ion:	142	Resp:	71341
Ion Ratio	Lower	Upper	
142	100		
141	94.0	70.9	106.3
116	4.1	3.1	4.7

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 03/10/2022
Supervised By :Yogesh Patel 03/11/2022



#39

1,2,4,5-Tetrachlorobenzene

Concen: 25.834 ng/ul




RT: 12.740 min Scan# 1658


Delta R.T. 0.000 min

Lab File: BN018893.D

Acq: 09 Mar 2022 20:21


Tgt Ion:	216	Resp:	367658
Ion Ratio	Lower	Upper	
216	100		
214	79.0	63.4	95.0
179	22.1	17.8	26.8
108	18.4	15.1	22.7





#40
Hexachlorocyclopentadiene
Concen: 25.255 ng/ul
RT: 12.728 min Scan# 1656
Delta R.T. 0.000 min
Lab File: BN018893.D
Acq: 09 Mar 2022 20:21




Instrument :
BNA_N
ClientSampleId :
SLCS227




Tgt Ion:237 Resp: 211159
Ion Ratio Lower Upper
237 100
235 59.7 48.8 73.2
272 13.2 9.3 13.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 03/10/2022
Supervised By :Yogesh Patel 03/11/2022




#41
2,4,6-Trichlorophenol
Concen: 26.727 ng/ul
RT: 12.975 min Scan# 1698
Delta R.T. 0.000 min
Lab File: BN018893.D
Acq: 09 Mar 2022 20:21



Tgt Ion:196 Resp: 245881
Ion Ratio Lower Upper
196 100
198 98.8 79.8 119.8
200 31.6 25.0 37.6





#42
2,4,5-Trichlorophenol
Concen: 27.143 ng/ul
RT: 13.046 min Scan# 1709
Delta R.T. 0.000 min
Lab File: BN018893.D
Acq: 09 Mar 2022 20:21

Instrument :

BNA_N

ClientSampleId :





SLCS227

Tgt Ion:196 Resp: 27581
Ion Ratio Lower Upper
196 100
198 102.4 77.6 116.4
200 31.8 24.4 36.6

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 03/10/2022
Supervised By :Yogesh Patel 03/11/2022



#43
1,1'-Biphenyl
Concen: 26.385 ng/ul
RT: 13.375 min Scan# 1766
Delta R.T. 0.000 min
Lab File: BN018893.D
Acq: 09 Mar 2022 20:21


Tgt Ion:154 Resp: 965304
Ion Ratio Lower Upper
154 100
153 40.8 31.5 47.3
76 13.9 10.5 15.7





#44
2-Chloronaphthalene
Concen: 26.051 ng/ul
RT: 13.416 min Scan# 1773
Delta R.T. -0.006 min
Lab File: BN018893.D
Acq: 09 Mar 2022 20:21




Instrument :
BNA_N
ClientSampleId :
SLCS227



Tgt Ion:162 Resp: 73397
Ion Ratio Lower Upper
162 100
164 32.3 25.8 38.6
127 41.2 31.4 47.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 03/10/2022
Supervised By :Yogesh Patel 03/11/2022



#45
2-Nitroaniline
Concen: 28.782 ng/ul
RT: 13.610 min Scan# 1806
Delta R.T. 0.000 min
Lab File: BN018893.D
Acq: 09 Mar 2022 20:21


Tgt Ion: 65 Resp: 241686
Ion Ratio Lower Upper
65 100
92 67.3 54.1 81.1
138 98.2 77.0 115.4





#47
Dimethylphthalate
Concen: 27.539 ng/ul
RT: 13.993 min Scan# 1871
Delta R.T. 0.000 min
Lab File: BN018893.D
Acq: 09 Mar 2022 20:21




Instrument :
BNA_N
ClientSampleId :
SLCS227



Tgt Ion:163 Resp: 995088
Ion Ratio Lower Upper
163 100
194 4.1 3.4 5.0
164 10.3 7.7 11.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 03/10/2022
Supervised By :Yogesh Patel 03/11/2022



#48
2,6-Dinitrotoluene
Concen: 29.067 ng/ul
RT: 14.110 min Scan# 1891
Delta R.T. 0.000 min
Lab File: BN018893.D
Acq: 09 Mar 2022 20:21


Tgt Ion:165 Resp: 201257
Ion Ratio Lower Upper
165 100
89 56.9 50.6 75.8
121 20.3 17.9 26.9





#50
Acenaphthylene
Concen: 26.866 ng/ul
RT: 14.263 min Scan# 1916
Delta R.T. 0.000 min
Lab File: BN018893.D
Acq: 09 Mar 2022 20:21




Instrument :
BNA_N
ClientSampleId :
SLCS227



Tgt Ion:152 Resp: 1232540
Ion Ratio Lower Upper
152 100
151 18.8 16.0 24.0
153 12.3 10.7 16.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 03/10/2022
Supervised By :Yogesh Patel 03/11/2022



#51
3-Nitroaniline
Concen: 28.841 ng/ul
RT: 14.434 min Scan# 1946
Delta R.T. 0.000 min
Lab File: BN018893.D
Acq: 09 Mar 2022 20:21


Tgt Ion:138 Resp: 193435
Ion Ratio Lower Upper
138 100
108 11.1 8.6 12.8
92 121.4 92.9 139.3





#52
Acenaphthene
Concen: 26.192 ng/ul
RT: 14.604 min Scan# 1975
Delta R.T. 0.000 min
Lab File: BN018893.D
Acq: 09 Mar 2022 20:21




Instrument :
BNA_N
ClientSampleId :
SLCS227




Tgt Ion:153 Resp: 78658
Ion Ratio Lower Upper
153 100
152 46.0 38.4 57.6
154 88.1 69.7 104.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 03/10/2022
Supervised By :Yogesh Patel 03/11/2022




#53
2,4-Dinitrophenol
Concen: 25.136 ng/ul
RT: 14.640 min Scan# 1981
Delta R.T. 0.000 min
Lab File: BN018893.D
Acq: 09 Mar 2022 20:21




Tgt Ion:184 Resp: 111264
Ion Ratio Lower Upper
184 100
63 65.9 55.9 83.9
154 65.1 53.1 79.7





#55
4-Nitrophenol
Concen: 27.689 ng/ul
RT: 14.734 min Scan# 1997
Delta R.T. 0.000 min
Lab File: BN018893.D
Acq: 09 Mar 2022 20:21

Instrument : BNA_N
ClientSampleId : SLCS227






Tgt Ion:109 Resp: 170160

Ion	Ratio	Lower	Upper
109	100		
139	126.6	100.1	150.1
65	132.4	104.6	157.0

Manual Integrations
APPROVED




Reviewed By :Jagrut Upadhyay 03/10/2022
Supervised By :Yogesh Patel 03/11/2022




#56
Dibenzofuran
Concen: 26.743 ng/ul
RT: 14.934 min Scan# 2031
Delta R.T. -0.006 min
Lab File: BN018893.D
Acq: 09 Mar 2022 20:21

Tgt Ion:168 Resp: 1160144


Ion	Ratio	Lower	Upper
168	100		
139	37.5	31.1	46.7





#57
2,4-Dinitrotoluene
Concen: 29.244 ng/ul
RT: 14.893 min Scan# 2070
Delta R.T. 0.000 min
Lab File: BN018893.D
Acq: 09 Mar 2022 20:21




Instrument :
BNA_N
ClientSampleId :
SLCS227




Tgt Ion:165 Resp: 304760
Ion Ratio Lower Upper
165 100
63 45.3 36.3 54.5
182 2.8 2.3 3.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 03/10/2022
Supervised By :Yogesh Patel 03/11/2022




#58
2,3,4,6-Tetrachlorophenol
Concen: 28.326 ng/ul
RT: 15.163 min Scan# 2070
Delta R.T. 0.000 min
Lab File: BN018893.D
Acq: 09 Mar 2022 20:21




Tgt Ion:232 Resp: 236750
Ion Ratio Lower Upper
232 100
131 44.5 38.5 57.7
130 2.4 2.2 3.2
166 31.0 25.8 38.8





#59
Diethylphthalate
Concen: 27.702 ng/ul
RT: 15.351 min Scan# 2102
Delta R.T. 0.000 min
Lab File: BN018893.D
Acq: 09 Mar 2022 20:21




Instrument :
BNA_N
ClientSampleId :
SLCS227




Tgt Ion:149 Resp: 996239
Ion Ratio Lower Upper
149 100
177 21.5 17.8 26.8
150 12.4 10.5 15.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 03/10/2022
Supervised By :Yogesh Patel 03/11/2022




#61
Fluorene
Concen: 26.473 ng/ul
RT: 15.587 min Scan# 2142
Delta R.T. 0.000 min
Lab File: BN018893.D
Acq: 09 Mar 2022 20:21




Tgt Ion:166 Resp: 904180
Ion Ratio Lower Upper
166 100
165 96.1 77.7 116.5
167 13.3 11.1 16.7





#62
4-Chlorophenyl-phenylether
Concen: 27.085 ng/ul
RT: 15.575 min Scan# 2140
Delta R.T. 0.000 min
Lab File: BN018893.D
Acq: 09 Mar 2022 20:21




Instrument :
BNA_N
ClientSampleId :
SLCS227




Tgt Ion:204 Resp: 455700
Ion Ratio Lower Upper
204 100
206 32.8 26.7 40.1
141 61.4 50.2 75.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 03/10/2022
Supervised By :Yogesh Patel 03/11/2022




#63
4-Nitroaniline
Concen: 34.838 ng/ul
RT: 15.592 min Scan# 2143
Delta R.T. 0.000 min
Lab File: BN018893.D
Acq: 09 Mar 2022 20:21




Tgt Ion:138 Resp: 217574
Ion Ratio Lower Upper
138 100
92 51.0 41.0 61.6
108 83.7 65.3 97.9





#66
4,6-Dinitro-2-methylphenol
Concen: 28.835 ng/ul
RT: 15.657 min Scan# 2154
Delta R.T. 0.000 min
Lab File: BN018893.D
Acq: 09 Mar 2022 20:21




Instrument :
BNA_N
ClientSampleId :
SLCS227



Tgt Ion:198 Resp: 17998
Ion Ratio Lower Upper
198 100
182 4.4 3.0 4.4
77 26.4 20.1 30.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 03/10/2022
Supervised By :Yogesh Patel 03/11/2022



#67
N-Nitrosodiphenylamine
Concen: 28.133 ng/ul
RT: 15.787 min Scan# 2176
Delta R.T. 0.000 min
Lab File: BN018893.D
Acq: 09 Mar 2022 20:21


Tgt Ion:169 Resp: 829192
Ion Ratio Lower Upper
169 100
168 65.4 52.4 78.6
167 35.6 27.5 41.3





#68
4-Bromophenyl-phenylether
Concen: 27.753 ng/ul
RT: 16.469 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN018893.D
Acq: 09 Mar 2022 20:21




Instrument :
BNA_N
ClientSampleId :
SLCS227




Tgt Ion:248 Resp: 282772
Ion Ratio Lower Upper
248 100
250 99.4 77.8 116.8
141 81.0 62.6 93.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 03/10/2022
Supervised By :Yogesh Patel 03/11/2022




#69
Hexachlorobenzene
Concen: 28.362 ng/ul
RT: 16.592 min Scan# 2313
Delta R.T. 0.000 min
Lab File: BN018893.D
Acq: 09 Mar 2022 20:21




Tgt Ion:284 Resp: 331343
Ion Ratio Lower Upper
284 100
142 37.5 27.8 41.8
249 32.9 25.7 38.5





#70
Atrazine
Concen: 28.549 ng/ul
RT: 16.734 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN018893.D
Acq: 09 Mar 2022 20:21




Instrument :
BNA_N
ClientSampleId :
SLCS227




Tgt Ion:200 Resp: 31527
Ion Ratio Lower Upper
200 100
173 25.6 20.5 30.7
215 46.0 38.9 58.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 03/10/2022
Supervised By :Yogesh Patel 03/11/2022




#71
Pentachlorophenol
Concen: 27.397 ng/ul
RT: 16.928 min Scan# 2370
Delta R.T. 0.000 min
Lab File: BN018893.D
Acq: 09 Mar 2022 20:21




Tgt Ion:266 Resp: 199423
Ion Ratio Lower Upper
266 100
264 63.7 50.2 75.2
268 62.0 49.3 73.9





#72
Phenanthrene
Concen: 27.916 ng/ul
RT: 17.322 min Scan# 2437
Delta R.T. 0.000 min
Lab File: BN018893.D
Acq: 09 Mar 2022 20:21




Instrument :
BNA_N
ClientSampleId :
SLCS227




Tgt Ion:178 Resp: 1571980
Ion Ratio Lower Upper
178 100
179 15.7 12.2 18.4
176 19.1 15.2 22.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 03/10/2022
Supervised By :Yogesh Patel 03/11/2022




#74
Anthracene
Concen: 28.415 ng/ul
RT: 17.410 min Scan# 2452
Delta R.T. 0.000 min
Lab File: BN018893.D
Acq: 09 Mar 2022 20:21



Tgt Ion:178 Resp: 1595590
Ion Ratio Lower Upper
178 100
179 15.3 13.0 19.4
176 18.2 15.5 23.3





#75

1,2,3,4-Tetrachlorobenzene

Concen: 25.984 ng/uL

RT: 13.346 min Scan# 1761

Delta R.T. 0.000 min

Lab File: BN018893.D

Acq: 09 Mar 2022 20:21

Instrument :

BNA_N

ClientSampleId :

SLCS227

Tgt Ion:216 Resp: 397160

Ion Ratio Lower Upper

216 100

214 78.2 63.2 94.8





218 48.7 37.9 56.9

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 03/10/2022

Supervised By :Yogesh Patel 03/11/2022



#76

Pentachlorobenzene

Concen: 27.926 ng/uL

RT: 14.863 min Scan# 2019

Delta R.T. 0.000 min

Lab File: BN018893.D

Acq: 09 Mar 2022 20:21




Tgt Ion:250 Resp: 386215


Ion Ratio Lower Upper

250 100

248 62.3 53.0 79.4


252 67.4 53.4 80.2





#77
 Carbazole
 Concen: 29.007 ng/ul
 RT: 17.675 min Scan# 2497
 Delta R.T. 0.000 min
 Lab File: BN018893.D
 Acq: 09 Mar 2022 20:21




Instrument :
 BNA_N
 ClientSampleId :
 SLCS227




Tgt Ion:167 Resp: 147076
 Ion Ratio Lower Upper
 167 100
 166 21.3 16.1 24.1
 139 13.3 10.2 15.2

Manual Integrations
 APPROVED



Reviewed By :Jagrut Upadhyay 03/10/2022
 Supervised By :Yogesh Patel 03/11/2022




#78
 Di-n-butylphthalate
 Concen: 29.608 ng/ul
 RT: 18.245 min Scan# 2594
 Delta R.T. 0.000 min
 Lab File: BN018893.D
 Acq: 09 Mar 2022 20:21




Tgt Ion:149 Resp: 1687120
 Ion Ratio Lower Upper
 149 100
 150 9.6 7.1 10.7
 104 5.4 4.1 6.1





#80
Fluoranthene
Concen: 28.615 ng/ul
RT: 19.333 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN018893.D
Acq: 09 Mar 2022 20:21




Instrument :
BNA_N
ClientSampleId :
SLCS227




Tgt Ion:202 Resp: 1928088
Ion Ratio Lower Upper
202 100
101 10.8 8.9 13.3
100 8.4 6.7 10.1

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 03/10/2022
Supervised By :Yogesh Patel 03/11/2022




#82
Pyrene
Concen: 29.164 ng/ul
RT: 19.698 min Scan# 2841
Delta R.T. 0.000 min
Lab File: BN018893.D
Acq: 09 Mar 2022 20:21




Tgt Ion:202 Resp: 2007416
Ion Ratio Lower Upper
202 100
101 12.0 9.7 14.5
100 9.8 8.0 12.0





#83
Butylbenzylphthalate
Concen: 30.998 ng/ul
RT: 20.586 min Scan# 2992
Delta R.T. -0.006 min
Lab File: BN018893.D
Acq: 09 Mar 2022 20:21




Instrument :
BNA_N
ClientSampleId :
SLCS227



Tgt Ion:149 Resp: 812292
Ion Ratio Lower Upper
149 100
91 73.2 58.8 88.2
206 20.7 18.1 27.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 03/10/2022
Supervised By :Yogesh Patel 03/11/2022



#84
3,3'-Dichlorobenzidine
Concen: 26.965 ng/ul
RT: 21.369 min Scan# 3125
Delta R.T. 0.000 min
Lab File: BN018893.D
Acq: 09 Mar 2022 20:21


Tgt Ion:252 Resp: 614614
Ion Ratio Lower Upper
252 100
254 60.9 50.8 76.2
126 10.8 8.2 12.4





#85
Benzo(a)anthracene
Concen: 28.353 ng/ul
RT: 21.439 min Scan# 3137
Delta R.T. 0.000 min
Lab File: BN018893.D
Acq: 09 Mar 2022 20:21




Instrument :
BNA_N
ClientSampleId :
SLCS227



Tgt Ion:228 Resp: 2014979
Ion Ratio Lower Upper
228 100
229 19.8 16.3 24.5
226 27.0 23.2 34.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 03/10/2022
Supervised By :Yogesh Patel 03/11/2022



#86
Bis(2-ethylhexyl)phthalate
Concen: 30.313 ng/ul
RT: 21.369 min Scan# 3125
Delta R.T. 0.000 min
Lab File: BN018893.D
Acq: 09 Mar 2022 20:21


Tgt Ion:149 Resp: 1195353
Ion Ratio Lower Upper
149 100
167 27.5 22.5 33.7
279 4.2 3.5 5.3





#87
Chrysene
Concen: 28.800 ng/ul
RT: 21.492 min Scan# 3146
Delta R.T. 0.000 min
Lab File: BN018893.D
Acq: 09 Mar 2022 20:21




Instrument :
BNA_N
ClientSampleId :
SLCS227



Tgt Ion:228 Resp: 1981960
Ion Ratio Lower Upper
228 100
226 31.6 24.2 36.2
229 20.9 16.2 24.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 03/10/2022
Supervised By :Yogesh Patel 03/11/2022



#89
Di-n-octyl phthalate
Concen: 29.965 ng/ul
RT: 22.286 min Scan# 3281
Delta R.T. -0.006 min
Lab File: BN018893.D
Acq: 09 Mar 2022 20:21


Tgt Ion:149 Resp: 2163884





#90
Benzo(b)fluoranthene
Concen: 27.558 ng/u1
RT: 23.122 min Scan# 3422
Delta R.T. 0.000 min
Lab File: BN018893.D
Acq: 09 Mar 2022 20:21




Instrument :
BNA_N
ClientSampleId :
SLCS227



Tgt Ion:252 Resp: 214815
Ion Ratio Lower Upper
252 100
253 21.3 16.5 24.7
125 9.3 7.8 11.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 03/10/2022
Supervised By :Yogesh Patel 03/11/2022



#91
Benzo(k)fluoranthene
Concen: 28.778 ng/u1
RT: 23.169 min Scan# 3431
Delta R.T. 0.000 min
Lab File: BN018893.D
Acq: 09 Mar 2022 20:21

Tgt Ion:252 Resp: 2081720
Ion Ratio Lower Upper
252 100
253 21.6 18.2 27.4
125 9.3 8.2 12.4





#93

Benzo(a)pyrene

Concen: 28.592 ng/u1

RT: 23.745 min Scan# 3529

Delta R.T. 0.000 min

Lab File: BN018893.D

Acq: 09 Mar 2022 20:21

Instrument :

BNA_N

ClientSampleId :

SLCS227

Tgt Ion:252 Resp: 215748

Ion Ratio Lower Upper

252 100

253 22.4 18.7 28.1





125 10.8 9.2 13.8

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 03/10/2022

Supervised By :Yogesh Patel 03/11/2022



#94

Indeno(1,2,3-cd)pyrene

Concen: 29.789 ng/u1

RT: 26.333 min Scan# 3969

Delta R.T. 0.000 min

Lab File: BN018893.D

Acq: 09 Mar 2022 20:21




Tgt Ion:276 Resp: 2489866


Ion Ratio Lower Upper

276 100

138 25.9 19.5 29.3

277 26.7 21.4 32.0






#95
Dibenzo(a,h)anthracene
Concen: 28.438 ng/ul
RT: 26.345 min Scan# 3972
Delta R.T. -0.012 min
Lab File: BN018893.D
Acq: 09 Mar 2022 20:21

Instrument :

BNA_N

ClientSampleId :

SLCS227






Tgt Ion:278 Resp: 210730
Ion Ratio Lower Upper
278 100
139 15.6 13.8 20.8
279 21.9 18.3 27.5

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 03/10/2022
Supervised By :Yogesh Patel 03/11/2022



#96
Benzo(g,h,i)perylene
Concen: 28.417 ng/ul
RT: 27.098 min Scan# 4099
Delta R.T. 0.000 min
Lab File: BN018893.D
Acq: 09 Mar 2022 20:21

Tgt Ion:276 Resp: 2113356
Ion Ratio Lower Upper
276 100
138 20.6 17.0 25.6
277 23.0 17.8 26.6

