


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031022\
 Data File : BN018895.D
 Acq On : 10 Mar 2022 09:28
 Operator : CG/JU
 Sample : DFTPP235
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP235

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN030822.M
 Title : SVOA CALIBRATION
 Last Update : Tue Mar 08 14:15:20 2022



AutoFind: Scans 2504, 2505, 2506; Background Corrected with Scan 2498

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	38.6	134959	PASS
68	69	0.00	2	1.8	2351	PASS
69	198	0.00	100	37.8	132128	PASS
70	69	0.00	2	0.5	667	PASS
127	198	10	80	46.8	163627	PASS
197	198	0.00	2	0.2	611	PASS
198	198	100	100	100.0	349888	PASS
199	198	5	9	6.9	24061	PASS
275	198	10	60	26.6	93107	PASS
365	198	1	100	3.3	11591	PASS
441	198	0.01	100	10.3	35976	PASS
442	198	50	100	64.4	225493	PASS
443	442	15	24	19.1	43176	PASS

m/z	Abundance
40.00	1486.0
43.00	182.0
43.90	1166.0
44.90	155.0
48.90	438.0
50.90	308.0
73.10	504.0
83.90	339.0
85.90	342.0
95.80	152.0
132.80	209.0
147.00	268.0
191.00	481.0
193.00	209.0
207.00	3040.0
208.00	635.0
209.00	486.0
281.00	230.0

Instrument :

BNA_N

ClientSampleId :

DFTPP235

m/z	Abundance
35.90	153.0
36.90	722.0
38.00	1783.0
39.00	10679.0
39.90	1677.0
41.00	464.0
43.10	195.0
43.90	1078.0
44.90	322.0
46.90	178.0
49.00	1632.0
50.00	36760.0
51.00	145344.0
52.00	7257.0
52.90	233.0
55.00	725.0
56.00	4340.0
57.00	9756.0
58.00	432.0
59.00	297.0
60.00	158.0
61.00	2062.0
62.00	2206.0
63.00	5330.0
64.00	829.0
65.00	3089.0
66.00	170.0
67.00	190.0
68.00	2409.0
69.00	139008.0
69.90	641.0
72.90	1633.0
74.00	15051.0
75.00	23168.0
76.00	7898.0
77.00	158208.0
78.00	12172.0
79.00	11399.0
80.00	7889.0
81.00	11251.0
82.00	2880.0
83.00	2777.0
84.00	626.0
85.00	2066.0
86.00	3390.0
87.00	1413.0
88.00	657.0
89.00	314.0
91.00	2446.0
92.00	3073.0
93.00	18368.0
94.00	1229.0
95.20	363.0
96.00	941.0
96.80	325.0
98.00	14731.0
99.00	12465.0
99.90	895.0
100.90	6272.0
101.90	423.0
103.00	2194.0
104.00	4093.0
104.90	3595.0
106.00	1306.0
107.00	47448.0
108.00	7875.0
109.00	1360.0
110.00	88112.0
111.00	13908.0
112.00	1806.0
113.10	564.0
116.00	2358.0
117.00	41168.0
118.00	2854.0
119.00	458.0
119.90	607.0
120.90	270.0
121.90	2868.0
123.00	4394.0
124.00	2551.0
124.90	2142.0
127.00	162368.0
128.00	12135.0
129.00	62904.0
130.00	5276.0
130.90	1323.0
132.00	428.0
133.10	469.0
134.00	1799.0
135.00	5737.0
136.00	2118.0
137.00	3508.0

Instrument :
BNA_N
ClientSampleId :
DFTPP235

137.90	555.0
138.90	369.0
139.90	847.0
141.00	8607.0
141.90	2760.0
143.00	1610.0
144.00	411.0
144.90	365.0
146.00	1610.0
147.00	4545.0
148.00	10179.0
149.00	2486.0
150.00	759.0
150.90	1040.0
151.80	618.0
153.00	2886.0
154.00	1947.0
155.00	4930.0
156.00	6978.0
157.00	1321.0
158.00	1783.0
158.90	1292.0
159.90	2677.0
161.00	3721.0
161.90	1368.0
162.90	260.0
164.00	341.0
164.90	3211.0
166.00	2262.0
167.00	16187.0
168.00	12360.0
169.00	2103.0
169.90	687.0
170.90	591.0
172.00	1595.0
173.00	1779.0
174.00	3174.0
175.00	6297.0
176.00	1656.0
177.00	2996.0
178.00	1021.0
178.90	12112.0
180.00	10045.0
181.00	4131.0
182.00	478.0
182.90	375.0
184.00	1181.0
185.00	5683.0
186.00	46056.0
187.00	13898.0
188.00	1729.0
189.00	3035.0
189.90	422.0
191.00	1613.0
192.00	4534.0
193.00	5276.0
194.00	1034.0
195.00	321.0
196.00	9222.0
197.90	324352.0
199.00	22400.0
200.00	1776.0
201.50	1166.0
203.00	2523.0
204.00	12507.0
205.00	20760.0
206.00	87480.0
207.00	14602.0
208.00	3407.0
209.00	1567.0
210.30	1386.0
211.00	3638.0
212.90	203.0
214.90	1006.0
216.00	1674.0
216.90	22624.0
218.00	3022.0
218.90	184.0
221.00	11528.0
221.80	3115.0
223.00	5336.0
224.00	48592.0
225.00	11480.0
226.00	1890.0
227.00	18992.0
228.00	3103.0
229.00	3648.0
229.90	573.0
231.00	1558.0
231.90	220.0
233.00	435.0
234.00	1138.0
235.00	1395.0
236.00	1122.0
237.00	2657.0
237.90	354.0

Instrument :
BNA_N
ClientSampleId :
DFTPP235

238.90	911.0
240.00	766.0
240.90	1295.0
242.00	2715.0
243.10	2959.0
244.00	34808.0
245.10	4791.0
246.00	6845.0
247.00	1502.0
248.00	298.0
248.90	1099.0
249.90	228.0
251.00	410.0
251.80	266.0
253.00	668.0
255.00	184064.0
256.00	28168.0
257.00	2403.0
258.00	9948.0
259.00	1908.0
260.00	396.0
261.00	239.0
264.10	268.0
265.00	4083.0
265.90	565.0
266.90	195.0
270.00	155.0
270.80	450.0
272.10	601.0
273.00	5297.0
274.00	14287.0
275.00	82272.0
276.00	10931.0
277.00	6879.0
278.00	1428.0
279.10	206.0
280.90	309.0
282.00	233.0
283.00	836.0
284.00	545.0
285.00	1037.0
285.90	178.0
288.90	263.0
290.00	305.0
291.90	259.0
292.90	1896.0
294.00	508.0
296.00	20808.0
297.00	2985.0
298.10	187.0
300.90	315.0
302.10	342.0
303.00	2354.0
304.00	661.0
305.00	202.0
308.00	437.0
309.00	208.0
309.90	343.0
312.90	163.0
314.00	905.0
314.90	2226.0
316.00	1552.0
317.00	195.0
321.00	819.0
321.80	336.0
323.00	7004.0
324.00	1355.0
327.00	1246.0
328.00	682.0
332.10	419.0
333.00	788.0
334.00	4419.0
335.00	1105.0
336.00	284.0
340.90	712.0
342.00	236.0
346.00	1484.0
347.00	326.0
351.90	1889.0
353.00	1329.0
354.00	2005.0
355.00	494.0
365.00	8393.0
365.90	1564.0
370.10	211.0
370.90	532.0
372.00	3592.0
372.90	890.0
383.00	918.0
383.90	294.0
384.90	189.0
390.00	505.0
391.10	325.0
392.00	278.0
402.00	1287.0
403.00	2022.0

Instrument :
BNA_N
ClientSampleId :
DFTPP235

404.00	878.0
405.00	414.0
421.00	1439.0
422.00	1881.0
423.00	11334.0
424.10	4466.0
424.90	650.0
439.80	155.0
441.00	26488.0
442.00	149888.0
443.00	32008.0
444.10	3074.0

Instrument :
BNA_N
ClientSampleId :
DFTPP235

m/z	Abundance
36.00	192.0
37.00	814.0
38.00	2241.0
39.00	11561.0
40.00	1566.0
41.10	364.0
43.00	169.0
43.90	1225.0
45.00	327.0
47.00	168.0
49.00	1700.0
50.00	40624.0
51.00	161088.0
52.00	7428.0
53.00	285.0
54.90	697.0
56.00	4968.0
57.00	10584.0
58.00	499.0
59.00	165.0
61.00	2147.0
62.00	2446.0
63.00	6921.0
63.90	679.0
65.00	2996.0
66.00	234.0
67.00	273.0
68.00	2782.0
69.00	156032.0
70.00	949.0
73.00	1806.0
74.00	15732.0
75.00	25872.0
76.00	9282.0
77.00	185600.0
78.00	13190.0
79.00	12522.0
80.00	9202.0
81.00	13499.0
82.00	3398.0
83.00	2798.0
83.90	786.0
85.00	2818.0
86.00	4275.0
87.00	1559.0
88.00	766.0
88.90	324.0
91.00	3058.0
92.00	3876.0
93.00	21208.0
93.90	1235.0
94.90	353.0
96.00	1160.0
97.00	430.0
98.00	16888.0
99.00	14679.0
100.00	1221.0
101.00	7220.0
101.90	439.0
102.90	2145.0
104.00	4167.0
105.00	4668.0
106.00	1605.0
107.00	57304.0
108.00	8201.0
109.00	1800.0
110.00	101968.0
111.00	15594.0
112.00	2063.0
112.90	658.0
114.00	157.0
114.90	280.0
116.00	2869.0
117.00	48664.0
117.90	3765.0
118.90	570.0
120.00	833.0
120.90	257.0
122.00	4022.0
123.00	5723.0
123.90	2388.0
125.00	2299.0
127.00	194496.0
128.00	14957.0
129.00	74008.0
130.00	6166.0
131.00	1008.0
132.00	705.0
132.80	413.0
133.90	2092.0
135.00	6315.0
136.00	2903.0

Instrument :
BNA_N
ClientSampleId :
DFTPP235

137.00	3387.0
137.90	793.0
138.90	461.0
139.90	955.0
141.00	10358.0
142.00	3150.0
143.00	2612.0
143.90	717.0
144.90	533.0
146.00	1841.0
147.00	5603.0
148.00	11892.0
149.00	2972.0
149.90	721.0
150.90	1213.0
151.70	756.0
153.00	3074.0
154.00	2565.0
155.00	6015.0
156.00	8328.0
157.10	1951.0
158.00	2314.0
158.90	1336.0
160.00	3142.0
161.00	4550.0
162.00	1067.0
163.10	501.0
163.90	480.0
165.00	3976.0
166.00	3049.0
167.00	21320.0
168.00	13112.0
169.00	2018.0
170.00	843.0
171.00	944.0
172.00	1942.0
173.00	2365.0
174.00	4128.0
175.00	8395.0
176.00	2270.0
177.00	3994.0
177.90	1367.0
178.90	15046.0
180.00	11465.0
181.00	4766.0
182.10	1072.0
182.80	313.0
183.90	1304.0
185.00	7839.0
186.00	58336.0
187.00	17872.0
188.00	2382.0
188.90	3984.0
190.00	649.0
191.00	1986.0
192.00	5262.0
193.00	5445.0
194.10	1099.0
194.90	795.0
196.00	12048.0
198.00	416384.0
199.00	29352.0
200.00	2377.0
201.50	1605.0
203.00	2939.0
204.00	15429.0
205.00	28264.0
206.00	112640.0
207.00	17248.0
208.00	4199.0
209.00	1813.0
210.00	1670.0
211.00	4124.0
211.90	745.0
213.00	325.0
215.00	1385.0
216.00	2530.0
217.00	29808.0
218.00	4010.0
218.90	403.0
221.00	16243.0
221.70	3878.0
223.00	6733.0
224.00	62256.0
225.00	15612.0
226.00	2153.0
227.00	28112.0
227.90	3654.0
229.00	5445.0
229.90	760.0
231.00	2216.0
232.00	341.0
232.90	547.0
233.90	1844.0
235.00	1917.0
236.00	1517.0

Instrument :
BNA_N
ClientSampleId :
DFTPP235

237.00	3338.0
237.90	458.0
239.00	814.0
239.90	698.0
240.90	1584.0
242.00	3384.0
243.00	3787.0
244.00	47936.0
245.00	6285.0
245.90	9818.0
247.00	2082.0
247.90	517.0
249.00	1779.0
249.90	423.0
250.90	382.0
252.00	359.0
252.90	1095.0
254.00	1967.0
255.00	258752.0
256.00	38120.0
257.00	3352.0
258.00	13589.0
259.00	2178.0
260.10	501.0
261.00	397.0
263.90	331.0
265.00	6282.0
266.00	635.0
266.90	194.0
269.90	324.0
270.90	437.0
272.00	574.0
273.00	7209.0
274.00	20192.0
275.00	109384.0
276.00	15158.0
277.00	9490.0
278.00	1829.0
279.00	444.0
281.00	415.0
282.00	332.0
283.00	1068.0
283.90	591.0
285.00	1450.0
286.00	426.0
288.90	403.0
289.90	242.0
290.70	190.0
291.90	447.0
293.00	2309.0
294.00	674.0
296.00	30528.0
297.00	4344.0
297.90	392.0
299.10	191.0
300.90	485.0
302.00	499.0
303.00	3639.0
304.00	963.0
305.00	185.0
308.00	432.0
309.00	416.0
310.00	449.0
313.10	205.0
314.00	1770.0
315.00	3735.0
316.10	2166.0
317.00	376.0
321.00	1007.0
322.00	598.0
323.00	10109.0
324.00	1571.0
324.90	170.0
325.90	217.0
327.00	1921.0
328.00	973.0
328.90	179.0
331.90	519.0
332.90	925.0
334.00	6420.0
335.00	1597.0
336.10	216.0
340.00	157.0
341.00	1305.0
342.10	370.0
346.00	2402.0
346.90	412.0
352.00	3432.0
353.10	2049.0
354.00	3197.0
355.00	688.0
358.90	201.0
365.00	14784.0
366.00	2033.0
369.90	339.0
371.00	985.0

Instrument :
BNA_N
ClientSampleId :
DFTPP235

372.00	4996.0
373.00	1137.0
377.10	156.0
383.00	1600.0
383.90	613.0
385.00	158.0
389.90	598.0
391.00	586.0
392.00	340.0
400.90	304.0
402.00	2241.0
403.00	3766.0
403.90	995.0
404.80	225.0
421.00	2691.0
422.00	2290.0
423.00	17616.0
424.00	6353.0
425.00	987.0
441.00	40760.0
442.00	258944.0
443.00	49224.0
444.10	5034.0
444.90	327.0

Instrument :
BNA_N
ClientSampleId :
DFTPP235

m/z	Abundance
36.00	200.0
36.90	379.0
38.10	945.0
39.10	7017.0
39.90	1709.0
41.00	305.0
43.10	201.0
43.90	1033.0
44.90	201.0
48.00	190.0
49.00	1130.0
50.00	24960.0
51.00	99368.0
52.00	5439.0
53.00	198.0
54.90	484.0
56.00	3125.0
57.00	6509.0
57.90	269.0
59.00	165.0
61.00	1274.0
62.00	1217.0
63.00	3772.0
64.00	654.0
65.00	1902.0
67.00	253.0
68.00	1861.0
69.00	101344.0
70.00	410.0
71.00	194.0
73.00	1009.0
74.00	10982.0
75.00	15600.0
76.00	5790.0
77.00	117424.0
78.00	7871.0
79.00	7544.0
80.00	5585.0
81.00	8605.0
82.00	2010.0
83.00	1949.0
83.90	577.0
85.00	2096.0
86.00	2540.0
87.00	1174.0
88.00	497.0
89.00	153.0
91.00	2012.0
91.90	2178.0
93.00	13973.0
94.00	1087.0
95.00	211.0
96.00	930.0
96.90	304.0
98.00	11105.0
99.00	9335.0
99.90	752.0
101.00	5228.0
101.90	289.0
102.90	1561.0
104.00	2978.0
104.90	2650.0
106.00	970.0
107.00	38600.0
108.00	5234.0
109.10	1289.0
110.00	66472.0
111.00	9856.0
111.90	1485.0
113.00	523.0
115.00	236.0
116.00	1983.0
117.00	33320.0
118.00	2383.0
119.10	319.0
119.90	641.0
122.00	2404.0
123.00	3830.0
124.00	1836.0
125.00	1599.0
126.10	881.0
127.00	134016.0
128.00	9416.0
129.00	50160.0
130.00	3885.0
131.00	890.0
132.00	474.0
132.90	483.0
133.90	1503.0
134.90	4242.0
135.90	1549.0
137.00	2656.0

Instrument :
BNA_N
ClientSampleId :
DFTPP235

138.00	484.0
139.10	302.0
140.00	525.0
141.00	7428.0
142.00	2019.0
143.00	1682.0
144.00	430.0
144.80	326.0
145.90	1238.0
147.00	3472.0
147.90	8857.0
149.00	1790.0
149.90	504.0
151.00	781.0
151.60	537.0
153.00	1983.0
153.90	1577.0
155.00	3983.0
156.00	6031.0
157.00	1165.0
157.90	1512.0
159.00	996.0
159.90	2198.0
161.00	3400.0
161.90	1202.0
162.80	312.0
164.00	448.0
164.90	2547.0
166.00	1996.0
167.00	14615.0
168.00	9325.0
169.00	1448.0
170.10	506.0
171.00	611.0
171.90	1145.0
173.00	1701.0
174.00	3266.0
174.90	5368.0
176.00	1547.0
177.00	3010.0
178.00	1121.0
179.00	11351.0
180.00	7784.0
181.00	3903.0
182.00	539.0
182.90	267.0
184.00	1021.0
185.00	5669.0
186.00	41576.0
187.00	11348.0
187.90	1337.0
188.90	2499.0
189.90	436.0
191.00	1555.0
191.90	3650.0
193.00	4278.0
194.00	1020.0
195.10	441.0
196.00	8064.0
197.00	1834.0
197.90	308928.0
199.00	20432.0
200.00	1431.0
201.50	1284.0
203.00	2344.0
204.00	11284.0
205.00	18688.0
206.00	80896.0
207.00	13010.0
208.00	3184.0
208.90	1141.0
210.00	1272.0
211.00	2869.0
211.80	796.0
213.00	184.0
215.00	1084.0
216.00	1938.0
217.00	21824.0
217.90	2588.0
219.00	267.0
221.00	11338.0
221.80	3486.0
223.00	4820.0
224.00	47064.0
225.00	11976.0
226.10	1399.0
227.00	19344.0
228.00	2485.0
229.00	4181.0
229.90	532.0
231.00	2045.0
231.90	214.0
233.10	336.0
233.90	1184.0
235.00	1442.0
235.90	1157.0

Instrument :
BNA_N
ClientSampleId :
DFTPP235

237.00	1971.0
238.10	241.0
238.90	780.0
239.90	827.0
241.00	1061.0
242.00	2468.0
243.00	2149.0
244.00	37160.0
245.00	4363.0
246.00	7500.0
247.00	1710.0
247.90	446.0
249.00	1609.0
249.90	288.0
251.00	440.0
251.90	338.0
253.00	845.0
255.00	191360.0
256.00	26920.0
257.00	1891.0
258.00	11308.0
259.00	1787.0
259.90	263.0
260.90	349.0
263.90	309.0
265.00	4274.0
265.90	563.0
270.90	397.0
272.00	643.0
273.00	5978.0
274.00	15760.0
275.00	87664.0
276.00	10778.0
277.00	6860.0
278.00	1383.0
278.90	247.0
281.00	430.0
282.00	239.0
283.00	955.0
284.00	628.0
285.00	1418.0
286.20	178.0
288.90	256.0
290.00	385.0
290.80	172.0
292.00	295.0
293.00	1563.0
294.00	453.0
295.10	551.0
296.00	25040.0
297.00	3564.0
297.80	214.0
301.00	323.0
302.00	583.0
303.10	3177.0
304.10	808.0
305.00	183.0
308.00	270.0
308.90	279.0
310.00	354.0
313.00	232.0
313.90	1317.0
315.00	2728.0
316.00	1614.0
317.10	380.0
320.90	815.0
322.00	493.0
323.00	8613.0
323.90	1335.0
325.90	150.0
326.90	1482.0
328.00	919.0
328.90	178.0
332.00	634.0
333.00	870.0
334.00	4770.0
335.00	1501.0
335.90	260.0
341.00	1071.0
341.90	299.0
345.90	2007.0
346.90	458.0
352.00	2600.0
353.00	1592.0
354.00	2791.0
355.00	547.0
365.00	11596.0
365.90	1799.0
369.90	303.0
371.00	678.0
372.00	4087.0
373.10	1005.0
376.90	154.0
383.00	1254.0
383.90	365.0
389.90	516.0

Instrument :
BNA_N
ClientSampleId :
DFTPP235

390.90	503.0
392.00	267.0
401.10	342.0
402.00	1934.0
403.00	2777.0
404.00	1058.0
405.00	299.0
421.00	2744.0
422.00	2309.0
423.00	17768.0
424.00	4684.0
425.00	581.0
441.00	40680.0
442.00	267648.0
443.00	48296.0
444.00	4599.0
444.90	283.0

Instrument :
BNA_N
ClientSampleId :
DFTPP235