

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031025\
 Data File : BN036560.D
 Acq On : 10 Mar 2025 13:31
 Operator : RC/JU
 Sample : SSTDICC0.8
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8

Quant Time: Mar 10 16:01:54 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 10 15:54:23 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.724	152	2495	0.400	ng	0.00
7) Naphthalene-d8	10.509	136	5884	0.400	ng	0.00
13) Acenaphthene-d10	14.366	164	3456	0.400	ng	0.00
19) Phenanthrene-d10	17.111	188	6971	0.400	ng	0.00
29) Chrysene-d12	21.295	240	4636	0.400	ng	0.00
35) Perylene-d12	23.554	264	4198	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.312	112	4381	0.753	ng	0.00
5) Phenol-d6	6.894	99	5324	0.741	ng	0.00
8) Nitrobenzene-d5	8.875	82	4717	0.737	ng	0.00
11) 2-Methylnaphthalene-d10	12.106	152	6616	0.756	ng	0.00
14) 2,4,6-Tribromophenol	15.858	330	1166	0.744	ng	0.00
15) 2-Fluorobiphenyl	12.988	172	16243	0.808	ng	0.00
27) Fluoranthene-d10	19.141	212	13330	0.746	ng	0.00
31) Terphenyl-d14	19.745	244	8571	0.772	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.239	88	2251	0.813	ng	99
3) n-Nitrosodimethylamine	3.550	42	4197	0.750	ng	96
6) bis(2-Chloroethyl)ether	7.147	93	5632	0.759	ng	99
9) Naphthalene	10.562	128	13078	0.756	ng	98
10) Hexachlorobutadiene	10.851	225	3147	0.772	ng	# 99
12) 2-Methylnaphthalene	12.177	142	8272	0.751	ng	98
16) Acenaphthylene	14.078	152	12403	0.760	ng	100
17) Acenaphthene	14.430	154	8096	0.758	ng	99
18) Fluorene	15.414	166	11120	0.770	ng	100
20) 4,6-Dinitro-2-methylph...	15.499	198	1039	0.724	ng	# 59
21) 4-Bromophenyl-phenylether	16.305	248	3324	0.761	ng	93
22) Hexachlorobenzene	16.416	284	4115	0.780	ng	99
23) Atrazine	16.578	200	2677	0.764	ng	94
24) Pentachlorophenol	16.764	266	1701	0.707	ng	98
25) Phenanthrene	17.149	178	15910	0.761	ng	99
26) Anthracene	17.236	178	14403	0.763	ng	99
28) Fluoranthene	19.174	202	17738	0.755	ng	99
30) Pyrene	19.536	202	17714	0.781	ng	100
32) Benzo(a)anthracene	21.277	228	12089	0.750	ng	98
33) Chrysene	21.331	228	13974	0.793	ng	100
34) Bis(2-ethylhexyl)phtha...	21.214	149	8021	0.699	ng	# 99
36) Indeno(1,2,3-cd)pyrene	25.838	276	11785	0.778	ng	99
37) Benzo(b)fluoranthene	22.873	252	11771	0.770	ng	# 91
38) Benzo(k)fluoranthene	22.917	252	12432	0.776	ng	# 90
39) Benzo(a)pyrene	23.455	252	10036	0.780	ng	# 87
40) Dibenzo(a,h)anthracene	25.858	278	9450	0.801	ng	91
41) Benzo(g,h,i)perylene	26.534	276	10494	0.778	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031025\
Data File : BN036560.D
Acq On : 10 Mar 2025 13:31
Operator : RC/JU
Sample : SSTDICC0.8
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Quant Time: Mar 10 16:01:54 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:54:23 2025
Response via : Initial Calibration








#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.724 min Scan# 64
 Delta R.T. 0.000 min
 Lab File: BN036560.D
 Acq: 10 Mar 2025 13:31

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8







Tgt Ion:152 Resp: 2495
 Ion Ratio Lower Upper
 152 100
 150 153.3 123.7 185.5
 115 65.1 54.3 81.5



#2
 1,4-Dioxane
 Concen: 0.813 ng
 RT: 3.239 min Scan# 22
 Delta R.T. 0.000 min
 Lab File: BN036560.D
 Acq: 10 Mar 2025 13:31

Tgt Ion: 88 Resp: 2251
 Ion Ratio Lower Upper
 88 100
 43 45.6 37.8 56.8
 58 84.5 67.4 101.2





#3

n-Nitrosodimethylamine

Concen: 0.750 ng

RT: 3.550 min Scan# 61

Delta R.T. 0.000 min

Lab File: BN036560.D

Acq: 10 Mar 2025 13:31

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion: 42 Resp: 4197

Ion Ratio Lower Upper

42 100

74 79.3 60.6 90.8

44 7.4 6.3 9.5



#4

2-Fluorophenol

Concen: 0.753 ng

RT: 5.312 min Scan# 309

Delta R.T. 0.000 min

Lab File: BN036560.D

Acq: 10 Mar 2025 13:31




Tgt Ion: 112 Resp: 4381


Ion Ratio Lower Upper

112 100

64 68.8 53.1 79.7

63 40.4 31.8 47.8









#5
Phenol-d6
Concen: 0.741 ng
RT: 6.894 min Scan# 51
Delta R.T. -0.007 min
Lab File: BN036560.D
Acq: 10 Mar 2025 13:31




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion: 99 Resp: 5324
Ion Ratio Lower Upper
99 100
42 31.3 26.5 39.7
71 42.6 34.1 51.1



#6
bis(2-Chloroethyl)ether
Concen: 0.759 ng
RT: 7.147 min Scan# 563
Delta R.T. -0.007 min
Lab File: BN036560.D
Acq: 10 Mar 2025 13:31

Tgt Ion: 93 Resp: 5632
Ion Ratio Lower Upper
93 100
63 84.0 67.7 101.5
95 31.6 25.6 38.4









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.509 min Scan# 930
Delta R.T. 0.000 min
Lab File: BN036560.D
Acq: 10 Mar 2025 13:31




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:136 Resp: 5884
Ion Ratio Lower Upper
136 100
137 11.9 10.3 15.5
54 12.4 11.5 17.3
68 7.5 7.0 10.4



#8
Nitrobenzene-d5
Concen: 0.737 ng
RT: 8.875 min Scan# 777
Delta R.T. 0.000 min
Lab File: BN036560.D
Acq: 10 Mar 2025 13:31

Tgt Ion: 82 Resp: 4717
Ion Ratio Lower Upper
82 100
128 37.9 30.6 45.8
54 63.2 52.2 78.4









#9
Naphthalene
Concen: 0.756 ng
RT: 10.562 min Scan# 935
Delta R.T. 0.000 min
Lab File: BN036560.D
Acq: 10 Mar 2025 13:31




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:128 Resp: 13078
Ion Ratio Lower Upper
128 100
129 11.1 9.8 14.6
127 13.9 11.8 17.8



#10
Hexachlorobutadiene
Concen: 0.772 ng
RT: 10.851 min Scan# 962
Delta R.T. 0.000 min
Lab File: BN036560.D
Acq: 10 Mar 2025 13:31

Tgt Ion:225 Resp: 3147
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.7 51.8 77.8









#11
2-Methylnaphthalene-d10
Concen: 0.756 ng
RT: 12.106 min Scan# 1141
Delta R.T. -0.005 min
Lab File: BN036560.D
Acq: 10 Mar 2025 13:31




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:152 Resp: 6616
Ion Ratio Lower Upper
152 100
151 20.8 17.0 25.6



#12
2-Methylnaphthalene
Concen: 0.751 ng
RT: 12.177 min Scan# 1155
Delta R.T. -0.005 min
Lab File: BN036560.D
Acq: 10 Mar 2025 13:31

Tgt Ion:142 Resp: 8272
Ion Ratio Lower Upper
142 100
141 91.0 71.7 107.5
115 46.2 38.3 57.5









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.366 min Scan# 1446
Delta R.T. 0.000 min
Lab File: BN036560.D
Acq: 10 Mar 2025 13:31




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:164 Resp: 3456
Ion Ratio Lower Upper
164 100
162 103.6 84.2 126.2
160 52.1 42.2 63.2



#14
2,4,6-Tribromophenol
Concen: 0.744 ng
RT: 15.858 min Scan# 1581
Delta R.T. 0.000 min
Lab File: BN036560.D
Acq: 10 Mar 2025 13:31

Tgt Ion:330 Resp: 1166
Ion Ratio Lower Upper
330 100
332 97.6 75.2 112.8
141 51.8 43.4 65.2









#15
2-Fluorobiphenyl
Concen: 0.808 ng
RT: 12.988 min Scan# 1315
Delta R.T. 0.000 min
Lab File: BN036560.D
Acq: 10 Mar 2025 13:31




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:172 Resp: 16243
Ion Ratio Lower Upper
172 100
171 36.2 29.5 44.3
170 24.3 20.2 30.4



#16
Acenaphthylene
Concen: 0.760 ng
RT: 14.078 min Scan# 1419
Delta R.T. 0.000 min
Lab File: BN036560.D
Acq: 10 Mar 2025 13:31

Tgt Ion:152 Resp: 12403
Ion Ratio Lower Upper
152 100
151 20.1 16.2 24.4
153 13.1 10.6 15.8





#17

Acenaphthene

Concen: 0.758 ng

RT: 14.430 min Scan# 1452

Delta R.T. 0.000 min

Lab File: BN036560.D

Acq: 10 Mar 2025 13:31

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:154 Resp: 8096

Ion Ratio Lower Upper

154 100

153 117.8 94.1 141.1

152 60.9 49.8 74.6



#18

Fluorene

Concen: 0.770 ng

RT: 15.414 min Scan# 1544

Delta R.T. 0.000 min

Lab File: BN036560.D

Acq: 10 Mar 2025 13:31




Tgt Ion:166 Resp: 11120


Ion Ratio Lower Upper

166 100

165 99.8 79.8 119.8

167 13.3 10.6 15.8









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.111 min Scan# 1682
Delta R.T. 0.000 min
Lab File: BN036560.D
Acq: 10 Mar 2025 13:31




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:188 Resp: 6971
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 10.3 8.8 13.2



#20
4,6-Dinitro-2-methylphenol
Concen: 0.724 ng
RT: 15.499 min Scan# 1552
Delta R.T. 0.000 min
Lab File: BN036560.D
Acq: 10 Mar 2025 13:31

Tgt Ion:198 Resp: 1039
Ion Ratio Lower Upper
198 100
51 76.5 107.9 161.9#
105 48.1 56.2 84.2#









#21
4-Bromophenyl-phenylether
Concen: 0.761 ng
RT: 16.305 min Scan# 1617
Delta R.T. 0.000 min
Lab File: BN036560.D
Acq: 10 Mar 2025 13:31




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:248 Resp: 3324
Ion Ratio Lower Upper
248 100
250 96.2 73.0 109.6
141 77.7 68.6 103.0



#22
Hexachlorobenzene
Concen: 0.780 ng
RT: 16.416 min Scan# 1626
Delta R.T. 0.000 min
Lab File: BN036560.D
Acq: 10 Mar 2025 13:31

Tgt Ion:284 Resp: 4115
Ion Ratio Lower Upper
284 100
142 47.2 37.0 55.4
249 34.7 28.1 42.1









#23
Atrazine
Concen: 0.764 ng
RT: 16.578 min Scan# 1639
Delta R.T. 0.000 min
Lab File: BN036560.D
Acq: 10 Mar 2025 13:31




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:200 Resp: 2677
Ion Ratio Lower Upper
200 100
173 28.6 27.3 40.9
215 43.8 36.8 55.2



#24
Pentachlorophenol
Concen: 0.707 ng
RT: 16.764 min Scan# 1654
Delta R.T. -0.012 min
Lab File: BN036560.D
Acq: 10 Mar 2025 13:31

Tgt Ion:266 Resp: 1701
Ion Ratio Lower Upper
266 100
264 63.5 49.6 74.4
268 65.1 50.9 76.3





#25

Phenanthrene

Concen: 0.761 ng

RT: 17.149 min Scan# 1685

Delta R.T. 0.000 min

Lab File: BN036560.D

Acq: 10 Mar 2025 13:31

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:178 Resp: 15910

Ion Ratio Lower Upper

178 100

176 19.6 15.9 23.9

179 15.1 12.2 18.4



#26

Anthracene

Concen: 0.763 ng

RT: 17.236 min Scan# 1692

Delta R.T. -0.012 min

Lab File: BN036560.D

Acq: 10 Mar 2025 13:31




Tgt Ion:178 Resp: 14403


Ion Ratio Lower Upper

178 100

176 18.8 15.4 23.2

179 15.1 12.6 18.8





#27

Fluoranthene-d10

Concen: 0.746 ng

RT: 19.141 min Scan# 1930

Delta R.T. 0.000 min

Lab File: BN036560.D

Acq: 10 Mar 2025 13:31

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:212 Resp: 13330

Ion Ratio Lower Upper

212 100

106 15.2 11.8 17.6

104 8.9 7.3 10.9



#28

Fluoranthene

Concen: 0.755 ng

RT: 19.174 min Scan# 1937

Delta R.T. 0.000 min

Lab File: BN036560.D

Acq: 10 Mar 2025 13:31




Tgt Ion:202 Resp: 17738


Ion Ratio Lower Upper

202 100

101 12.1 9.4 14.0

203 17.0 13.5 20.3





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.295 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN036560.D

Acq: 10 Mar 2025 13:31

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:240 Resp: 4636

Ion Ratio Lower Upper

240 100

120 17.7 14.6 22.0

236 30.5 24.1 36.1



#30

Pyrene

Concen: 0.781 ng

RT: 19.536 min Scan# 2015

Delta R.T. 0.000 min

Lab File: BN036560.D

Acq: 10 Mar 2025 13:31




Tgt Ion:202 Resp: 17714


Ion Ratio Lower Upper

202 100

200 21.4 17.1 25.7

203 17.5 14.1 21.1









#31
 Terphenyl-d14
 Concen: 0.772 ng
 RT: 19.745 min Scan# 2060
 Delta R.T. 0.000 min
 Lab File: BN036560.D
 Acq: 10 Mar 2025 13:31




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8


Tgt Ion:244	Resp:	8571
Ion Ratio	Lower	Upper
244	100	
212	10.1	9.6 14.4
122	15.8	13.9 20.9



#32
 Benzo(a)anthracene
 Concen: 0.750 ng
 RT: 21.277 min Scan# 2258
 Delta R.T. -0.009 min
 Lab File: BN036560.D
 Acq: 10 Mar 2025 13:31

Tgt Ion:228	Resp:	12089
Ion Ratio	Lower	Upper
228	100	
226	29.3	22.5 33.7
229	20.1	16.6 25.0









#33
Chrysene
Concen: 0.793 ng
RT: 21.331 min Scan# 2264
Delta R.T. 0.000 min
Lab File: BN036560.D
Acq: 10 Mar 2025 13:31




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:228 Resp: 13974
Ion Ratio Lower Upper
228 100
226 31.5 25.3 37.9
229 19.7 15.8 23.8



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.699 ng
RT: 21.214 min Scan# 2251
Delta R.T. 0.000 min
Lab File: BN036560.D
Acq: 10 Mar 2025 13:31

Tgt Ion:149 Resp: 8021
Ion Ratio Lower Upper
149 100
167 25.9 20.7 31.1
279 3.2 3.6 5.4#









#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.554 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: BN036560.D
 Acq: 10 Mar 2025 13:31




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8


Tgt Ion:	264	Resp:	4198
Ion Ratio	Lower	Upper	
264	100		
260	27.0	22.6	33.8
265	91.3	88.1	132.1



#36
 Indeno(1,2,3-cd)pyrene
 Concen: 0.778 ng
 RT: 25.838 min Scan# 3558
 Delta R.T. 0.000 min
 Lab File: BN036560.D
 Acq: 10 Mar 2025 13:31

Tgt Ion:	276	Resp:	11785
Ion Ratio	Lower	Upper	
276	100		
138	29.6	23.4	35.2
277	24.7	20.0	30.0









#37
Benzo(b)fluoranthene
Concen: 0.770 ng
RT: 22.873 min Scan# 2544
Delta R.T. 0.000 min
Lab File: BN036560.D
Acq: 10 Mar 2025 13:31




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:252 Resp: 11771
Ion Ratio Lower Upper
252 100
253 25.5 23.9 35.9
125 16.7 17.4 26.2#



#38
Benzo(k)fluoranthene
Concen: 0.776 ng
RT: 22.917 min Scan# 2559
Delta R.T. 0.000 min
Lab File: BN036560.D
Acq: 10 Mar 2025 13:31

Tgt Ion:252 Resp: 12432
Ion Ratio Lower Upper
252 100
253 25.9 24.6 36.8
125 16.2 17.8 26.8#





#39

Benzo(a)pyrene

Concen: 0.780 ng

RT: 23.455 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN036560.D


Acq: 10 Mar 2025 13:31

Instrument :




BNA_N

ClientSampleId :

SSTDICC0.8



Tgt Ion:	252	Resp:	10036
Ion Ratio	Lower	Upper	
252	100		
253	27.8	27.8	41.8#
125	20.7	22.7	34.1#



#40

Dibenzo(a,h)anthracene

Concen: 0.801 ng




RT: 25.858 min Scan# 3565


Delta R.T. 0.003 min

Lab File: BN036560.D

Acq: 10 Mar 2025 13:31


Tgt Ion:	278	Resp:	9450
Ion Ratio	Lower	Upper	
278	100		
139	22.3	20.8	31.2
279	29.6	28.8	43.2





#41
Benzo(g,h,i)perylene
Concen: 0.778 ng
RT: 26.534 min Scan# 31
Delta R.T. 0.003 min
Lab File: BN036560.D
Acq: 10 Mar 2025 13:31

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8



Tgt Ion:276 Resp: 10494
Ion Ratio Lower Upper
276 100
277 24.4 22.2 33.4
138 29.1 24.1 36.1

