

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031025\
 Data File : BN036561.D
 Acq On : 10 Mar 2025 14:07
 Operator : RC/JU
 Sample : SSTDICC1.6
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6

Quant Time: Mar 10 16:02:21 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 10 15:54:23 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.724	152	2537	0.400	ng	0.00
7) Naphthalene-d8	10.509	136	6200	0.400	ng	0.00
13) Acenaphthene-d10	14.366	164	3827	0.400	ng	0.00
19) Phenanthrene-d10	17.111	188	8149	0.400	ng	0.00
29) Chrysene-d12	21.295	240	5977	0.400	ng	# 0.00
35) Perylene-d12	23.552	264	5048	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.312	112	9276	1.569	ng	0.00
5) Phenol-d6	6.894	99	11493	1.574	ng	0.00
8) Nitrobenzene-d5	8.875	82	9959	1.477	ng	0.00
11) 2-Methylnaphthalene-d10	12.101	152	14319	1.553	ng	0.00
14) 2,4,6-Tribromophenol	15.858	330	2872	1.654	ng	0.00
15) 2-Fluorobiphenyl	12.983	172	36192	1.626	ng	0.00
27) Fluoranthene-d10	19.141	212	33414	1.600	ng	0.00
31) Terphenyl-d14	19.740	244	21872	1.527	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.239	88	4464	1.586	ng	98
3) n-Nitrosodimethylamine	3.550	42	8625	1.515	ng	# 96
6) bis(2-Chloroethyl)ether	7.147	93	11485	1.521	ng	99
9) Naphthalene	10.562	128	27473	1.506	ng	97
10) Hexachlorobutadiene	10.851	225	6466	1.506	ng	# 99
12) 2-Methylnaphthalene	12.177	142	18206	1.569	ng	98
16) Acenaphthylene	14.078	152	28080	1.555	ng	100
17) Acenaphthene	14.420	154	18355	1.553	ng	98
18) Fluorene	15.414	166	25565	1.599	ng	99
20) 4,6-Dinitro-2-methylph...	15.489	198	2879	1.488	ng	# 64
21) 4-Bromophenyl-phenylether	16.305	248	7859	1.539	ng	# 85
22) Hexachlorobenzene	16.416	284	9216	1.495	ng	100
23) Atrazine	16.578	200	6530	1.595	ng	# 91
24) Pentachlorophenol	16.764	266	4395	1.563	ng	99
25) Phenanthrene	17.149	178	37989	1.554	ng	99
26) Anthracene	17.235	178	35054	1.589	ng	99
28) Fluoranthene	19.169	202	44451	1.619	ng	99
30) Pyrene	19.531	202	44705	1.530	ng	100
32) Benzo(a)anthracene	21.277	228	32205	1.550	ng	98
33) Chrysene	21.331	228	34953	1.539	ng	99
34) Bis(2-ethylhexyl)phtha...	21.214	149	22621	1.529	ng	# 99
36) Indeno(1,2,3-cd)pyrene	25.832	276	28605	1.570	ng	99
37) Benzo(b)fluoranthene	22.870	252	29819	1.623	ng	# 86
38) Benzo(k)fluoranthene	22.917	252	30710	1.593	ng	# 85
39) Benzo(a)pyrene	23.452	252	24696	1.596	ng	# 79
40) Dibenzo(a,h)anthracene	25.850	278	22248	1.569	ng	# 85
41) Benzo(g,h,i)perylene	26.531	276	24906	1.535	ng	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031025\
Data File : BN036561.D
Acq On : 10 Mar 2025 14:07
Operator : RC/JU
Sample : SSTDICC1.6
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Quant Time: Mar 10 16:02:21 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:54:23 2025
Response via : Initial Calibration








#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.724 min Scan# 64
 Delta R.T. 0.000 min
 Lab File: BN036561.D
 Acq: 10 Mar 2025 14:07

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6







Tgt Ion:152 Resp: 2537
 Ion Ratio Lower Upper
 152 100
 150 155.4 123.7 185.5
 115 66.4 54.3 81.5



#2
 1,4-Dioxane
 Concen: 1.586 ng
 RT: 3.239 min Scan# 22
 Delta R.T. 0.000 min
 Lab File: BN036561.D
 Acq: 10 Mar 2025 14:07

Tgt Ion: 88 Resp: 4464
 Ion Ratio Lower Upper
 88 100
 43 45.1 37.8 56.8
 58 85.6 67.4 101.2









#3
n-Nitrosodimethylamine
Concen: 1.515 ng
RT: 3.550 min Scan# 61
Delta R.T. 0.000 min
Lab File: BN036561.D
Acq: 10 Mar 2025 14:07




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion: 42 Resp: 8625
Ion Ratio Lower Upper
42 100
74 79.2 60.6 90.8
44 5.8 6.3 9.5#



#4
2-Fluorophenol
Concen: 1.569 ng
RT: 5.312 min Scan# 309
Delta R.T. 0.000 min
Lab File: BN036561.D
Acq: 10 Mar 2025 14:07

Tgt Ion: 112 Resp: 9276
Ion Ratio Lower Upper
112 100
64 68.3 53.1 79.7
63 40.1 31.8 47.8









#5
Phenol-d6
Concen: 1.574 ng
RT: 6.894 min Scan# 51
Delta R.T. -0.007 min
Lab File: BN036561.D
Acq: 10 Mar 2025 14:07




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion: 99 Resp: 11493
Ion Ratio Lower Upper
99 100
42 32.6 26.5 39.7
71 43.3 34.1 51.1



#6
bis(2-Chloroethyl)ether
Concen: 1.521 ng
RT: 7.147 min Scan# 563
Delta R.T. -0.007 min
Lab File: BN036561.D
Acq: 10 Mar 2025 14:07

Tgt Ion: 93 Resp: 11485
Ion Ratio Lower Upper
93 100
63 85.8 67.7 101.5
95 31.8 25.6 38.4









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.509 min Scan# 930
Delta R.T. 0.000 min
Lab File: BN036561.D
Acq: 10 Mar 2025 14:07




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:	136	Resp:	6200
Ion Ratio	Lower	Upper	
136	100		
137	11.9	10.3	15.5
54	12.3	11.5	17.3
68	7.7	7.0	10.4



#8
Nitrobenzene-d5
Concen: 1.477 ng
RT: 8.875 min Scan# 777
Delta R.T. 0.000 min
Lab File: BN036561.D
Acq: 10 Mar 2025 14:07

Tgt Ion:	82	Resp:	9959
Ion Ratio	Lower	Upper	
82	100		
128	39.4	30.6	45.8
54	60.4	52.2	78.4









#9
Naphthalene
Concen: 1.506 ng
RT: 10.562 min Scan# 935
Delta R.T. 0.000 min
Lab File: BN036561.D
Acq: 10 Mar 2025 14:07




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:128 Resp: 27473
Ion Ratio Lower Upper
128 100
129 11.1 9.8 14.6
127 13.5 11.8 17.8



#10
Hexachlorobutadiene
Concen: 1.506 ng
RT: 10.851 min Scan# 962
Delta R.T. 0.000 min
Lab File: BN036561.D
Acq: 10 Mar 2025 14:07

Tgt Ion:225 Resp: 6466
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.8 51.8 77.8









#11
2-Methylnaphthalene-d10
Concen: 1.553 ng
RT: 12.101 min Scan# 1142
Delta R.T. -0.010 min
Lab File: BN036561.D
Acq: 10 Mar 2025 14:07




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:152 Resp: 14319
Ion Ratio Lower Upper
152 100
151 21.3 17.0 25.6



#12
2-Methylnaphthalene
Concen: 1.569 ng
RT: 12.177 min Scan# 1155
Delta R.T. -0.005 min
Lab File: BN036561.D
Acq: 10 Mar 2025 14:07

Tgt Ion:142 Resp: 18206
Ion Ratio Lower Upper
142 100
141 89.7 71.7 107.5
115 44.7 38.3 57.5





#13

Acenaphthene-d10

Concen: 0.400 ng

RT: 14.366 min Scan# 1446

Delta R.T. 0.000 min

Lab File: BN036561.D

Acq: 10 Mar 2025 14:07

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:164 Resp: 3827

Ion Ratio Lower Upper

164 100

162 102.9 84.2 126.2

160 52.1 42.2 63.2



#14

2,4,6-Tribromophenol

Concen: 1.654 ng

RT: 15.858 min Scan# 1581

Delta R.T. 0.000 min

Lab File: BN036561.D

Acq: 10 Mar 2025 14:07




Tgt Ion:330 Resp: 2872


Ion Ratio Lower Upper

330 100

332 96.3 75.2 112.8

141 53.8 43.4 65.2





#15

2-Fluorobiphenyl

Concen: 1.626 ng

RT: 12.983 min Scan# 1419

Delta R.T. -0.005 min

Lab File: BN036561.D

Acq: 10 Mar 2025 14:07

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:172 Resp: 36192

Ion Ratio Lower Upper

172 100

171 36.5 29.5 44.3

170 24.9 20.2 30.4



#16

Acenaphthylene

Concen: 1.555 ng

RT: 14.078 min Scan# 1419

Delta R.T. 0.000 min

Lab File: BN036561.D

Acq: 10 Mar 2025 14:07




Tgt Ion:152 Resp: 28080


Ion Ratio Lower Upper

152 100

151 20.3 16.2 24.4

153 13.0 10.6 15.8









#17
Acenaphthene
Concen: 1.553 ng
RT: 14.420 min Scan# 1451
Delta R.T. -0.011 min
Lab File: BN036561.D
Acq: 10 Mar 2025 14:07




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:154 Resp: 18355
Ion Ratio Lower Upper
154 100
153 115.7 94.1 141.1
152 60.3 49.8 74.6



#18
Fluorene
Concen: 1.599 ng
RT: 15.414 min Scan# 1544
Delta R.T. 0.000 min
Lab File: BN036561.D
Acq: 10 Mar 2025 14:07

Tgt Ion:166 Resp: 25565
Ion Ratio Lower Upper
166 100
165 100.5 79.8 119.8
167 13.1 10.6 15.8









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.111 min Scan# 1
Delta R.T. 0.000 min
Lab File: BN036561.D
Acq: 10 Mar 2025 14:07




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:188 Resp: 8149
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 8.8 8.8 13.2



#20
4,6-Dinitro-2-methylphenol
Concen: 1.488 ng
RT: 15.489 min Scan# 1551
Delta R.T. -0.010 min
Lab File: BN036561.D
Acq: 10 Mar 2025 14:07

Tgt Ion:198 Resp: 2879
Ion Ratio Lower Upper
198 100
51 81.2 107.9 161.9#
105 55.5 56.2 84.2#









#21
4-Bromophenyl-phenylether
Concen: 1.539 ng
RT: 16.305 min Scan# 1617
Delta R.T. 0.000 min
Lab File: BN036561.D
Acq: 10 Mar 2025 14:07




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:248 Resp: 7859
Ion Ratio Lower Upper
248 100
250 97.5 73.0 109.6
141 63.7 68.6 103.0#



#22
Hexachlorobenzene
Concen: 1.495 ng
RT: 16.416 min Scan# 1626
Delta R.T. 0.000 min
Lab File: BN036561.D
Acq: 10 Mar 2025 14:07

Tgt Ion:284 Resp: 9216
Ion Ratio Lower Upper
284 100
142 46.6 37.0 55.4
249 35.1 28.1 42.1









#23
Atrazine
Concen: 1.595 ng
RT: 16.578 min Scan# 1639
Delta R.T. 0.000 min
Lab File: BN036561.D
Acq: 10 Mar 2025 14:07




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:200 Resp: 6530
Ion Ratio Lower Upper
200 100
173 24.6 27.3 40.9#
215 43.5 36.8 55.2



#24
Pentachlorophenol
Concen: 1.563 ng
RT: 16.764 min Scan# 1654
Delta R.T. -0.012 min
Lab File: BN036561.D
Acq: 10 Mar 2025 14:07

Tgt Ion:266 Resp: 4395
Ion Ratio Lower Upper
266 100
264 63.5 49.6 74.4
268 64.4 50.9 76.3









#25
 Phenanthrene
 Concen: 1.554 ng
 RT: 17.149 min Scan# 1685
 Delta R.T. 0.000 min
 Lab File: BN036561.D
 Acq: 10 Mar 2025 14:07

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6

Tgt Ion:178 Resp: 37989




Ion	Ratio	Lower	Upper
178	100		
176	19.7	15.9	23.9
179	15.1	12.2	18.4




#26
 Anthracene
 Concen: 1.589 ng
 RT: 17.235 min Scan# 1692
 Delta R.T. -0.012 min
 Lab File: BN036561.D
 Acq: 10 Mar 2025 14:07

Tgt Ion:178 Resp: 35054

Ion	Ratio	Lower	Upper
178	100		
176	19.0	15.4	23.2
179	15.3	12.6	18.8









#27
Fluoranthene-d10
Concen: 1.600 ng
RT: 19.141 min Scan# 1930
Delta R.T. 0.000 min
Lab File: BN036561.D
Acq: 10 Mar 2025 14:07




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:212 Resp: 33414
Ion Ratio Lower Upper
212 100
106 15.0 11.8 17.6
104 9.1 7.3 10.9



#28
Fluoranthene
Concen: 1.619 ng
RT: 19.169 min Scan# 1936
Delta R.T. -0.004 min
Lab File: BN036561.D
Acq: 10 Mar 2025 14:07

Tgt Ion:202 Resp: 44451
Ion Ratio Lower Upper
202 100
101 12.0 9.4 14.0
203 17.2 13.5 20.3





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.295 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN036561.D

Acq: 10 Mar 2025 14:07

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:240 Resp: 5977

Ion Ratio Lower Upper

240 100

120 14.0 14.6 22.0#

236 28.9 24.1 36.1



#30

Pyrene

Concen: 1.530 ng

RT: 19.531 min Scan# 2014

Delta R.T. -0.004 min

Lab File: BN036561.D

Acq: 10 Mar 2025 14:07




Tgt Ion:202 Resp: 44705


Ion Ratio Lower Upper

202 100

200 21.3 17.1 25.7

203 17.5 14.1 21.1









#31
 Terphenyl-d14
 Concen: 1.527 ng
 RT: 19.740 min Scan# 2060
 Delta R.T. -0.004 min
 Lab File: BN036561.D
 Acq: 10 Mar 2025 14:07




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6


Tgt Ion:244 Resp: 21872
 Ion Ratio Lower Upper
 244 100
 212 10.3 9.6 14.4
 122 17.0 13.9 20.9



#32
 Benzo(a)anthracene
 Concen: 1.550 ng
 RT: 21.277 min Scan# 2258
 Delta R.T. -0.009 min
 Lab File: BN036561.D
 Acq: 10 Mar 2025 14:07

Tgt Ion:228 Resp: 32205
 Ion Ratio Lower Upper
 228 100
 226 28.0 22.5 33.7
 229 19.2 16.6 25.0





#33

Chrysene

Concen: 1.539 ng

RT: 21.331 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN036561.D

Acq: 10 Mar 2025 14:07

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:228 Resp: 34953

Ion Ratio Lower Upper

228 100

226 30.6 25.3 37.9

229 19.3 15.8 23.8



#34

Bis(2-ethylhexyl)phthalate

Concen: 1.529 ng

RT: 21.214 min Scan# 2251

Delta R.T. 0.000 min

Lab File: BN036561.D

Acq: 10 Mar 2025 14:07




Tgt Ion:149 Resp: 22621


Ion Ratio Lower Upper

149 100

167 26.3 20.7 31.1

279 2.7 3.6 5.4#





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.552 min Scan# 21

Delta R.T. -0.003 min

Lab File: BN036561.D

Acq: 10 Mar 2025 14:07

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:264 Resp: 5048

Ion Ratio Lower Upper

264 100

260 27.2 22.6 33.8

265 85.7 88.1 132.1#



#36

Indeno(1,2,3-cd)pyrene

Concen: 1.570 ng

RT: 25.832 min Scan# 3556

Delta R.T. -0.006 min

Lab File: BN036561.D

Acq: 10 Mar 2025 14:07




Tgt Ion:276 Resp: 28605


Ion Ratio Lower Upper

276 100

138 30.4 23.4 35.2

277 24.9 20.0 30.0





#37

Benzo(b)fluoranthene

Concen: 1.623 ng

RT: 22.870 min Scan# 2544

Delta R.T. -0.003 min

Lab File: BN036561.D


Acq: 10 Mar 2025 14:07

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6






Tgt Ion:252 Resp: 29819

Ion Ratio Lower Upper

252 100

253 23.3 23.9 35.9#

125 14.0 17.4 26.2#



#38

Benzo(k)fluoranthene

Concen: 1.593 ng

RT: 22.917 min Scan# 2559

Delta R.T. 0.000 min

Lab File: BN036561.D

Acq: 10 Mar 2025 14:07




Tgt Ion:252 Resp: 30710


Ion Ratio Lower Upper

252 100

253 23.0 24.6 36.8#

125 14.1 17.8 26.8#





#39

Benzo(a)pyrene

Concen: 1.596 ng

RT: 23.452 min Scan# 21

Delta R.T. -0.003 min

Lab File: BN036561.D

Acq: 10 Mar 2025 14:07

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:252 Resp: 24696

Ion Ratio Lower Upper

252 100

253 23.9 27.8 41.8#

125 16.3 22.7 34.1#



#40

Dibenzo(a,h)anthracene

Concen: 1.569 ng

RT: 25.850 min Scan# 3562

Delta R.T. -0.006 min

Lab File: BN036561.D

Acq: 10 Mar 2025 14:07




Tgt Ion:278 Resp: 22248


Ion Ratio Lower Upper

278 100

139 21.1 20.8 31.2

279 25.1 28.8 43.2#





#41
 Benzo(g,h,i)perylene
 Concen: 1.535 ng
 RT: 26.531 min Scan# 31
 Delta R.T. 0.000 min
 Lab File: BN036561.D
 Acq: 10 Mar 2025 14:07

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6

Tgt Ion:	276	Resp:	24906
Ion Ratio	Lower	Upper	
276	100		
277	23.9	22.2	33.4
138	26.9	24.1	36.1

