

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031025\
 Data File : BN036569.D
 Acq On : 10 Mar 2025 19:38
 Operator : RC/JU
 Sample : Q1531-06
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DUP01-20250306

Manual Integrations
 APPROVED

Reviewed By :Anahy Claudio 03/11/2025
 Supervised By :Jagrut Upadhyay 03/11/2025

Quant Time: Mar 11 00:52:53 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 10 16:06:28 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.724	152	2291	0.400	ng	0.00
7) Naphthalene-d8	10.519	136	4880	0.400	ng	0.01
13) Acenaphthene-d10	14.366	164	2910	0.400	ng	0.00
19) Phenanthrene-d10	17.111	188	5566	0.400	ng	# 0.00
29) Chrysene-d12	21.304	240	3909	0.400	ng	0.00
35) Perylene-d12	23.557	264	3415	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.319	112	829	0.155	ng	0.00
5) Phenol-d6	6.901	99	554	0.084	ng	0.00
8) Nitrobenzene-d5	8.875	82	1737	0.327	ng	0.00
11) 2-Methylnaphthalene-d10	12.116	152	2265	0.312	ng	0.00
14) 2,4,6-Tribromophenol	15.870	330	500	0.379	ng	0.01
15) 2-Fluorobiphenyl	12.993	172	5235	0.309	ng	0.00
27) Fluoranthene-d10	19.141	212	6371	0.447	ng	0.00
31) Terphenyl-d14	19.740	244	5588	0.597	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.239	88	9449	3.718	ng	98
34) Bis(2-ethylhexyl)phtha...	21.214	149	544m	0.056	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031025\
Data File : BN036569.D
Acq On : 10 Mar 2025 19:38
Operator : RC/JU
Sample : Q1531-06
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DUP01-20250306

Quant Time: Mar 11 00:52:53 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 16:06:28 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Anahy Claudio 03/11/2025
Supervised By :Jagrut Upadhyay 03/11/2025

