

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN031120\
 Data File : BN010193.D
 Acq On : 11 Mar 2020 15:41
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 12 07:26:12 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN021920.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 11 17:37:44 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	0.02
2	1,4-Dioxane	0.410	0.353	13.9	84	0.00
3	n-Nitrosodimethylamine	0.673	0.844	-25.4#	139	0.04
4 S	2-Fluorophenol	0.889	0.667	25.0	76	0.02
5 S	Phenol-d6	1.082	1.207	-11.6	114	0.07
6	bis(2-Chloroethyl)ether	1.052	0.857	18.5	82	0.04
7 I	Naphthalene-d8	1.000	1.000	0.0	72	0.03
8 S	Nitrobenzene-d5	0.333	0.316	5.1	73	0.05
9	Naphthalene	1.091	1.106	-1.4	75	0.03
10	Hexachlorobutadiene	0.240	0.263	-9.6	79	0.00
11 SURR	2-Methylnaphthalene-d10	0.774	0.817	-5.6	79	0.02
12	2-Methylnaphthalene	0.748	0.775	-3.6	77	0.03
13 I	Acenaphthene-d10	1.000	1.000	0.0	73	0.01
14 S	2,4,6-Tribromophenol	0.154	0.142	7.8	75	0.02
15 S	2-Fluorobiphenyl	1.510	1.406	6.9	71	0.02
16	Acenaphthylene	1.705	1.620	5.0	77	0.01
17	Acenaphthene	1.199	1.218	-1.6	77	0.01
18	Fluorene	1.582	1.460	7.7	69	0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	70	0.01
20	4-Bromophenyl-phenylether	0.279	0.135	51.6#	36#	0.00
21	Hexachlorobenzene	0.249	0.247	0.8	72	0.01
22	Pentachlorophenol	0.064	0.051	20.3	75	0.02
23	Phenanthrene	1.190	1.139	4.3	72	0.01
24	Anthracene	1.006	0.933	7.3	72	0.02
25 SURR	Fluoranthene-d10	1.384	1.278	7.7	69	0.00
26	Fluoranthene	1.425	1.296	9.1	69	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	73	0.01
28	Pvrene	1.519	1.413	7.0	70	0.00
29 S	Terphenyl-d14	0.956	0.884	7.5	69	0.01
30	Benzo(a)anthracene	1.327	1.200	9.6	74	0.01
31	Chrysene	1.536	1.603	-4.4	78	0.01
32	Bis(2-ethylhexyl)phthalate	0.648	0.675	-4.2	84	0.00
33	Indeno(1,2,3-cd)pyrene	1.599	1.637	-2.4	75	0.01
34 I	Perylene-d12	1.000	1.000	0.0	75	0.00
35	Benzo(b)fluoranthene	1.462	1.370	6.3	77	0.00
36	Benzo(k)fluoranthene	1.592	1.705	-7.1	84	0.00
37 C	Benzo(a)pyrene	1.334	1.269	4.9	73	0.01
38	Dibenzo(a,h)anthracene	1.314	1.173	10.7	71	0.02
39	Benzo(a,h,i)perylene	1.400	1.406	-0.4	80	0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN031120\
Data File : BN010193.D
Acq On : 11 Mar 2020 15:41
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 12 07:26:12 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN021920.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 11 17:37:44 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			