

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN031122\  
 Data File : BN018912.D  
 Acq On : 10 Mar 2022 20:49  
 Operator : CG/JU  
 Sample : N1862-01  
 Misc :  
 ALS Vial : 9 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 ClientSampleId :  
 V-2-0309-1A

Quant Time: Mar 11 03:05:24 2022  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_N\METHODS\8270-SIM-BN030322.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Fri Mar 11 03:02:57 2022  
 Response via : Initial Calibration


| Compound                      | R.T.   | QIon | Response | Conc  | Units | Dev(Min) |
|-------------------------------|--------|------|----------|-------|-------|----------|
| -----                         |        |      |          |       |       |          |
| Internal Standards            |        |      |          |       |       |          |
| 1) 1,4-Dichlorobenzene-d4     | 7.905  | 152  | 4302     | 0.400 | ng    | 0.00     |
| 7) Naphthalene-d8             | 10.712 | 136  | 12731    | 0.400 | ng    | # 0.00   |
| 13) Acenaphthene-d10          | 14.538 | 164  | 8191     | 0.400 | ng    | 0.00     |
| 19) Phenanthrene-d10          | 17.277 | 188  | 18852    | 0.400 | ng    | # 0.00   |
| 29) Chrysene-d12              | 21.449 | 240  | 18866    | 0.400 | ng    | # 0.00   |
| 35) Perylene-d12              | 23.840 | 264  | 17206    | 0.400 | ng    | 0.00     |
|                               |        |      |          |       |       |          |
| System Monitoring Compounds   |        |      |          |       |       |          |
| 4) 2-Fluorophenol             | 5.464  | 112  | 3056     | 0.325 | ng    | 0.00     |
| 5) Phenol-d6                  | 7.060  | 99   | 4395     | 0.389 | ng    | 0.00     |
| 8) Nitrobenzene-d5            | 9.057  | 82   | 2606     | 0.281 | ng    | 0.00     |
| 11) 2-Methylnaphthalene-d10   | 12.299 | 152  | 5676     | 0.274 | ng    | 0.00     |
| 14) 2,4,6-Tribromophenol      | 16.026 | 330  | 892      | 0.236 | ng    | 0.00     |
| 15) 2-Fluorobiphenyl          | 13.170 | 172  | 9143     | 0.260 | ng    | 0.00     |
| 27) Fluoranthene-d10          | 19.303 | 212  | 16420    | 0.296 | ng    | 0.00     |
| 31) Terphenyl-d14             | 19.893 | 244  | 10145    | 0.252 | ng    | 0.00     |
|                               |        |      |          |       |       |          |
| Target Compounds              |        |      |          |       |       | Qvalue   |
| 2) 1,4-Dioxane                | 3.297  | 88   | 111      | 0.023 | ng    | # 81     |
| 9) Naphthalene                | 10.765 | 128  | 2768     | 0.076 | ng    | # 92     |
| 12) 2-Methylnaphthalene       | 12.371 | 142  | 906      | 0.036 | ng    | # 93     |
| 16) Acenaphthylene            | 14.260 | 152  | 981      | 0.026 | ng    | 95       |
| 17) Acenaphthene              | 14.602 | 154  | 2118     | 0.080 | ng    | 98       |
| 18) Fluorene                  | 15.586 | 166  | 3055     | 0.089 | ng    | # 94     |
| 24) Pentachlorophenol         | 16.928 | 266  | 53       | 0.226 | ng    | # 70     |
| 25) Phenanthrene              | 17.318 | 178  | 45323    | 0.724 | ng    | 100      |
| 26) Anthracene                | 17.410 | 178  | 9096     | 0.173 | ng    | # 90     |
| 28) Fluoranthene              | 19.333 | 202  | 67909    | 0.954 | ng    | 100      |
| 30) Pyrene                    | 19.691 | 202  | 60168    | 0.714 | ng    | # 95     |
| 32) Benzo(a)anthracene        | 21.431 | 228  | 30325    | 0.428 | ng    | 99       |
| 33) Chrysene                  | 21.485 | 228  | 26473    | 0.340 | ng    | 97       |
| 34) Bis(2-ethylhexyl)phtha... | 21.360 | 149  | 2114     | 0.048 | ng    | # 99     |
| 36) Indeno(1,2,3-cd)pyrene    | 26.316 | 276  | 14256    | 0.184 | ng    | # 93     |
| 37) Benzo(b)fluoranthene      | 23.112 | 252  | 34090    | 0.426 | ng    | 96       |
| 39) Benzo(a)pyrene            | 23.732 | 252  | 23362    | 0.380 | ng    | 97       |
| 40) Dibenzo(a,h)anthracene    | 26.325 | 278  | 3587     | 0.061 | ng    | # 71     |
| 41) Benzo(g,h,i)perylene      | 27.077 | 276  | 15070    | 0.237 | ng    | 99       |
| -----                         |        |      |          |       |       |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN031122\  
Data File : BN018912.D  
Acq On : 10 Mar 2022 20:49  
Operator : CG/JU  
Sample : N1862-01  
Misc :  
ALS Vial : 9 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :  
V-2-0309-1A


Quant Time: Mar 11 03:05:24 2022  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_N\METHODS\8270-SIM-BN030322.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Fri Mar 11 03:02:57 2022  
Response via : Initial Calibration








#1  
 1,4-Dichlorobenzene-d4  
 Concen: 0.400 ng  
 RT: 7.905 min Scan# 6  
 Delta R.T. -0.007 min  
 Lab File: BN018912.D  
 Acq: 10 Mar 2022 20:49


Instrument :  
 BNA\_N  
 ClientSampleId :  
 V-2-0309-1A





Tgt Ion: 152 Resp: 4302  
 Ion Ratio Lower Upper  
 152 100  
 150 155.9 123.9 185.9  
 115 63.4 48.2 72.4




#2  
 1,4-Dioxane  
 Concen: 0.023 ng  
 RT: 3.297 min Scan# 30  
 Delta R.T. -0.014 min  
 Lab File: BN018912.D  
 Acq: 10 Mar 2022 20:49



Tgt Ion: 88 Resp: 111  
 Ion Ratio Lower Upper  
 88 100  
 43 0.0 24.6 37.0#  
 58 86.5 65.4 98.2









#4  
2-Fluorophenol  
Concen: 0.325 ng  
RT: 5.464 min Scan# 31  
Delta R.T. 0.007 min  
Lab File: BN018912.D  
Acq: 10 Mar 2022 20:49




Instrument :  
BNA\_N  
ClientSampleId :  
V-2-0309-1A


|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 112   | Resp: | 3056  |
| Ion      | Ratio | Lower | Upper |
| 112      | 100   |       |       |
| 64       | 61.1  | 47.8  | 71.8  |
| 63       | 31.8  | 25.7  | 38.5  |



#5  
Phenol-d6  
Concen: 0.389 ng  
RT: 7.060 min Scan# 551  
Delta R.T. -0.000 min  
Lab File: BN018912.D  
Acq: 10 Mar 2022 20:49

|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 99    | Resp: | 4395  |
| Ion      | Ratio | Lower | Upper |
| 99       | 100   |       |       |
| 42       | 28.6  | 16.2  | 24.4# |
| 71       | 35.8  | 27.7  | 41.5  |









#7  
Naphthalene-d8  
Concen: 0.400 ng  
RT: 10.712 min Scan# 94  
Delta R.T. -0.000 min  
Lab File: BN018912.D  
Acq: 10 Mar 2022 20:49




Instrument :  
BNA\_N  
ClientSampleId :  
V-2-0309-1A


|             |       |          |
|-------------|-------|----------|
| Tgt Ion:136 | Resp: | 12731    |
| Ion Ratio   | Lower | Upper    |
| 136         | 100   |          |
| 137         | 11.4  | 9.0 13.4 |
| 54          | 9.4   | 5.8 8.6# |
| 68          | 7.1   | 4.6 7.0# |



#8  
Nitrobenzene-d5  
Concen: 0.281 ng  
RT: 9.057 min Scan# 794  
Delta R.T. -0.000 min  
Lab File: BN018912.D  
Acq: 10 Mar 2022 20:49

|             |       |           |
|-------------|-------|-----------|
| Tgt Ion: 82 | Resp: | 2606      |
| Ion Ratio   | Lower | Upper     |
| 82          | 100   |           |
| 128         | 36.8  | 33.0 49.6 |
| 54          | 58.9  | 42.5 63.7 |









#9  
Naphthalene  
Concen: 0.076 ng  
RT: 10.765 min Scan# 91  
Delta R.T. -0.000 min  
Lab File: BN018912.D  
Acq: 10 Mar 2022 20:49




Instrument :  
BNA\_N  
ClientSampleId :  
V-2-0309-1A


Tgt Ion:128 Resp: 2768  
Ion Ratio Lower Upper  
128 100  
129 14.9 9.0 13.4#  
127 16.3 10.7 16.1#



#11  
2-Methylnaphthalene-d10  
Concen: 0.274 ng  
RT: 12.299 min Scan# 1232  
Delta R.T. -0.000 min  
Lab File: BN018912.D  
Acq: 10 Mar 2022 20:49

Tgt Ion:152 Resp: 5676  
Ion Ratio Lower Upper  
152 100  
151 20.5 16.5 24.7





#12

2-Methylnaphthalene

Concen: 0.036 ng

RT: 12.371 min Scan# 1251

Delta R.T. -0.000 min

Lab File: BN018912.D

Acq: 10 Mar 2022 20:49

Instrument :

BNA\_N

ClientSampleId :

V-2-0309-1A





Tgt Ion:142 Resp: 906

Ion Ratio Lower Upper

142 100

141 89.4 70.0 105.0

115 49.2 29.8 44.8#



#13

Acenaphthene-d10

Concen: 0.400 ng

RT: 14.538 min Scan# 1477

Delta R.T. -0.000 min

Lab File: BN018912.D

Acq: 10 Mar 2022 20:49




Tgt Ion:164 Resp: 8191


Ion Ratio Lower Upper

164 100

162 102.6 85.2 127.8

160 47.8 41.5 62.3









#14  
2,4,6-Tribromophenol  
Concen: 0.236 ng  
RT: 16.026 min Scan# 1617  
Delta R.T. -0.000 min  
Lab File: BN018912.D  
Acq: 10 Mar 2022 20:49




Instrument :  
BNA\_N  
ClientSampleId :  
V-2-0309-1A


Tgt Ion: 330 Resp: 892  
Ion Ratio Lower Upper  
330 100  
332 94.3 76.8 115.2  
141 39.8 32.2 48.2



#15  
2-Fluorobiphenyl  
Concen: 0.260 ng  
RT: 13.170 min Scan# 1349  
Delta R.T. -0.000 min  
Lab File: BN018912.D  
Acq: 10 Mar 2022 20:49

Tgt Ion: 172 Resp: 9143  
Ion Ratio Lower Upper  
172 100  
171 35.1 28.2 42.2  
170 22.6 18.7 28.1









#16  
Acenaphthylene  
Concen: 0.026 ng  
RT: 14.260 min Scan# 1451  
Delta R.T. -0.000 min  
Lab File: BN018912.D  
Acq: 10 Mar 2022 20:49




Instrument :  
BNA\_N  
ClientSampleId :  
V-2-0309-1A


|             |       |             |
|-------------|-------|-------------|
| Tgt Ion:152 | Resp: | 981         |
| Ion         | Ratio | Lower Upper |
| 152         | 100   |             |
| 151         | 22.7  | 15.9 23.9   |
| 153         | 14.3  | 10.6 15.8   |



#17  
Acenaphthene  
Concen: 0.080 ng  
RT: 14.602 min Scan# 1483  
Delta R.T. -0.000 min  
Lab File: BN018912.D  
Acq: 10 Mar 2022 20:49

|             |       |             |
|-------------|-------|-------------|
| Tgt Ion:154 | Resp: | 2118        |
| Ion         | Ratio | Lower Upper |
| 154         | 100   |             |
| 153         | 114.4 | 89.7 134.5  |
| 152         | 56.2  | 44.7 67.1   |









#18  
Fluorene  
Concen: 0.089 ng  
RT: 15.586 min Scan# 1575  
Delta R.T. -0.000 min  
Lab File: BN018912.D  
Acq: 10 Mar 2022 20:49




Instrument :  
BNA\_N  
ClientSampleId :  
V-2-0309-1A


Tgt Ion:166 Resp: 3055  
Ion Ratio Lower Upper  
166 100  
165 93.7 79.2 118.8  
167 16.6 10.7 16.1#



#19  
Phenanthrene-d10  
Concen: 0.400 ng  
RT: 17.277 min Scan# 1739  
Delta R.T. -0.000 min  
Lab File: BN018912.D  
Acq: 10 Mar 2022 20:49

Tgt Ion:188 Resp: 18852  
Ion Ratio Lower Upper  
188 100  
94 0.0 0.0 0.0  
80 10.3 11.8 17.8#









#24  
Pentachlorophenol  
Concen: 0.226 ng  
RT: 16.928 min Scan# 1705  
Delta R.T. -0.000 min  
Lab File: BN018912.D  
Acq: 10 Mar 2022 20:49




Instrument :  
BNA\_N  
ClientSampleId :  
V-2-0309-1A


Tgt Ion:266 Resp: 53  
Ion Ratio Lower Upper  
266 100  
264 43.4 50.2 75.4#  
268 37.7 51.9 77.9#



#25  
Phenanthrene  
Concen: 0.724 ng  
RT: 17.318 min Scan# 1743  
Delta R.T. -0.000 min  
Lab File: BN018912.D  
Acq: 10 Mar 2022 20:49

Tgt Ion:178 Resp: 45323  
Ion Ratio Lower Upper  
178 100  
176 19.2 15.4 23.0  
179 15.3 12.2 18.2









#26  
 Anthracene  
 Concen: 0.173 ng  
 RT: 17.410 min Scan# 11  
 Delta R.T. -0.000 min  
 Lab File: BN018912.D  
 Acq: 10 Mar 2022 20:49




Instrument :  
 BNA\_N  
 ClientSampleId :  
 V-2-0309-1A


Tgt Ion:178 Resp: 9096  
 Ion Ratio Lower Upper  
 178 100  
 176 21.7 14.8 22.2  
 179 21.2 12.2 18.4#



#27  
 Fluoranthene-d10  
 Concen: 0.296 ng  
 RT: 19.303 min Scan# 2118  
 Delta R.T. -0.000 min  
 Lab File: BN018912.D  
 Acq: 10 Mar 2022 20:49

Tgt Ion:212 Resp: 16420  
 Ion Ratio Lower Upper  
 212 100  
 106 15.6 12.7 19.1  
 104 9.0 7.1 10.7









#28  
Fluoranthene  
Concen: 0.954 ng  
RT: 19.333 min Scan# 2125  
Delta R.T. -0.000 min  
Lab File: BN018912.D  
Acq: 10 Mar 2022 20:49




Instrument :  
BNA\_N  
ClientSampleId :  
V-2-0309-1A


Tgt Ion:202 Resp: 67909  
Ion Ratio Lower Upper  
202 100  
101 11.9 9.8 14.6  
203 17.2 13.7 20.5



#29  
Chrysene-d12  
Concen: 0.400 ng  
RT: 21.449 min Scan# 2446  
Delta R.T. -0.000 min  
Lab File: BN018912.D  
Acq: 10 Mar 2022 20:49

Tgt Ion:240 Resp: 18866  
Ion Ratio Lower Upper  
240 100  
120 13.8 8.5 12.7#  
236 28.8 22.4 33.6









#30  
 Pyrene  
 Concen: 0.714 ng  
 RT: 19.691 min Scan# 21  
 Delta R.T. -0.004 min  
 Lab File: BN018912.D  
 Acq: 10 Mar 2022 20:49




Instrument :  
 BNA\_N  
 ClientSampleId :  
 V-2-0309-1A


Tgt Ion:202 Resp: 60168  
 Ion Ratio Lower Upper  
 202 100  
 200 21.1 16.9 25.3  
 203 22.2 14.2 21.2#



#31  
 Terphenyl-d14  
 Concen: 0.252 ng  
 RT: 19.893 min Scan# 2258  
 Delta R.T. -0.004 min  
 Lab File: BN018912.D  
 Acq: 10 Mar 2022 20:49

Tgt Ion:244 Resp: 10145  
 Ion Ratio Lower Upper  
 244 100  
 212 10.9 8.2 12.4  
 122 0.0 0.0 0.0









#32  
Benzo(a)anthracene  
Concen: 0.428 ng  
RT: 21.431 min Scan# 2444  
Delta R.T. -0.009 min  
Lab File: BN018912.D  
Acq: 10 Mar 2022 20:49




Instrument :  
BNA\_N  
ClientSampleId :  
V-2-0309-1A


Tgt Ion:228 Resp: 30325  
Ion Ratio Lower Upper  
228 100  
226 28.1 21.9 32.9  
229 20.4 16.0 24.0



#33  
Chrysene  
Concen: 0.340 ng  
RT: 21.485 min Scan# 2450  
Delta R.T. -0.000 min  
Lab File: BN018912.D  
Acq: 10 Mar 2022 20:49

Tgt Ion:228 Resp: 26473  
Ion Ratio Lower Upper  
228 100  
226 30.6 24.0 36.0  
229 22.3 15.9 23.9









#34  
 Bis(2-ethylhexyl)phthalate  
 Concen: 0.048 ng  
 RT: 21.360 min Scan# 2436  
 Delta R.T. -0.009 min  
 Lab File: BN018912.D  
 Acq: 10 Mar 2022 20:49




Instrument :  
 BNA\_N  
 ClientSampleId :  
 V-2-0309-1A


Tgt Ion:149 Resp: 2114  
 Ion Ratio Lower Upper  
 149 100  
 167 27.0 21.8 32.8  
 279 3.6 2.0 3.0#



#35  
 Perylene-d12  
 Concen: 0.400 ng  
 RT: 23.840 min Scan# 3043  
 Delta R.T. -0.003 min  
 Lab File: BN018912.D  
 Acq: 10 Mar 2022 20:49

Tgt Ion:264 Resp: 17206  
 Ion Ratio Lower Upper  
 264 100  
 260 27.8 22.4 33.6  
 265 29.9 25.3 37.9









#36  
Indeno(1,2,3-cd)pyrene  
Concen: 0.184 ng  
RT: 26.316 min Scan# 31  
Delta R.T. -0.003 min  
Lab File: BN018912.D  
Acq: 10 Mar 2022 20:49




Instrument :  
BNA\_N  
ClientSampleId :  
V-2-0309-1A


Tgt Ion:276 Resp: 14256  
Ion Ratio Lower Upper  
276 100  
138 24.3 24.5 36.7#  
277 24.6 19.8 29.8



#37  
Benzo(b)fluoranthene  
Concen: 0.426 ng  
RT: 23.112 min Scan# 2794  
Delta R.T. -0.003 min  
Lab File: BN018912.D  
Acq: 10 Mar 2022 20:49

Tgt Ion:252 Resp: 34090  
Ion Ratio Lower Upper  
252 100  
253 25.7 18.9 28.3  
125 16.7 12.1 18.1





#39

Benzo(a)pyrene

Concen: 0.380 ng

RT: 23.732 min Scan# 3014

Delta R.T. -0.006 min

Lab File: BN018912.D


Acq: 10 Mar 2022 20:49

Instrument :

BN018912.D

ClientSampleId :

V-2-0309-1A






Tgt Ion:252 Resp: 23362

Ion Ratio Lower Upper

252 100

253 26.3 20.1 30.1

125 17.1 15.2 22.8



#40

Dibenzo(a,h)anthracene

Concen: 0.061 ng

RT: 26.325 min Scan# 3893

Delta R.T. -0.012 min

Lab File: BN018912.D

Acq: 10 Mar 2022 20:49




Tgt Ion:278 Resp: 3587


Ion Ratio Lower Upper

278 100

139 32.8 17.2 25.8#


279 41.3 19.9 29.9#





#41  
 Benzo(g,h,i)perylene  
 Concen: 0.237 ng  
 RT: 27.077 min Scan# 41  
 Delta R.T. -0.006 min  
 Lab File: BN018912.D  
 Acq: 10 Mar 2022 20:49

Instrument :  
 BNA\_N  
 ClientSampleId :  
 V-2-0309-1A



|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 276   | Resp: | 15070 |
| Ion      | Ratio | Lower | Upper |
| 276      | 100   |       |       |
| 277      | 25.0  | 19.4  | 29.0  |
| 138      | 26.3  | 21.0  | 31.6  |

