

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031122\
 Data File : BN018920.D
 Acq On : 11 Mar 2022 01:35
 Operator : CG/JU
 Sample : SP5839
 Misc : 8270 SIM SURROGATE
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SP5839

Quant Time: Mar 11 03:06:15 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN030322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 11 03:02:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.905	152	4937	0.400	ng	0.00
7) Naphthalene-d8	10.712	136	14561	0.400	ng	# 0.00
13) Acenaphthene-d10	14.538	164	9054	0.400	ng	0.00
19) Phenanthrene-d10	17.277	188	19817	0.400	ng	# 0.00
29) Chrysene-d12	21.449	240	19102	0.400	ng	# 0.00
35) Perylene-d12	23.840	264	15518	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.464	112	5681	0.526	ng	0.00
5) Phenol-d6	7.060	99	6676	0.515	ng	0.00
8) Nitrobenzene-d5	9.057	82	4013	0.378	ng	0.00
11) 2-Methylnaphthalene-d10	12.299	152	8487	0.358	ng	0.00
14) 2,4,6-Tribromophenol	16.025	330	1741	0.417	ng	0.00
15) 2-Fluorobiphenyl	13.169	172	12253	0.316	ng	0.00
27) Fluoranthene-d10	19.303	212	24086	0.413	ng	0.00
31) Terphenyl-d14	19.893	244	14430	0.354	ng	0.00

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031122\
Data File : BN018920.D
Acq On : 11 Mar 2022 01:35
Operator : CG/JU
Sample : SP5839
Misc : 8270 SIM SURROGATE
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SP5839

Quant Time: Mar 11 03:06:15 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN030322.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 11 03:02:57 2022
Response via : Initial Calibration

