




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031125\
Data File : BN036573.D
Acq On : 11 Mar 2025 08:50
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Mar 11 17:55:30 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 25 04:23:53 2024
Response via : Initial Calibration



TIC: BN036573.D\data.ms

(70) Pentachlorophenol (C)




16.763min (-0.003) 22415.21 ng

response 337921

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	64.83
264.00	61.60	67.30
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031125\
Data File : BN036573.D
Acq On : 11 Mar 2025 08:50
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Mar 11 17:55:30 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 25 04:23:53 2024
Response via : Initial Calibration



TIC: BN036573.D\data.ms

(77) Benzdine

19.363min (+ 0.009) 0.00 ng

response 1813573

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	13.66
183.00	13.20	12.80
0.00	0.00	0.00