

Data Path : Z:\HPCHEM1\BNA_N\Data\BN031218\
 Data File : BN000301.D
 Acq On : 12 Mar 2018 08:57
 Operator : JU/SJ
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTD02094

Quant Time: Mar 12 11:45:06 2018
 Quant Method : Z:\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN030918MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Mar 09 23:47:40 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	112	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	112	0.00
3	Naphthalene	1.055	1.021	3.2	110	0.00
4	2-Methylnaphthalene	0.743	0.717	3.5	109	0.00
5	1-Methylnaphthalene	0.746	0.720	3.5	109	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	111	0.00
7 S	Acenaphthylene-d8	1.985	1.979	0.3	110	0.00
8	Acenaphthylene	1.995	1.960	1.8	110	0.00
9 C	Acenaphthene	1.388	1.340	3.5	109	0.00
10 S	Fluorene-d10	1.344	1.291	3.9	109	0.00
11	Fluorene	1.531	1.475	3.7	108	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	111	0.00
13	Phenanthrene	1.147	1.109	3.3	108	0.00
14 S	Anthracene-d10	0.934	0.923	1.2	109	0.00
15	Anthracene	1.164	1.140	2.1	107	0.00
16 C	Fluoranthene	1.239	1.219	1.6	110	0.00
17 I	Chrysene-d12	1.000	1.000	0.0	108	0.00
18 S	Pyrene-d10	0.978	0.985	-0.7	108	0.00
19	Pyrene	1.263	1.278	-1.2	109	0.00
20	Benzo(a)anthracene	1.243	1.213	2.4	106	0.00
21	Chrysene	1.180	1.151	2.5	107	0.00
22 I	Perylene-d12	1.000	1.000	0.0	100	0.00
23	Benzo(b)fluoranthene	1.193	1.226	-2.8	104	0.00
24	Benzo(k)fluoranthene	1.116	1.150	-3.0	100	0.00
25 S	Benzo(a)pyrene-d12	0.913	0.910	0.3	101	0.00
26 C	Benzo(a)pyrene	1.135	1.122	1.1	99	0.00
27	Indeno(1,2,3-cd)pyrene	1.224	1.145	6.5	97	0.00
28	Dibenzo(a,h)anthracene	1.003	0.946	5.7	96	0.00
29	Benzo(g,h,i)perylene	1.022	0.944	7.6	98	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0