

Data Path : Z:\HPCHEM1\BNA_N\DATA\BN031218\
 Data File : BN000313.D
 Acq On : 12 Mar 2018 16:38
 Operator : JU/SJ
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTD02095

Quant Time: Mar 12 17:12:13 2018
 Quant Method : Z:\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN030918MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Mar 12 15:04:53 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	100	0.00
3	Naphthalene	1.055	1.069	-1.3	103	0.00
4	2-Methylnaphthalene	0.743	0.741	0.3	101	0.00
5	1-Methylnaphthalene	0.746	0.741	0.7	100	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	98	0.00
7 S	Acenaphthylene-d8	1.985	2.088	-5.2	102	0.00
8	Acenaphthylene	1.995	2.062	-3.4	102	0.00
9 C	Acenaphthene	1.388	1.410	-1.6	101	0.00
10 S	Fluorene-d10	1.344	1.346	-0.1	100	0.00
11	Fluorene	1.531	1.551	-1.3	100	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	99	0.00
13	Phenanthrene	1.147	1.163	-1.4	101	0.00
14 S	Anthracene-d10	0.934	0.964	-3.2	101	0.00
15	Anthracene	1.164	1.195	-2.7	100	0.00
16 C	Fluoranthene	1.239	1.305	-5.3	104	0.00
17 I	Chrysene-d12	1.000	1.000	0.0	100	0.00
18 S	Pyrene-d10	0.978	1.015	-3.8	104	0.00
19	Pyrene	1.263	1.305	-3.3	104	0.00
20	Benzo(a)anthracene	1.243	1.281	-3.1	104	0.00
21	Chrysene	1.180	1.206	-2.2	105	0.00
22 I	Perylene-d12	1.000	1.000	0.0	93	0.00
23	Benzo(b)fluoranthene	1.193	1.302	-9.1	102	0.00
24	Benzo(k)fluoranthene	1.116	1.212	-8.6	98	0.00
25 S	Benzo(a)pyrene-d12	0.913	0.963	-5.5	99	0.00
26 C	Benzo(a)pyrene	1.135	1.179	-3.9	96	0.00
27	Indeno(1,2,3-cd)pyrene	1.224	1.190	2.8	94	0.00
28	Dibenzo(a,h)anthracene	1.003	1.004	-0.1	95	0.00
29	Benzo(g,h,i)perylene	1.022	0.960	6.1	92	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0