


Data Path : Z:\HPCHEM1\BNA_N\DATA\BN031218\
 Data File : BN000300.D
 Acq On : 12 Mar 2018 08:19
 Operator : JU/SJ
 Sample : DFTPP76
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP76

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN030918MA.M
 Title : SVOA CALIBRATION
 Last Update : Tue Mar 13 15:51:19 2018



AutoFind: Scans 2581, 2582, 2583; Background Corrected with Scan 2574

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	35.7	91496	PASS
68	69	0.00	2	1.8	1953	PASS
69	198	0.00	100	42.7	109520	PASS
70	69	0.00	2	0.5	596	PASS
127	198	10	80	52.3	134090	PASS
197	198	0.00	2	0.5	1295	PASS
198	198	100	100	100.0	256426	PASS
199	198	5	9	6.8	17480	PASS
275	198	10	60	23.2	59600	PASS
365	198	1	100	2.8	7136	PASS
441	443	0.01	100	84.1	28253	PASS
442	198	50	100	69.9	179336	PASS
443	442	15	24	18.7	33576	PASS

m/z	Abundance
37.00	504.0
38.10	1147.0
39.10	6973.0
40.00	1106.0
41.00	443.0
44.00	822.0
49.10	802.0
50.10	25496.0
51.10	99664.0
52.10	5466.0
55.00	704.0
56.00	3271.0
57.00	6701.0
61.00	1330.0
62.00	1536.0
63.00	4639.0
64.00	589.0
65.10	2032.0
66.10	304.0
68.00	2081.0
69.00	117208.0
70.10	591.0
73.10	1302.0
74.00	12377.0
75.00	19768.0
76.10	6728.0
77.10	143104.0
78.10	10149.0
79.00	8996.0
80.00	7622.0
81.00	9188.0
82.00	2563.0
83.00	2243.0
84.00	346.0
85.00	2015.0
86.00	3079.0
87.00	1150.0
88.00	583.0
91.00	2219.0
92.00	2779.0
93.00	15692.0
94.10	1284.0
95.00	360.0
96.00	1137.0
96.90	327.0
98.00	11690.0
99.00	11831.0
100.10	1145.0
101.00	4989.0
102.00	385.0
103.00	1847.0
104.00	3662.0
105.00	3165.0
106.10	1360.0
107.00	41488.0
108.00	6518.0
109.00	1224.0
110.00	74672.0
111.00	10890.0
112.00	1465.0
113.10	503.0
115.10	373.0
116.00	2413.0
117.00	35128.0
118.00	2963.0
119.00	497.0
120.10	441.0
122.00	2521.0
123.00	3917.0
124.00	1825.0
125.00	1528.0
126.20	697.0
127.00	136640.0
128.00	10831.0
129.00	51448.0
130.00	5291.0
131.00	868.0
133.00	536.0
134.00	1428.0
135.00	3931.0
136.00	1569.0
137.00	3014.0
138.10	599.0
140.00	518.0
141.00	6176.0
142.00	1925.0
143.00	1417.0
144.10	396.0
145.00	382.0
146.00	1203.0
147.00	3619.0

Instrument :
BNA_N
ClientSampleId :
DFTPP76

148.00	8357.0
149.00	1955.0
150.00	624.0
151.10	777.0
153.00	1861.0
153.90	1523.0
155.00	3830.0
156.10	4983.0
157.00	1090.0
157.90	1220.0
159.00	1042.0
160.00	1964.0
161.10	3158.0
162.00	897.0
163.10	312.0
164.10	338.0
165.10	2635.0
166.00	2050.0
167.00	13614.0
168.00	9899.0
169.00	1560.0
170.00	361.0
171.00	485.0
172.00	1125.0
173.10	1323.0
174.00	2574.0
175.10	5432.0
176.00	1606.0
177.00	2390.0
178.00	887.0
179.00	9093.0
180.00	6591.0
181.00	3444.0
182.00	535.0
184.10	632.0
185.00	4209.0
186.10	34616.0
187.10	10718.0
188.00	1280.0
189.00	2030.0
190.00	432.0
191.00	1226.0
192.00	3103.0
193.00	3634.0
194.10	662.0
195.10	465.0
196.00	7046.0
197.10	1563.0
198.00	243456.0
199.00	16280.0
200.10	1148.0
201.60	1036.0
203.00	1708.0
204.10	8765.0
205.10	14587.0
206.10	61424.0
207.00	9569.0
208.00	2768.0
209.00	1132.0
210.20	986.0
211.10	2354.0
215.00	613.0
216.00	1402.0
217.00	15443.0
218.00	2322.0
221.00	8906.0
222.10	1949.0
223.00	3769.0
224.10	33040.0
225.10	8762.0
226.10	966.0
227.00	14524.0
228.00	1908.0
229.00	2809.0
230.00	386.0
231.00	1004.0
233.10	301.0
234.00	871.0
235.00	896.0
236.00	694.0
237.00	1466.0
239.00	580.0
240.00	504.0
241.10	715.0
242.10	1526.0
243.00	1636.0
244.10	25728.0
245.10	3372.0
246.00	5257.0
247.00	842.0
249.00	1057.0
253.10	665.0
255.00	121256.0
256.00	17704.0
257.00	1415.0

Instrument :
BNA_N
ClientSampleId :
DFTPP76

258.00	6714.0
259.00	1050.0
264.20	314.0
265.00	3074.0
266.00	380.0
273.00	3203.0
274.10	10128.0
275.10	52312.0
276.10	7236.0
277.10	4232.0
278.00	1045.0
283.00	348.0
285.10	717.0
293.00	918.0
294.00	327.0
296.10	14195.0
297.00	1983.0
303.00	1674.0
304.10	548.0
314.00	677.0
315.00	1488.0
316.10	850.0
321.00	414.0
322.10	307.0
323.10	4272.0
324.00	731.0
327.00	903.0
327.90	405.0
332.00	416.0
333.10	434.0
334.10	2437.0
335.10	737.0
341.00	813.0
346.00	853.0
352.10	1106.0
353.00	892.0
354.10	1447.0
355.10	324.0
365.00	5572.0
366.00	901.0
371.00	358.0
372.10	2153.0
373.00	464.0
383.10	500.0
390.00	323.0
402.10	960.0
403.10	1189.0
404.00	327.0
421.00	939.0
421.90	1033.0
423.10	7212.0
424.10	2261.0
425.10	346.0
441.10	19928.0
442.10	128216.0
443.10	25208.0
444.10	2589.0

Instrument :
BNA_N
ClientSampleId :
DFTPP76

m/z	Abundance
37.00	521.0
38.00	1325.0
39.10	7857.0
40.00	980.0
41.10	389.0
43.00	308.0
44.00	690.0
49.00	970.0
50.10	28264.0
51.10	110160.0
52.10	5400.0
55.00	848.0
56.00	3396.0
57.00	7795.0
61.00	1387.0
62.00	1766.0
63.00	4956.0
64.00	646.0
65.10	2674.0
67.00	347.0
68.10	2186.0
69.00	129712.0
70.00	685.0
73.10	1616.0
74.00	14804.0
75.00	21384.0
76.10	7605.0
77.10	159744.0
78.10	10975.0
79.00	10488.0
80.00	7523.0
81.00	11097.0
82.00	2880.0
83.00	2585.0
84.00	577.0
85.00	2514.0
86.00	3299.0
87.00	1478.0
88.00	524.0
88.90	320.0
91.00	2597.0
92.00	2919.0
93.00	16896.0
94.00	1156.0
95.10	335.0
96.00	906.0
96.90	448.0
98.00	13855.0
99.00	12489.0
100.00	1055.0
101.00	5980.0
102.00	379.0
103.00	2240.0
104.00	3723.0
105.00	3823.0
106.00	1171.0
107.00	47896.0
108.00	7398.0
109.00	1344.0
110.00	87632.0
111.00	12496.0
112.00	1725.0
113.10	622.0
116.00	2631.0
117.00	40160.0
118.00	3219.0
119.00	512.0
120.10	718.0
122.00	3217.0
123.00	3979.0
124.00	2112.0
125.10	1904.0
126.10	956.0
127.00	160640.0
128.00	11597.0
129.00	59912.0
130.00	4682.0
131.00	1152.0
132.00	654.0
133.00	494.0
134.00	1809.0
135.00	4714.0
136.00	1988.0
137.00	3065.0
138.00	594.0
139.00	431.0
140.00	1032.0
141.00	7635.0
142.10	2969.0
143.00	1847.0
144.00	449.0

Instrument :
BNA_N
ClientSampleId :
DFTPP76

144.90	437.0
146.00	1453.0
147.00	4450.0
148.00	9365.0
149.00	2193.0
150.10	720.0
150.90	946.0
151.70	678.0
153.00	2609.0
154.00	1943.0
155.00	4514.0
156.10	6061.0
157.10	1420.0
158.00	1465.0
159.00	1112.0
160.00	2175.0
161.00	3408.0
162.00	1069.0
163.90	485.0
165.00	2921.0
166.00	2040.0
167.00	16648.0
168.00	10296.0
169.10	1709.0
170.00	449.0
170.90	750.0
172.00	1401.0
173.00	1835.0
174.00	3140.0
175.10	6307.0
176.00	1653.0
177.00	2543.0
178.10	1053.0
179.00	11154.0
180.10	7418.0
181.10	4305.0
182.00	577.0
182.90	436.0
184.00	875.0
185.00	5568.0
186.00	42872.0
187.10	12052.0
188.00	1325.0
189.00	2633.0
190.00	542.0
191.00	1286.0
192.00	4230.0
193.00	4279.0
194.10	841.0
195.00	399.0
196.00	9007.0
197.20	2322.0
198.00	308672.0
199.00	21664.0
200.00	1724.0
201.50	1440.0
203.00	2238.0
204.10	11092.0
205.10	19088.0
206.10	78760.0
207.10	13176.0
208.00	2999.0
209.10	1194.0
210.20	1147.0
211.00	3258.0
215.00	1042.0
216.00	1824.0
217.00	20128.0
218.00	2577.0
221.00	11612.0
221.90	2438.0
223.00	4776.0
224.10	42520.0
225.00	11041.0
226.10	1336.0
227.00	17496.0
228.00	2704.0
229.00	3914.0
230.00	436.0
231.00	1525.0
234.00	948.0
235.00	1403.0
236.00	1023.0
237.00	1614.0
239.10	513.0
240.10	514.0
241.00	974.0
242.10	2278.0
243.10	2501.0
244.10	35392.0
245.10	4360.0
246.00	6841.0
247.00	1396.0
248.00	391.0
249.00	1433.0

Instrument :
BNA_N
ClientSampleId :
DFTPP76

250.00	356.0
251.90	406.0
253.00	739.0
255.00	158272.0
256.10	23896.0
257.00	2110.0
258.00	9388.0
259.10	1648.0
261.00	306.0
264.10	319.0
265.00	3523.0
266.00	498.0
271.00	330.0
272.00	486.0
273.10	5136.0
274.10	12697.0
275.10	72768.0
276.10	9406.0
277.00	6149.0
278.00	1036.0
283.00	687.0
284.00	516.0
285.10	1140.0
292.10	397.0
293.00	1243.0
294.10	421.0
295.00	534.0
296.00	19352.0
297.00	2907.0
298.10	305.0
303.10	1919.0
304.00	598.0
314.00	920.0
315.00	2096.0
316.10	1345.0
321.10	611.0
322.00	351.0
323.10	6697.0
324.00	1042.0
327.00	1465.0
328.00	624.0
332.10	550.0
333.00	632.0
334.10	3800.0
335.10	919.0
341.00	837.0
342.00	318.0
346.00	1607.0
347.10	356.0
352.00	1922.0
353.10	1279.0
354.10	1942.0
355.10	455.0
365.00	8750.0
366.00	1281.0
371.10	435.0
372.10	3331.0
373.00	665.0
383.00	865.0
390.00	391.0
391.00	405.0
402.00	1470.0
403.00	1990.0
404.00	833.0
421.10	1570.0
422.10	1581.0
423.10	11381.0
424.10	3179.0
425.00	497.0
441.10	34152.0
442.10	215168.0
443.10	39760.0
444.10	3846.0

Instrument :
BNA_N
ClientSampleId :
DFTPP76

m/z	Abundance
37.00	370.0
38.00	762.0
39.10	4804.0
40.00	815.0
41.00	462.0
44.00	651.0
49.10	701.0
50.10	17320.0
51.10	64664.0
52.00	3493.0
55.00	522.0
56.00	1926.0
57.00	4495.0
61.00	911.0
62.00	1155.0
63.00	3293.0
64.00	407.0
65.00	1499.0
67.10	352.0
68.10	1594.0
69.00	81640.0
70.10	514.0
73.00	1158.0
74.00	8650.0
75.00	13714.0
76.10	4483.0
77.10	99728.0
78.10	6830.0
79.00	6517.0
80.00	4927.0
81.00	7151.0
82.00	1850.0
83.00	1681.0
84.00	409.0
85.00	1616.0
86.00	1944.0
87.00	1094.0
87.90	385.0
91.00	1872.0
92.00	1734.0
93.00	11502.0
94.00	942.0
95.00	401.0
96.00	784.0
96.90	321.0
98.00	8918.0
99.00	8166.0
100.00	679.0
101.00	4043.0
102.00	334.0
102.90	1247.0
104.00	2776.0
105.00	2556.0
106.00	746.0
107.00	31872.0
108.00	5126.0
109.00	916.0
110.00	55640.0
111.00	8660.0
112.00	1079.0
113.10	308.0
116.00	1625.0
117.00	27296.0
118.00	2373.0
119.00	406.0
120.10	441.0
122.00	2196.0
123.00	2941.0
124.00	1616.0
125.00	1132.0
126.20	468.0
127.00	104992.0
128.00	7707.0
129.00	38504.0
130.00	3287.0
131.00	660.0
132.10	332.0
132.90	383.0
134.00	1183.0
135.00	3276.0
136.00	1161.0
137.10	2061.0
137.90	451.0
140.10	439.0
141.00	4755.0
142.00	1877.0
142.90	1249.0
144.00	401.0
146.00	884.0
147.00	2884.0
148.00	6736.0

Instrument :
BNA_N
ClientSampleId :
DFTPP76

149.00	1634.0
150.00	476.0
151.00	889.0
151.90	433.0
153.00	1666.0
154.10	1112.0
155.00	3127.0
156.10	4593.0
156.90	885.0
157.90	871.0
159.00	745.0
160.00	1610.0
161.00	2352.0
161.90	924.0
165.00	1783.0
166.10	1389.0
167.00	12176.0
168.00	6891.0
169.00	908.0
170.00	436.0
170.80	547.0
171.90	874.0
173.00	1540.0
174.00	2696.0
175.10	4042.0
176.10	1079.0
177.10	1909.0
178.00	807.0
179.00	7976.0
180.00	5995.0
181.10	2544.0
182.00	423.0
184.10	627.0
185.00	3896.0
186.00	29848.0
187.00	7921.0
188.00	902.0
189.00	2028.0
189.90	352.0
190.90	1106.0
192.00	2670.0
193.00	3623.0
194.10	615.0
195.00	443.0
196.00	5912.0
198.00	217152.0
199.00	14496.0
200.00	1290.0
201.50	821.0
203.00	1310.0
204.00	7275.0
205.00	13122.0
206.10	54256.0
207.10	8925.0
208.00	2232.0
209.00	954.0
210.20	886.0
211.10	2019.0
215.00	527.0
216.00	1324.0
217.00	14184.0
218.00	1964.0
221.10	8510.0
221.90	1958.0
223.00	3234.0
224.10	29528.0
225.10	7872.0
226.10	942.0
227.00	13345.0
228.00	1909.0
229.00	2669.0
230.00	444.0
231.10	1165.0
234.10	786.0
235.00	979.0
236.00	825.0
237.10	1384.0
239.00	576.0
239.90	488.0
241.00	693.0
242.10	1972.0
243.10	1775.0
244.10	23480.0
245.10	3545.0
246.00	5214.0
247.00	1069.0
248.90	546.0
252.10	339.0
253.10	599.0
254.10	1038.0
255.00	123128.0
256.10	17336.0
257.10	1400.0
258.00	7694.0
259.10	1087.0

Instrument :

BNA_N

ClientSampleId :

DFTPP76

264.10	312.0
265.00	3185.0
266.00	443.0
271.00	310.0
272.10	389.0
273.00	3491.0
274.10	9866.0
275.10	53720.0
276.00	6804.0
277.00	4172.0
278.00	889.0
281.10	382.0
283.00	643.0
284.00	362.0
285.00	878.0
292.00	394.0
293.00	1138.0
294.00	321.0
295.00	413.0
296.10	14838.0
297.00	2082.0
303.00	1896.0
304.00	558.0
310.00	311.0
314.00	847.0
315.00	1843.0
316.00	889.0
321.10	637.0
323.10	4400.0
324.00	1018.0
327.00	1345.0
327.90	664.0
332.00	468.0
334.10	3520.0
335.10	766.0
341.10	559.0
346.00	1442.0
352.00	1578.0
353.10	871.0
354.10	1653.0
365.00	7086.0
366.00	1062.0
371.10	357.0
372.00	2661.0
373.10	700.0
383.00	886.0
390.00	398.0
391.10	330.0
402.10	1269.0
403.10	1614.0
404.00	628.0
421.00	1318.0
422.00	1580.0
423.10	10311.0
424.10	2502.0
441.10	30680.0
442.10	194624.0
443.10	35760.0
444.10	3113.0

Instrument :
BNA_N
ClientSampleId :
DFTPP76

Scan 2574 (18.128 min): BN000300.D
DFTPP76

m/z	Abundance
40.00	624.0
41.00	304.0
44.00	726.0
73.10	451.0
91.10	371.0
96.00	453.0
207.10	2293.0
208.00	386.0

Instrument :
BNA_N
ClientSampleId :
DFTPP76