

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN031221\
 Data File : BN013930.D
 Acq On : 12 Mar 2021 13:41
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SICV010

Quant Time: Mar 12 14:12:19 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SFAM-EPA-SIM-BN031221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 12 13:33:49 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.72	152	4172	0.40	ng/ul	0.00
4) Naphthalene-d8	10.50	136	13827	0.40	ng/ul	0.00
9) Acenaphthene-d10	14.35	164	6928	0.40	ng/ul	0.00
13) Phenanthrene-d10	17.10	188	14923	0.40	ng/ul	0.00
17) Chrysene-d12	21.27	240	13718	0.40	ng/ul	0.00
23) Perylene-d12	23.51	264	10278	0.40	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.22	96	1756	0.39	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.09	152	8035	0.40	ng/ul	0.00
18) Fluoranthene-d10	19.13	212	15627	0.38	ng/ul	0.00
Target Compounds					Ovalue	
2) 1,4-Dioxane	3.25	88	1846	0.395	ng/ul	94
5) Naphthalene	10.54	128	16053	0.398	ng/ul	100
7) 2-Methylnaphthalene	12.17	142	10465	0.399	ng/ul	100
8) 1-Methylnaphthalene	12.39	142	10121	0.400	ng/ul	100
10) Acenaphthylene	14.07	152	11947	0.382	ng/ul	99
11) Acenaphthene	14.42	153	10284	0.396	ng/ul	99
12) Fluorene	15.41	166	11472	0.396	ng/ul	99
14) Pentachlorophenol	16.75	266	839	0.329	ng/ul	99
15) Phenanthrene	17.14	178	19593	0.393	ng/ul	100
16) Anthracene	17.23	178	15872	0.386	ng/ul	99
19) Fluoranthene	19.15	202	20718	0.376	ng/ul	99
20) Pyrene	19.52	202	21265	0.372	ng/ul	99
21) Benzo(a)anthracene	21.26	228	19218	0.393	ng/ul	99
22) Chrysene	21.31	228	21448	0.397	ng/ul	99
24) Benzo(b)fluoranthene	22.83	252	19189	0.399	ng/ul	100
25) Benzo(k)fluoranthene	22.88	252	19977	0.411	ng/ul	100
26) Benzo(a)pyrene	23.40	252	15464	0.397	ng/ul	99
27) Indeno(1,2,3-cd)pyrene	25.74	276	21053	0.407	ng/ul#	99
28) Dibenzo(a,h)anthracene	25.76	278	16900	0.405	ng/ul	98
29) Benzo(g,h,i)perylene	26.43	276	18767	0.407	ng/ul	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN031221\
Data File : BN013930.D
Acq On : 12 Mar 2021 13:41
Operator : CG/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SICV010

Quant Time: Mar 12 14:12:19 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SFAM-EPA-SIM-BN031221.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 12 13:33:49 2021
Response via : Initial Calibration

