

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN031221\
 Data File : BN013931.D
 Acq On : 12 Mar 2021 15:00
 Operator : CG/JU
 Sample : M1496-04
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 X3166

Quant Time: Mar 12 15:31:02 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SFAM-EPA-SIM-BN031221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 12 13:33:49 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.71	152	4521	0.40	ng/ul	0.00
4) Naphthalene-d8	10.49	136	15143	0.40	ng/ul	0.00
9) Acenaphthene-d10	14.35	164	7404	0.40	ng/ul	0.00
13) Phenanthrene-d10	17.10	188	15595	0.40	ng/ul	0.00
17) Chrysene-d12	21.28	240	13491	0.40	ng/ul	0.00
23) Perylene-d12	23.51	264	11416	0.40	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.22	96	3032	0.63	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.09	152	7772	0.36	ng/ul	0.00
18) Fluoranthene-d10	19.13	212	14247	0.35	ng/ul	0.00
Target Compounds				Ovalue		
5) Naphthalene	10.54	128	28254	0.640	ng/ul	100
7) 2-Methylnaphthalene	12.17	142	12283	0.428	ng/ul	100
12) Fluorene	15.40	166	18379	0.594	ng/ul	97
14) Pentachlorophenol	16.75	266	7292	2.734	ng/ul	98
16) Anthracene	17.23	178	36116	0.840	ng/ul	99
19) Fluoranthene	19.16	202	37373	0.689	ng/ul	97
20) Pyrene	19.52	202	42748	0.761	ng/ul	98
22) Chrysene	21.31	228	30141	0.568	ng/ul	99
24) Benzo(b)fluoranthene	22.83	252	25178	0.471	ng/ul	99
26) Benzo(a)pyrene	23.41	252	32742	0.756	ng/ul	99
28) Dibenzo(a,h)anthracene	25.76	278	66144	1.426	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN031221\
Data File : BN013931.D
Acq On : 12 Mar 2021 15:00
Operator : CG/JU
Sample : M1496-04
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
X3166

Quant Time: Mar 12 15:31:02 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SFAM-EPA-SIM-BN031221.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 12 13:33:49 2021
Response via : Initial Calibration

