

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN031221\
 Data File : BN013932.D
 Acq On : 12 Mar 2021 16:33
 Operator : CG/JU
 Sample : PB134983BS
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS983

Quant Time: Mar 12 17:10:49 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SFAM-EPA-SIM-BN031221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 12 13:33:49 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.71	152	4435	0.40	ng/ul	0.00
4) Naphthalene-d8	10.49	136	14895	0.40	ng/ul	0.00
9) Acenaphthene-d10	14.35	164	7416	0.40	ng/ul	0.00
13) Phenanthrene-d10	17.10	188	15856	0.40	ng/ul	0.00
17) Chrysene-d12	21.27	240	14526	0.40	ng/ul	0.00
23) Perylene-d12	23.50	264	11412	0.40	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.22	96	3537	0.74	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.09	152	8484	0.39	ng/ul	0.00
18) Fluoranthene-d10	19.13	212	16554	0.38	ng/ul	0.00
Target Compounds				Ovalue		
2) 1,4-Dioxane	3.25	88	9311	1.872	ng/ul#	85
5) Naphthalene	10.54	128	16537	0.381	ng/ul	100
7) 2-Methylnaphthalene	12.16	142	10490	0.371	ng/ul	100
8) 1-Methylnaphthalene	12.39	142	10109	0.370	ng/ul	100
10) Acenaphthylene	14.07	152	11359	0.340	ng/ul	100
11) Acenaphthene	14.42	153	10426	0.375	ng/ul	99
12) Fluorene	15.40	166	11510	0.371	ng/ul	99
14) Pentachlorophenol	16.75	266	1509	0.557	ng/ul	98
15) Phenanthrene	17.14	178	19497	0.368	ng/ul	99
16) Anthracene	17.23	178	15720	0.360	ng/ul	100
19) Fluoranthene	19.16	202	21221	0.363	ng/ul	96
20) Pyrene	19.52	202	21586	0.357	ng/ul	98
21) Benzo(a)anthracene	21.26	228	20271	0.392	ng/ul	99
22) Chrysene	21.31	228	22933	0.401	ng/ul	99
24) Benzo(b)fluoranthene	22.83	252	22161	0.415	ng/ul	99
25) Benzo(k)fluoranthene	22.88	252	22133	0.410	ng/ul	100
26) Benzo(a)pyrene	23.41	252	17973	0.415	ng/ul	97
27) Indeno(1,2,3-cd)pyrene	25.75	276	24483	0.426	ng/ul#	99
28) Dibenzo(a,h)anthracene	25.76	278	19721	0.425	ng/ul	99
29) Benzo(a,h,i)perylene	26.43	276	21278	0.416	ng/ul	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN031221\
Data File : BN013932.D
Acq On : 12 Mar 2021 16:33
Operator : CG/JU
Sample : PB134983BS
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SLCS983

Quant Time: Mar 12 17:10:49 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SFAM-EPA-SIM-BN031221.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 12 13:33:49 2021
Response via : Initial Calibration

