

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN031221\  
 Data File : BN013933.D  
 Acq On : 12 Mar 2021 17:17  
 Operator : CG/JU  
 Sample : PB134983BL  
 Misc :  
 ALS Vial : 12 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 ClientSampleId :  
 SBLK983

Quant Time: Mar 12 17:47:08 2021  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SFAM-EPA-SIM-BN031221.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Fri Mar 12 13:33:49 2021  
 Response via : Initial Calibration

| Internal Standards        | R.T.  | QIon | Response | Conc | Units | Dev(Min) |
|---------------------------|-------|------|----------|------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4 | 7.71  | 152  | 3688     | 0.40 | ng/ul | 0.00     |
| 4) Naphthalene-d8         | 10.49 | 136  | 12132    | 0.40 | ng/ul | 0.00     |
| 9) Acenaphthene-d10       | 14.35 | 164  | 5959     | 0.40 | ng/ul | 0.00     |
| 13) Phenanthrene-d10      | 17.10 | 188  | 12728    | 0.40 | ng/ul | 0.00     |
| 17) Chrysene-d12          | 21.27 | 240  | 11047    | 0.40 | ng/ul | 0.00     |
| 23) Perylene-d12          | 23.50 | 264  | 9003     | 0.40 | ng/ul | 0.00     |

| System Monitoring Compounds | R.T.  | QIon | Response | Conc | Units | Dev(Min) |
|-----------------------------|-------|------|----------|------|-------|----------|
| 3) 1,4-Dioxane-d8           | 3.22  | 96   | 2939     | 0.74 | ng/ul | 0.00     |
| 6) 2-Methylnaphthalene-d10  | 12.09 | 152  | 6996     | 0.40 | ng/ul | 0.00     |
| 18) Fluoranthene-d10        | 19.13 | 212  | 13217    | 0.40 | ng/ul | 0.00     |

| Target Compounds | Ovalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN031221\  
Data File : BN013933.D  
Acq On : 12 Mar 2021 17:17  
Operator : CG/JU  
Sample : PB134983BL  
Misc :  
ALS Vial : 12 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :  
SBLK983

Quant Time: Mar 12 17:47:08 2021  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SFAM-EPA-SIM-BN031221.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Fri Mar 12 13:33:49 2021  
Response via : Initial Calibration

