

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN031221\
 Data File : BN013934.D
 Acq On : 12 Mar 2021 17:54
 Operator : CG/JU
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD0.4011

Quant Time: Mar 12 18:23:52 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SFAM-EPA-SIM-BN031221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 12 13:33:49 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.71	152	4215	0.40	ng/ul	0.00
4) Naphthalene-d8	10.49	136	14068	0.40	ng/ul	0.00
9) Acenaphthene-d10	14.35	164	6931	0.40	ng/ul	0.00
13) Phenanthrene-d10	17.10	188	14789	0.40	ng/ul	0.00
17) Chrysene-d12	21.27	240	13078	0.40	ng/ul	0.00
23) Perylene-d12	23.50	264	10381	0.40	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.22	96	1788	0.40	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.09	152	8101	0.40	ng/ul	0.00
18) Fluoranthene-d10	19.13	212	15333	0.39	ng/ul	0.00
Target Compounds				Ovalue		
2) 1,4-Dioxane	3.25	88	1856	0.393	ng/ul	95
5) Naphthalene	10.54	128	16396	0.400	ng/ul	100
7) 2-Methylnaphthalene	12.16	142	10627	0.398	ng/ul	99
8) 1-Methylnaphthalene	12.39	142	10242	0.397	ng/ul	100
10) Acenaphthylene	14.07	152	11939	0.382	ng/ul	100
11) Acenaphthene	14.42	153	10354	0.399	ng/ul	98
12) Fluorene	15.40	166	11479	0.396	ng/ul	99
14) Pentachlorophenol	16.75	266	782	0.309	ng/ul	100
15) Phenanthrene	17.14	178	19272	0.390	ng/ul	100
16) Anthracene	17.23	178	15602	0.383	ng/ul	100
19) Fluoranthene	19.15	202	20264	0.386	ng/ul	99
20) Pyrene	19.52	202	20460	0.375	ng/ul	100
21) Benzo(a)anthracene	21.26	228	17884	0.384	ng/ul	99
22) Chrysene	21.31	228	20518	0.399	ng/ul	99
24) Benzo(b)fluoranthene	22.83	252	19191	0.395	ng/ul	98
25) Benzo(k)fluoranthene	22.88	252	20020	0.408	ng/ul	99
26) Benzo(a)pyrene	23.41	252	16019	0.407	ng/ul#	95
27) Indeno(1,2,3-cd)pyrene	25.74	276	21692	0.415	ng/ul#	99
28) Dibenzo(a,h)anthracene	25.75	278	17335	0.411	ng/ul	100
29) Benzo(g,h,i)perylene	26.43	276	18843	0.405	ng/ul	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN031221\
Data File : BN013934.D
Acq On : 12 Mar 2021 17:54
Operator : CG/JU
Sample : SSTDCCC0.4EC
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD0.4011

Quant Time: Mar 12 18:23:52 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SFAM-EPA-SIM-BN031221.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 12 13:33:49 2021
Response via : Initial Calibration

