

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031225\
 Data File : BN036598.D
 Acq On : 12 Mar 2025 13:29
 Operator : RC/JU
 Sample : Q1531-17DL 5X
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE125D2-20250307DL

Quant Time: Mar 12 14:14:23 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 10 16:06:28 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.724	152	2304	0.400	ng	0.00
7) Naphthalene-d8	10.519	136	5150	0.400	ng	0.01
13) Acenaphthene-d10	14.366	164	3131	0.400	ng	0.00
19) Phenanthrene-d10	17.111	188	5911	0.400	ng	# 0.00
29) Chrysene-d12	21.295	240	3996	0.400	ng	0.00
35) Perylene-d12	23.551	264	3315	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.312	112	163	0.030	ng	0.00
5) Phenol-d6	6.901	99	418	0.063	ng	0.00
8) Nitrobenzene-d5	8.886	82	352	0.063	ng	0.01
11) 2-Methylnaphthalene-d10	12.126	152	471	0.061	ng	0.02
14) 2,4,6-Tribromophenol	15.870	330	112	0.079	ng	0.01
15) 2-Fluorobiphenyl	13.003	172	930	0.051	ng	0.02
27) Fluoranthene-d10	19.146	212	1393	0.092	ng	0.00
31) Terphenyl-d14	19.745	244	934	0.098	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.239	88	4295	1.680	ng	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031225\
Data File : BN036598.D
Acq On : 12 Mar 2025 13:29
Operator : RC/JU
Sample : Q1531-17DL 5X
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE125D2-20250307DL

Quant Time: Mar 12 14:14:23 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 16:06:28 2025
Response via : Initial Calibration

