


Data Path : Z:\HPCHEM1\BNA_N\Data\BN031318\
 Data File : BN000337.D
 Acq On : 13 Mar 2018 08:20
 Operator : JU/SJ
 Sample : DFTPP77
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP77

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN030918MA.M
 Title : SVOA CALIBRATION
 Last Update : Wed Mar 14 08:11:44 2018



AutoFind: Scans 2581, 2582, 2583; Background Corrected with Scan 2576

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	34.1	140554	PASS
68	69	0.00	2	1.8	3035	PASS
69	198	0.00	100	41.4	170553	PASS
70	69	0.00	2	0.6	1037	PASS
127	198	10	80	51.6	212544	PASS
197	198	0.00	2	0.4	1699	PASS
198	198	100	100	100.0	411733	PASS
199	198	5	9	6.7	27520	PASS
275	198	10	60	22.9	94370	PASS
365	198	1	100	2.7	11274	PASS
441	443	0.01	100	77.5	42373	PASS
442	198	50	100	68.3	281152	PASS
443	442	15	24	19.5	54704	PASS

m/z	Abundance
37.00	522.0
38.00	1454.0
39.10	8741.0
40.00	970.0
41.00	733.0
43.00	431.0
44.00	932.0
45.00	314.0
49.00	953.0
50.10	32488.0
51.10	126240.0
52.10	6724.0
53.10	468.0
55.00	1155.0
56.00	3836.0
57.00	9078.0
58.00	429.0
61.00	1601.0
62.00	1877.0
63.00	5985.0
64.10	1018.0
65.00	2862.0
67.10	650.0
68.00	2983.0
69.00	152640.0
70.00	986.0
73.10	1606.0
74.00	16031.0
75.00	24184.0
76.10	9167.0
77.10	181760.0
78.10	13250.0
79.00	11771.0
80.00	8821.0
81.00	12461.0
82.00	3576.0
83.00	3245.0
84.00	554.0
85.00	3016.0
86.00	3445.0
87.00	1803.0
88.00	731.0
89.10	335.0
91.00	2744.0
92.00	3135.0
93.00	20544.0
94.10	1445.0
95.00	745.0
96.10	1159.0
97.00	691.0
98.00	15372.0
99.00	13189.0
100.00	1144.0
101.00	6703.0
101.90	458.0
103.00	2578.0
104.00	4086.0
105.00	4190.0
106.10	1441.0
107.00	52552.0
108.00	8006.0
109.00	1595.0
110.00	98344.0
111.00	15434.0
112.00	1718.0
113.00	506.0
114.90	341.0
116.10	2872.0
117.00	43096.0
118.00	3249.0
119.00	601.0
120.00	557.0
121.10	575.0
122.00	3529.0
123.00	5299.0
124.00	2431.0
125.00	2252.0
126.20	924.0
127.00	175424.0
128.00	13899.0
129.00	65184.0
130.00	6007.0
131.00	1262.0
132.10	712.0
133.10	417.0
134.00	1889.0
135.00	5623.0
136.00	2364.0
137.00	3030.0
138.00	681.0
139.00	497.0

Instrument :
BNA_N
ClientSampleId :
DFTPP77

139.90	765.0
141.00	7707.0
142.00	2859.0
143.00	1993.0
144.00	514.0
145.10	587.0
146.00	1549.0
147.00	4596.0
148.00	10881.0
149.00	2246.0
150.00	839.0
151.10	940.0
153.00	2647.0
154.00	2127.0
155.00	4866.0
156.10	6893.0
157.10	1603.0
158.00	1685.0
159.10	1244.0
160.00	2509.0
161.10	3637.0
162.10	1085.0
163.10	551.0
164.00	335.0
165.00	3316.0
166.00	2768.0
167.00	16520.0
168.00	10341.0
169.00	1662.0
170.00	663.0
171.00	715.0
172.10	1434.0
173.00	1877.0
174.00	3250.0
175.00	6371.0
176.10	1842.0
177.00	3034.0
178.00	1102.0
179.00	12370.0
180.10	8297.0
181.00	3716.0
182.10	694.0
184.00	906.0
185.10	5109.0
186.10	42464.0
187.10	13157.0
188.00	1133.0
189.00	2988.0
190.00	831.0
191.10	1694.0
192.00	3855.0
193.10	4531.0
194.10	816.0
195.10	645.0
196.00	9233.0
197.10	2043.0
198.00	309568.0
199.00	21336.0
200.00	1506.0
201.60	1091.0
203.00	2135.0
204.00	10642.0
205.10	17840.0
206.10	74696.0
207.10	13751.0
208.00	3478.0
209.10	1580.0
210.00	1021.0
211.10	3108.0
211.90	543.0
215.00	866.0
216.00	1834.0
217.00	19832.0
218.00	2783.0
221.10	11733.0
222.00	2312.0
223.00	4357.0
224.10	41272.0
225.10	10378.0
226.10	1124.0
227.00	17176.0
228.00	2644.0
229.00	3797.0
230.10	622.0
231.10	1705.0
233.00	429.0
234.00	1153.0
235.00	1260.0
236.00	787.0
237.00	1255.0
239.00	733.0
240.00	413.0
241.00	1061.0
242.10	2139.0
243.10	2369.0

Instrument :
BNA_N
ClientSampleId :
DFTPP77

244.10	30800.0
245.00	4227.0
246.00	5900.0
247.10	1169.0
249.00	944.0
251.10	371.0
253.20	635.0
254.10	1506.0
255.00	150464.0
256.00	22328.0
257.10	1943.0
258.10	8559.0
259.00	1534.0
260.00	340.0
265.00	3534.0
266.00	673.0
271.00	312.0
272.00	470.0
273.00	4492.0
274.10	11604.0
275.10	65224.0
276.00	9115.0
277.00	5656.0
278.10	886.0
281.10	400.0
283.00	665.0
283.90	332.0
285.00	1103.0
293.10	1132.0
295.00	418.0
296.10	17680.0
297.00	2449.0
302.00	423.0
303.10	1727.0
304.10	417.0
308.00	329.0
314.10	716.0
315.00	1831.0
316.00	1040.0
321.00	511.0
322.10	324.0
323.10	4964.0
324.00	911.0
327.00	965.0
328.00	662.0
332.00	416.0
333.00	389.0
334.00	3077.0
335.00	839.0
341.00	887.0
345.90	1316.0
352.00	1462.0
353.10	1062.0
354.00	1785.0
355.10	485.0
365.00	7314.0
366.00	1302.0
371.00	450.0
372.10	2854.0
373.00	667.0
383.00	608.0
390.00	390.0
402.00	1012.0
403.10	1285.0
404.00	612.0
421.10	1115.0
422.00	1330.0
423.10	9049.0
424.10	2041.0
441.10	24496.0
442.10	150848.0
443.10	29184.0
444.10	2832.0

Instrument :
BNA_N
ClientSampleId :
DFTPP77

m/z	Abundance
37.10	826.0
38.00	1993.0
39.10	11408.0
40.00	1083.0
41.00	759.0
43.00	506.0
44.00	762.0
45.00	421.0
49.10	1368.0
50.10	41208.0
51.10	162560.0
52.10	8733.0
53.00	366.0
55.00	1226.0
56.00	4745.0
57.00	12002.0
57.90	478.0
61.10	2283.0
62.00	2849.0
63.10	7219.0
64.00	1008.0
65.00	3886.0
66.10	575.0
67.00	697.0
68.10	3437.0
69.00	197184.0
70.10	1109.0
73.00	2139.0
74.00	20728.0
75.00	32088.0
76.10	11941.0
77.10	245888.0
78.10	17208.0
79.00	16616.0
80.00	12168.0
81.00	16072.0
82.00	4042.0
83.00	3772.0
84.00	554.0
85.00	4243.0
86.00	4881.0
87.00	2659.0
87.90	916.0
89.00	463.0
91.00	3812.0
92.00	4328.0
93.00	26008.0
94.00	1971.0
95.00	775.0
96.10	1909.0
97.10	737.0
98.00	20536.0
99.00	18176.0
100.10	1500.0
101.00	9307.0
101.90	553.0
103.00	3261.0
104.00	5819.0
105.00	5186.0
106.00	2044.0
107.00	72456.0
108.00	10878.0
109.10	2345.0
110.00	135616.0
111.00	20808.0
112.00	2502.0
113.00	762.0
115.00	436.0
116.00	3861.0
117.00	61928.0
118.00	4614.0
119.00	780.0
120.10	988.0
120.90	441.0
122.00	4777.0
123.00	8056.0
124.00	3535.0
125.00	2638.0
126.20	1599.0
127.00	248128.0
128.00	19432.0
129.00	94520.0
130.00	7870.0
131.00	1779.0
132.00	799.0
132.90	738.0
134.00	2815.0
135.00	7580.0
136.00	3001.0
137.00	4275.0
138.00	914.0

Instrument :
BNA_N
ClientSampleId :
DFTPP77

139.00	691.0
140.00	1259.0
141.00	12211.0
142.10	4266.0
143.00	2928.0
144.00	770.0
145.00	644.0
146.10	2365.0
147.00	6578.0
148.00	14700.0
149.00	3368.0
150.00	954.0
151.10	1655.0
151.70	1129.0
153.00	3932.0
154.00	2870.0
155.00	6708.0
156.00	9715.0
157.10	2113.0
158.00	2126.0
159.00	1582.0
160.00	4117.0
161.00	5609.0
162.10	1453.0
163.10	646.0
164.00	515.0
165.00	4420.0
166.00	3461.0
167.00	26712.0
168.00	14649.0
169.00	2071.0
170.00	941.0
171.00	1084.0
172.00	2193.0
173.00	2936.0
174.00	4955.0
175.10	9262.0
176.10	2681.0
177.00	4094.0
178.00	1569.0
179.00	18224.0
180.00	12541.0
181.00	5615.0
182.10	1032.0
182.90	636.0
184.00	1514.0
185.00	8166.0
186.10	63976.0
187.10	18176.0
188.00	1951.0
189.00	4579.0
190.00	689.0
191.00	2207.0
192.00	5939.0
193.00	6642.0
194.00	1165.0
195.20	920.0
196.10	13502.0
198.00	488896.0
199.00	31080.0
200.10	2668.0
201.40	2013.0
203.00	3333.0
204.00	17112.0
205.10	28128.0
206.10	119800.0
207.10	17712.0
208.10	4669.0
209.10	1639.0
210.00	1992.0
211.00	3991.0
213.00	456.0
215.00	1610.0
216.00	2612.0
217.00	30888.0
218.00	4178.0
219.20	388.0
221.10	18696.0
221.80	4463.0
223.00	7157.0
224.10	64936.0
225.10	17256.0
226.10	2140.0
227.00	26904.0
228.10	3775.0
229.00	5868.0
230.00	1014.0
231.10	2208.0
231.90	481.0
233.00	694.0
234.10	1856.0
235.00	1875.0
236.00	1405.0
237.00	2227.0
238.00	374.0

Instrument :
BNA_N
ClientSampleId :
DFTPP77

239.00	1192.0
240.10	720.0
241.00	1532.0
242.10	3095.0
243.10	3830.0
244.10	50200.0
245.10	7084.0
246.00	9441.0
247.00	2043.0
248.10	498.0
249.00	1747.0
250.00	319.0
251.00	430.0
251.90	349.0
253.00	1083.0
254.20	2659.0
255.10	250368.0
256.00	37720.0
257.00	2753.0
258.00	14554.0
259.00	2320.0
260.10	518.0
261.00	431.0
265.10	5434.0
266.00	814.0
271.00	544.0
272.00	615.0
273.10	7065.0
274.10	20120.0
275.10	109520.0
276.10	14935.0
277.00	8761.0
278.10	1551.0
279.00	321.0
281.20	315.0
282.10	388.0
283.10	1177.0
284.10	862.0
285.00	1625.0
286.10	366.0
289.10	330.0
292.10	433.0
293.10	2082.0
294.00	582.0
295.00	705.0
296.00	28952.0
297.10	4212.0
298.20	334.0
301.10	414.0
301.90	397.0
303.10	3374.0
304.00	979.0
308.10	373.0
309.00	374.0
310.10	413.0
314.00	1238.0
315.00	2808.0
316.10	2079.0
317.00	413.0
321.10	759.0
322.00	703.0
323.10	9550.0
324.10	2215.0
325.00	317.0
327.00	1923.0
328.00	957.0
332.00	695.0
333.00	1037.0
334.10	6357.0
335.10	1812.0
341.00	1231.0
346.00	2479.0
347.00	306.0
352.10	2860.0
353.10	2047.0
354.10	3383.0
355.10	465.0
365.00	13103.0
366.00	2329.0
371.10	804.0
372.10	5259.0
373.10	1286.0
383.00	1318.0
384.00	322.0
390.00	794.0
391.00	518.0
392.00	309.0
400.90	328.0
402.00	1917.0
403.10	2726.0
403.90	1055.0
421.00	2494.0
422.10	2464.0
423.10	15918.0
424.10	3581.0

Instrument :
BNA_N
ClientSampleId :
DFTPP77

425.10	397.0
441.10	47800.0
442.10	319872.0
443.10	64416.0
444.10	5372.0
445.00	335.0

Instrument :
BNA_N
ClientSampleId :
DFTPP77

m/z	Abundance
37.00	679.0
38.10	1600.0
39.10	9532.0
40.00	1138.0
41.00	712.0
43.00	379.0
44.00	966.0
49.00	1144.0
50.10	34672.0
51.10	132864.0
52.10	6980.0
53.00	374.0
55.10	1054.0
56.00	3880.0
57.00	9023.0
58.00	430.0
61.00	1886.0
62.00	2141.0
63.10	6106.0
64.00	942.0
65.00	3481.0
66.00	314.0
67.00	433.0
68.10	2687.0
69.00	163264.0
70.00	1017.0
73.00	1923.0
74.10	17408.0
75.00	27976.0
76.10	9658.0
77.10	201600.0
78.10	13559.0
79.00	12777.0
80.00	9714.0
81.00	13683.0
82.10	4048.0
83.00	3688.0
83.90	554.0
85.00	3481.0
86.00	4160.0
87.00	2176.0
88.00	740.0
89.00	417.0
91.00	3491.0
92.00	3603.0
93.00	22872.0
94.00	1695.0
95.00	614.0
96.00	1423.0
97.10	664.0
98.00	18400.0
99.00	15003.0
100.00	1320.0
101.00	8220.0
101.90	464.0
103.00	2813.0
104.00	5089.0
105.00	4504.0
106.00	1718.0
107.00	62192.0
108.00	8997.0
109.10	1911.0
110.00	114136.0
111.00	17008.0
112.00	2059.0
113.00	755.0
114.90	384.0
116.00	3028.0
117.00	51000.0
118.00	4015.0
119.00	801.0
120.10	767.0
121.00	386.0
122.00	4078.0
123.00	6303.0
124.00	2624.0
125.00	2372.0
126.20	1237.0
127.00	214080.0
128.00	15693.0
129.00	78328.0
130.00	6659.0
131.00	1716.0
132.10	859.0
133.00	731.0
134.00	2444.0
135.00	6854.0
136.00	2515.0
137.00	3575.0
137.90	832.0
139.00	455.0

Instrument :
BNA_N
ClientSampleId :
DFTPP77

140.00	1003.0
141.00	10148.0
142.00	3375.0
143.10	2324.0
144.00	872.0
145.00	581.0
146.00	2030.0
147.00	5667.0
148.00	13075.0
149.00	2903.0
150.00	787.0
151.00	1326.0
151.90	774.0
153.00	3637.0
154.10	2485.0
155.00	5828.0
156.10	8768.0
157.10	1741.0
158.00	1836.0
159.00	1472.0
160.00	3223.0
161.00	4699.0
162.00	1601.0
163.00	477.0
164.00	669.0
165.00	4343.0
166.10	3298.0
167.00	22376.0
168.00	12252.0
169.00	1814.0
170.00	676.0
171.00	1170.0
172.00	2344.0
173.00	2484.0
174.00	4605.0
175.10	8191.0
176.00	2344.0
177.00	3633.0
178.00	1576.0
179.00	16432.0
180.10	11600.0
181.00	5252.0
182.10	793.0
183.00	511.0
184.00	1298.0
185.10	7382.0
186.00	59600.0
187.00	16656.0
188.10	1582.0
189.00	3962.0
190.00	707.0
191.00	2036.0
192.00	5735.0
193.00	5481.0
194.00	1213.0
195.00	988.0
196.10	12352.0
197.10	3056.0
198.00	436736.0
199.00	30144.0
200.00	2681.0
201.50	1924.0
202.20	596.0
203.00	3217.0
204.00	15259.0
205.10	26512.0
206.10	111160.0
207.10	16584.0
208.00	4196.0
209.00	1556.0
211.10	3981.0
212.00	783.0
213.00	349.0
215.00	1313.0
216.00	2345.0
217.00	28496.0
218.00	3680.0
219.00	419.0
221.10	17280.0
221.90	3773.0
223.00	6580.0
224.10	58640.0
225.10	16145.0
226.10	1827.0
227.00	24560.0
228.00	3770.0
229.00	4947.0
230.00	700.0
231.10	2448.0
232.00	407.0
232.90	473.0
234.00	1573.0
235.00	1644.0
236.00	1202.0
237.10	1828.0

Instrument :
BNA_N
ClientSampleId :
DFTPP77

238.10	477.0
239.10	1087.0
240.00	863.0
241.00	1202.0
242.10	3128.0
243.00	3444.0
244.10	48592.0
245.00	6598.0
246.00	9302.0
247.00	2038.0
248.10	503.0
249.00	1848.0
250.00	339.0
251.00	481.0
252.00	420.0
253.00	1069.0
255.00	240256.0
256.00	34448.0
257.00	2768.0
258.00	13655.0
259.10	2133.0
260.00	343.0
261.00	370.0
264.10	400.0
265.00	5841.0
266.00	826.0
270.00	395.0
271.10	537.0
272.10	744.0
273.10	6804.0
274.10	19440.0
275.10	108368.0
276.10	14329.0
277.10	9164.0
278.00	1831.0
281.10	430.0
283.10	1043.0
284.00	848.0
285.00	1875.0
286.00	329.0
289.10	412.0
290.00	318.0
291.10	311.0
292.10	375.0
293.00	2120.0
294.00	563.0
295.10	730.0
296.00	29632.0
297.00	4283.0
301.10	443.0
302.20	505.0
303.10	3274.0
304.10	1113.0
308.00	514.0
309.00	369.0
309.90	319.0
310.10	310.0
314.10	1488.0
315.00	3427.0
316.10	1674.0
317.00	302.0
321.00	937.0
322.00	663.0
323.10	10144.0
324.10	1840.0
327.00	2131.0
328.00	1023.0
332.00	741.0
333.00	856.0
334.10	6115.0
335.00	1882.0
341.00	1322.0
341.90	382.0
346.00	2123.0
347.00	535.0
352.00	3264.0
353.10	1700.0
354.10	3356.0
355.10	666.0
365.00	13406.0
366.00	2603.0
370.10	338.0
371.00	901.0
372.10	5081.0
373.00	1534.0
383.10	1591.0
384.00	469.0
390.00	677.0
391.00	625.0
392.00	531.0
401.00	450.0
402.00	2207.0
403.10	3128.0
404.10	1324.0
421.10	2984.0

Instrument :
BNA_N
ClientSampleId :
DFTPP77

422.00	2439.0
423.10	20048.0
424.10	3981.0
425.00	544.0
439.70	377.0
441.10	54824.0
442.10	372736.0
443.10	70512.0
444.10	6568.0
445.10	429.0

Instrument :
BNA_N
ClientSampleId :
DFTPP77

m/z	Abundance
40.00	701.0
41.10	701.0
43.00	591.0
44.00	969.0
55.00	553.0
67.10	311.0
69.10	476.0
73.00	697.0
77.00	346.0
79.00	308.0
81.00	479.0
91.00	316.0
93.00	322.0
95.10	434.0
96.00	464.0
133.00	311.0
190.90	452.0
207.10	2567.0
208.10	572.0
209.00	415.0
281.10	410.0

Instrument :
BNA_N
ClientSampleId :
DFTPP77