

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN031320\
 Data File : BN010213.D
 Acq On : 13 Mar 2020 22:46
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN031320

Quant Time: Mar 16 02:19:20 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN031320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 16 02:17:23 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	116	0.00
2	1,4-Dioxane	0.390	0.409	-4.9	113	0.00
3	n-Nitrosodimethylamine	0.346	0.302	12.7	113	0.00
4 S	2-Fluorophenol	0.939	0.944	-0.5	110	0.00
5 S	Phenol-d6	1.173	1.151	1.9	106	0.00
6	bis(2-Chloroethyl)ether	0.980	1.036	-5.7	125	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	116	0.00
8 S	Nitrobenzene-d5	0.240	0.229	4.6	111	0.00
9	Naphthalene	1.002	0.993	0.9	114	0.00
10	Hexachlorobutadiene	0.231	0.232	-0.4	117	-0.01
11 SURR	2-Methylnaphthalene-d10	0.661	0.648	2.0	116	0.00
12	2-Methylnaphthalene	0.720	0.704	2.2	113	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	113	0.00
14 S	2,4,6-Tribromophenol	0.143	0.140	2.1	113	0.00
15 S	2-Fluorobiphenyl	1.424	1.438	-1.0	112	0.00
16	Acenaphthylene	1.847	1.851	-0.2	114	0.00
17	Acenaphthene	1.120	1.100	1.8	112	0.00
18	Fluorene	1.527	1.534	-0.5	114	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	115	0.00
20	4-Bromophenyl-phenylether	0.230	0.226	1.7	114	0.00
21	Hexachlorobenzene	0.224	0.229	-2.2	117	0.00
22	Pentachlorophenol	0.069	0.065	5.8	119	0.00
23	Phenanthrene	1.061	1.060	0.1	113	0.00
24	Anthracene	1.058	1.045	1.2	111	0.00
25 SURR	Fluoranthene-d10	1.235	1.228	0.6	116	0.00
26	Fluoranthene	1.344	1.347	-0.2	113	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	113	0.00
28	Pvrene	1.336	1.345	-0.7	114	0.00
29 S	Terphenyl-d14	0.912	0.911	0.1	111	0.00
30	Benzo(a)anthracene	1.360	1.368	-0.6	115	-0.01
31	Chrysene	1.339	1.333	0.4	113	0.00
32	Bis(2-ethylhexyl)phthalate	0.579	0.516	10.9	108	0.00
33	Indeno(1,2,3-cd)pyrene	1.588	1.648	-3.8	116	0.00
34 I	Perylene-d12	1.000	1.000	0.0	116	0.00
35	Benzo(b)fluoranthene	1.230	1.193	3.0	113	0.00
36	Benzo(k)fluoranthene	1.243	1.239	0.3	115	0.00
37 C	Benzo(a)pyrene	1.161	1.126	3.0	113	0.00
38	Dibenzo(a,h)anthracene	1.164	1.175	-0.9	117	0.00
39	Benzo(a,h,i)perylene	1.163	1.165	-0.2	116	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN031320\
Data File : BN010213.D
Acq On : 13 Mar 2020 22:46
Operator : CG/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
ICVBN031320

Quant Time: Mar 16 02:19:20 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN031320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 16 02:17:23 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			