

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031323\
 Data File : BN024264.D
 Acq On : 14 Mar 2023 04:33
 Operator : CG/JU
 Sample : 01849-04
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 CB-10-22.0-030823

Manual Integrations
 APPROVED

Reviewed By : Christian Giraldo 03/14/2023
 Supervised By : Jagrut Upadhyay 03/14/2023

Quant Time: Mar 14 05:02:56 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 14 03:11:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.047	152	7116	0.400	ng	0.00
7) Naphthalene-d8	10.866	136	20564	0.400	ng	0.00
13) Acenaphthene-d10	14.675	164	10234	0.400	ng	0.00
19) Phenanthrene-d10	17.425	188	21172	0.400	ng	0.00
29) Chrysene-d12	21.610	240	18169	0.400	ng	# 0.00
35) Perylene-d12	24.111	264	16516	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.584	112	3991	0.253	ng	0.00
5) Phenol-d6	7.195	99	5219	0.258	ng	0.00
8) Nitrobenzene-d5	9.211	82	2417m	0.176	ng	0.00
11) 2-Methylnaphthalene-d10	12.444	152	6023	0.238	ng	0.00
14) 2,4,6-Tribromophenol	16.159	330	440	0.084	ng	0.00
15) 2-Fluorobiphenyl	13.307	172	8923	0.231	ng	0.00
27) Fluoranthene-d10	19.450	212	11332	0.252	ng	0.00
31) Terphenyl-d14	20.043	244	6468	0.211	ng	0.00
Target Compounds						
34) Bis(2-ethylhexyl)phtha...	21.511	149	3102	0.109	ng	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031323\
Data File : BN024264.D
Acq On : 14 Mar 2023 04:33
Operator : CG/JU
Sample : 01849-04
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
CB-10-22.0-030823

Manual Integrations
APPROVED

Reviewed By : Christian Giraldo 03/14/2023
Supervised By : Jagrut Upadhyay 03/14/2023

Quant Time: Mar 14 05:02:56 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031323.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 14 03:11:52 2023
Response via : Initial Calibration

