

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031323\
 Data File : BN024268.D
 Acq On : 14 Mar 2023 07:00
 Operator : CG/JU
 Sample : 01849-03
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 CB-10-17.5-030823

Quant Time: Mar 14 07:36:59 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 14 03:11:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.047	152	7827	0.400	ng	0.00
7) Naphthalene-d8	10.866	136	22896	0.400	ng	0.00
13) Acenaphthene-d10	14.675	164	11973	0.400	ng	0.00
19) Phenanthrene-d10	17.425	188	24994	0.400	ng	0.00
29) Chrysene-d12	21.610	240	20521	0.400	ng	# 0.00
35) Perylene-d12	24.132	264	18384	0.400	ng	0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.584	112	3826	0.221	ng	0.00
5) Phenol-d6	7.195	99	5637	0.253	ng	0.00
8) Nitrobenzene-d5	9.211	82	938	0.061	ng	0.00
11) 2-Methylnaphthalene-d10	12.444	152	6682	0.237	ng	0.00
14) 2,4,6-Tribromophenol	16.159	330	655	0.107	ng	0.00
15) 2-Fluorobiphenyl	13.307	172	10144	0.224	ng	0.00
27) Fluoranthene-d10	19.450	212	13181	0.249	ng	0.00
31) Terphenyl-d14	20.042	244	7717	0.223	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.432	88	1126	0.132	ng	# 60
34) Bis(2-ethylhexyl)phtha...	21.511	149	2346	0.073	ng	# 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031323\
Data File : BN024268.D
Acq On : 14 Mar 2023 07:00
Operator : CG/JU
Sample : 01849-03
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
CB-10-17.5-030823

Quant Time: Mar 14 07:36:59 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031323.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 14 03:11:52 2023
Response via : Initial Calibration

