

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031323\
 Data File : BN024260.D
 Acq On : 14 Mar 2023 01:30
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 14 03:34:33 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 14 03:11:52 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	97	0.00
2	1,4-Dioxane	0.400	0.412	-3.0	95	0.00
3	n-Nitrosodimethylamine	0.400	0.412	-3.0	98	0.00
4 S	2-Fluorophenol	0.400	0.393	1.8	95	0.00
5 S	Phenol-d6	0.400	0.385	3.8	95	0.00
6	bis(2-Chloroethyl)ether	0.400	0.403	-0.8	95	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	95	0.00
8 S	Nitrobenzene-d5	0.400	0.396	1.0	100	0.00
9	Naphthalene	0.400	0.397	0.8	95	-0.01
10	Hexachlorobutadiene	0.400	0.397	0.8	94	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.382	4.5	93	0.00
12	2-Methylnaphthalene	0.400	0.380	5.0	93	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	90	0.00
14 S	2,4,6-Tribromophenol	0.400	0.339	15.3	83	0.00
15 S	2-Fluorobiphenyl	0.400	0.400	0.0	91	0.00
16	Acenaphthylene	0.400	0.376	6.0	88	0.00
17	Acenaphthene	0.400	0.386	3.5	88	0.00
18	Fluorene	0.400	0.376	6.0	87	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	86	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.383	4.3	78	0.00
21	4-Bromophenyl-phenylether	0.400	0.372	7.0	83	-0.01
22	Hexachlorobenzene	0.400	0.390	2.5	86	0.00
23	Atrazine	0.400	0.346	13.5	84	0.00
24	Pentachlorophenol	0.400	0.378	5.5	82	0.00
25	Phenanthrene	0.400	0.393	1.8	87	0.00
26	Anthracene	0.400	0.372	7.0	85	0.00
27 SURR	Fluoranthene-d10	0.400	0.387	3.3	87	0.00
28	Fluoranthene	0.400	0.385	3.8	86	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	90	0.00
30	Pyrene	0.400	0.391	2.3	87	0.00
31 S	Terphenyl-d14	0.400	0.360	10.0	80	0.00
32	Benzo(a)anthracene	0.400	0.376	6.0	88	0.00
33	Chrysene	0.400	0.402	-0.5	90	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.358	10.5	88	0.00
35 I	Perylene-d12	0.400	0.400	0.0	98	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.405	-1.3	104	0.00
37	Benzo(b)fluoranthene	0.400	0.369	7.8	96	0.00
38	Benzo(k)fluoranthene	0.400	0.374	6.5	97	0.00
39 C	Benzo(a)pyrene	0.400	0.378	5.5	99	0.00
40	Dibenzo(a,h)anthracene	0.400	0.395	1.3	103	0.00
41	Benzo(g,h,i)perylene	0.400	0.405	-1.3	104	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4