

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031323\
 Data File : BN024273.D
 Acq On : 14 Mar 2023 10:03
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 15 01:43:06 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 14 03:11:52 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	113	0.00
2	1,4-Dioxane	0.400	0.402	-0.5	108	0.00
3	n-Nitrosodimethylamine	0.400	0.406	-1.5	112	0.00
4 S	2-Fluorophenol	0.400	0.395	1.3	111	0.00
5 S	Phenol-d6	0.400	0.386	3.5	112	0.00
6	bis(2-Chloroethyl)ether	0.400	0.398	0.5	109	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	112	0.00
8 S	Nitrobenzene-d5	0.400	0.395	1.3	117	0.00
9	Naphthalene	0.400	0.382	4.5	109	0.00
10	Hexachlorobutadiene	0.400	0.383	4.3	107	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.366	8.5	105	0.00
12	2-Methylnaphthalene	0.400	0.366	8.5	105	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	106	0.00
14 S	2,4,6-Tribromophenol	0.400	0.322	19.5	94	0.00
15 S	2-Fluorobiphenyl	0.400	0.383	4.3	103	0.00
16	Acenaphthylene	0.400	0.358	10.5	100	0.00
17	Acenaphthene	0.400	0.366	8.5	99	0.00
18	Fluorene	0.400	0.359	10.3	98	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	104	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.356	11.0	82	0.00
21	4-Bromophenyl-phenylether	0.400	0.347	13.3	93	0.00
22	Hexachlorobenzene	0.400	0.359	10.3	96	0.00
23	Atrazine	0.400	0.334	16.5	98	0.00
24	Pentachlorophenol	0.400	0.344	14.0	85	0.00
25	Phenanthrene	0.400	0.367	8.3	98	0.00
26	Anthracene	0.400	0.344	14.0	95	0.00
27 SURR	Fluoranthene-d10	0.400	0.354	11.5	96	0.00
28	Fluoranthene	0.400	0.355	11.3	96	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	103	0.00
30	Pyrene	0.400	0.380	5.0	97	0.00
31 S	Terphenyl-d14	0.400	0.365	8.8	93	0.00
32	Benzo(a)anthracene	0.400	0.358	10.5	96	0.00
33	Chrysene	0.400	0.371	7.3	95	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.351	12.3	98	0.00
35 I	Perylene-d12	0.400	0.400	0.0	105	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.340	15.0	94	0.00
37	Benzo(b)fluoranthene	0.400	0.348	13.0	97	0.00
38	Benzo(k)fluoranthene	0.400	0.362	9.5	101	0.00
39 C	Benzo(a)pyrene	0.400	0.356	11.0	100	0.00
40	Dibenzo(a,h)anthracene	0.400	0.334	16.5	93	0.00
41	Benzo(g,h,i)perylene	0.400	0.359	10.3	99	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4