

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031323\
 Data File : BN024243.D
 Acq On : 13 Mar 2023 13:11
 Operator : CG/JU
 Sample : SSTDICC0.1
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1

Quant Time: Mar 14 03:06:28 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 14 03:05:51 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.047	152	9667	0.400	ng	0.00
7) Naphthalene-d8	10.866	136	27655	0.400	ng	0.00
13) Acenaphthene-d10	14.675	164	14835	0.400	ng	0.00
19) Phenanthrene-d10	17.425	188	31371	0.400	ng	0.00
29) Chrysene-d12	21.610	240	23906	0.400	ng	0.00
35) Perylene-d12	24.123	264	19588	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.584	112	2206	0.103	ng	0.00
5) Phenol-d6	7.195	99	2732	0.099	ng	0.00
8) Nitrobenzene-d5	9.211	82	1704	0.092	ng	0.00
11) 2-Methylnaphthalene-d10	12.444	152	3297	0.097	ng	0.00
14) 2,4,6-Tribromophenol	16.159	330	697	0.092	ng	0.00
15) 2-Fluorobiphenyl	13.307	172	5541	0.099	ng	0.00
27) Fluoranthene-d10	19.450	212	6219	0.094	ng	0.00
31) Terphenyl-d14	20.042	244	4076	0.101	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.439	88	1121	0.107	ng	95
3) n-Nitrosodimethylamine	3.764	42	1514	0.098	ng	# 95
6) bis(2-Chloroethyl)ether	7.462	93	2760	0.101	ng	99
9) Naphthalene	10.909	128	6973	0.099	ng	96
10) Hexachlorobutadiene	11.208	225	1330	0.100	ng	# 100
12) 2-Methylnaphthalene	12.516	142	4569	0.096	ng	99
16) Acenaphthylene	14.397	152	6511	0.093	ng	99
17) Acenaphthene	14.739	154	4214	0.094	ng	99
18) Fluorene	15.725	166	4991	0.093	ng	99
21) 4-Bromophenyl-phenylether	16.606	248	1817	0.097	ng	# 71
22) Hexachlorobenzene	16.730	284	2275	0.099	ng	99
23) Atrazine	16.867	200	1032	0.087	ng	# 89
25) Phenanthrene	17.462	178	8830	0.095	ng	100
26) Anthracene	17.549	178	7741	0.090	ng	100
28) Fluoranthene	19.480	202	9239	0.091	ng	100
30) Pyrene	19.842	202	9449	0.098	ng	100
32) Benzo(a)anthracene	21.592	228	7431	0.093	ng	98
33) Chrysene	21.646	228	8381	0.098	ng	99
34) Bis(2-ethylhexyl)phtha...	21.511	149	3752	0.100	ng	# 98
36) Indeno(1,2,3-cd)pyrene	26.754	276	6279	0.086	ng	99
37) Benzo(b)fluoranthene	23.351	252	6976	0.092	ng	# 85
38) Benzo(k)fluoranthene	23.404	252	6846	0.087	ng	# 83
39) Benzo(a)pyrene	24.009	252	6071	0.087	ng	# 81
40) Dibenzo(a,h)anthracene	26.772	278	4709	0.083	ng	# 87
41) Benzo(g,h,i)perylene	27.564	276	5900	0.089	ng	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031323\
Data File : BN024243.D
Acq On : 13 Mar 2023 13:11
Operator : CG/JU
Sample : SSTDICC0.1
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1

Quant Time: Mar 14 03:06:28 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031323.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 14 03:05:51 2023
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 8.047 min Scan# 660
 Delta R.T. 0.000 min
 Lab File: BN024243.D
 Acq: 13 Mar 2023 13:11




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1


Tgt Ion:152 Resp: 9667
 Ion Ratio Lower Upper
 152 100
 150 155.8 124.7 187.1
 115 60.6 49.1 73.7



#2
 1,4-Dioxane
 Concen: 0.107 ng
 RT: 3.439 min Scan# 22
 Delta R.T. 0.000 min
 Lab File: BN024243.D
 Acq: 13 Mar 2023 13:11

Tgt Ion: 88 Resp: 1121
 Ion Ratio Lower Upper
 88 100
 43 41.0 29.0 43.4
 58 81.4 67.3 100.9





#3

n-Nitrosodimethylamine

Concen: 0.098 ng

RT: 3.764 min Scan# 61

Delta R.T. 0.000 min

Lab File: BN024243.D

Acq: 13 Mar 2023 13:11

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion: 42 Resp: 1514

Ion Ratio Lower Upper

42 100

74 105.8 81.6 122.4

44 14.6 7.3 10.9#



#4

2-Fluorophenol

Concen: 0.103 ng

RT: 5.584 min Scan# 319

Delta R.T. 0.000 min

Lab File: BN024243.D

Acq: 13 Mar 2023 13:11




Tgt Ion: 112 Resp: 2206


Ion Ratio Lower Upper

112 100

64 61.1 49.0 73.6

63 32.3 25.7 38.5









#5
Phenol-d6
Concen: 0.099 ng
RT: 7.195 min Scan# 54
Delta R.T. 0.000 min
Lab File: BN024243.D
Acq: 13 Mar 2023 13:11




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1

Tgt Ion: 99 Resp: 2732
Ion Ratio Lower Upper
99 100
42 24.7 20.1 30.1
71 32.5 25.8 38.8



#6
bis(2-Chloroethyl)ether
Concen: 0.101 ng
RT: 7.462 min Scan# 579
Delta R.T. 0.000 min
Lab File: BN024243.D
Acq: 13 Mar 2023 13:11

Tgt Ion: 93 Resp: 2760
Ion Ratio Lower Upper
93 100
63 81.0 65.4 98.2
95 33.4 25.8 38.8









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.866 min Scan# 936
Delta R.T. 0.000 min
Lab File: BN024243.D
Acq: 13 Mar 2023 13:11




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:	136	Resp:	27655
Ion	Ratio	Lower	Upper
136	100		
137	11.2	8.9	13.3
54	9.2	7.1	10.7
68	6.9	5.3	7.9



#8
Nitrobenzene-d5
Concen: 0.092 ng
RT: 9.211 min Scan# 781
Delta R.T. 0.000 min
Lab File: BN024243.D
Acq: 13 Mar 2023 13:11

Tgt Ion:	82	Resp:	1704
Ion	Ratio	Lower	Upper
82	100		
128	45.1	32.4	48.6
54	63.0	45.2	67.8









#9
Naphthalene
Concen: 0.099 ng
RT: 10.909 min Scan# 941
Delta R.T. -0.011 min
Lab File: BN024243.D
Acq: 13 Mar 2023 13:11




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:	128	Resp:	6973
Ion	Ratio	Lower	Upper
128	100		
129	12.7	9.0	13.6
127	14.5	10.2	15.4



#10
Hexachlorobutadiene
Concen: 0.100 ng
RT: 11.208 min Scan# 968
Delta R.T. 0.000 min
Lab File: BN024243.D
Acq: 13 Mar 2023 13:11

Tgt Ion:	225	Resp:	1330
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	63.5	51.0	76.4









#11
2-Methylnaphthalene-d10
Concen: 0.097 ng
RT: 12.444 min Scan# 11
Delta R.T. 0.000 min
Lab File: BN024243.D
Acq: 13 Mar 2023 13:11




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:152 Resp: 3297
Ion Ratio Lower Upper
152 100
151 20.8 16.8 25.2



#12
2-Methylnaphthalene
Concen: 0.096 ng
RT: 12.516 min Scan# 1263
Delta R.T. 0.000 min
Lab File: BN024243.D
Acq: 13 Mar 2023 13:11

Tgt Ion:142 Resp: 4569
Ion Ratio Lower Upper
142 100
141 89.3 71.8 107.6
115 39.3 30.0 45.0









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.675 min Scan# 1549
Delta R.T. 0.000 min
Lab File: BN024243.D
Acq: 13 Mar 2023 13:11




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:164 Resp: 14835
Ion Ratio Lower Upper
164 100
162 106.5 84.7 127.1
160 54.5 44.6 66.8



#14
2,4,6-Tribromophenol
Concen: 0.092 ng
RT: 16.159 min Scan# 1682
Delta R.T. 0.000 min
Lab File: BN024243.D
Acq: 13 Mar 2023 13:11

Tgt Ion:330 Resp: 697
Ion Ratio Lower Upper
330 100
332 96.6 78.0 117.0
141 38.7 30.8 46.2





#15

2-Fluorobiphenyl

Concen: 0.099 ng

RT: 13.307 min Scan# 1421

Delta R.T. 0.000 min

Lab File: BN024243.D

Acq: 13 Mar 2023 13:11

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:172 Resp: 5541

Ion Ratio Lower Upper

172 100

171 35.7 27.8 41.8

170 25.0 19.0 28.6



#16

Acenaphthylene

Concen: 0.093 ng

RT: 14.397 min Scan# 1523

Delta R.T. 0.000 min

Lab File: BN024243.D

Acq: 13 Mar 2023 13:11




Tgt Ion:152 Resp: 6511


Ion Ratio Lower Upper

152 100

151 19.6 15.4 23.0

153 13.0 10.4 15.6









#17
Acenaphthene
Concen: 0.094 ng
RT: 14.739 min Scan# 1555
Delta R.T. 0.000 min
Lab File: BN024243.D
Acq: 13 Mar 2023 13:11




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:154 Resp: 4214
Ion Ratio Lower Upper
154 100
153 112.8 88.4 132.6
152 60.7 48.7 73.1



#18
Fluorene
Concen: 0.093 ng
RT: 15.725 min Scan# 1647
Delta R.T. 0.000 min
Lab File: BN024243.D
Acq: 13 Mar 2023 13:11

Tgt Ion:166 Resp: 4991
Ion Ratio Lower Upper
166 100
165 99.0 80.2 120.2
167 13.6 10.8 16.2





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.425 min Scan# 1784

Delta R.T. 0.000 min

Lab File: BN024243.D

Acq: 13 Mar 2023 13:11

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:188 Resp: 31371

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 9.8 8.9 13.3



#21

4-Bromophenyl-phenylether

Concen: 0.097 ng

RT: 16.606 min Scan# 1718

Delta R.T. -0.012 min

Lab File: BN024243.D

Acq: 13 Mar 2023 13:11




Tgt Ion:248 Resp: 1817


Ion Ratio Lower Upper

248 100

250 92.6 81.9 122.9

141 84.4 28.2 42.2#









#22
Hexachlorobenzene
Concen: 0.099 ng
RT: 16.730 min Scan# 1728
Delta R.T. 0.000 min
Lab File: BN024243.D
Acq: 13 Mar 2023 13:11




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:284 Resp: 2275
Ion Ratio Lower Upper
284 100
142 37.0 30.2 45.4
249 30.5 24.5 36.7



#23
Atrazine
Concen: 0.087 ng
RT: 16.867 min Scan# 1739
Delta R.T. -0.012 min
Lab File: BN024243.D
Acq: 13 Mar 2023 13:11

Tgt Ion:200 Resp: 1032
Ion Ratio Lower Upper
200 100
173 38.0 18.9 28.3#
215 47.1 36.7 55.1









#25
Phenanthrene
Concen: 0.095 ng
RT: 17.462 min Scan# 1787
Delta R.T. 0.000 min
Lab File: BN024243.D
Acq: 13 Mar 2023 13:11




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:178 Resp: 8830
Ion Ratio Lower Upper
178 100
176 19.4 15.4 23.0
179 15.3 12.2 18.2



#26
Anthracene
Concen: 0.090 ng
RT: 17.549 min Scan# 1794
Delta R.T. 0.000 min
Lab File: BN024243.D
Acq: 13 Mar 2023 13:11

Tgt Ion:178 Resp: 7741
Ion Ratio Lower Upper
178 100
176 18.8 15.1 22.7
179 15.3 12.1 18.1









#27
Fluoranthene-d10
Concen: 0.094 ng
RT: 19.450 min Scan# 2173
Delta R.T. 0.000 min
Lab File: BN024243.D
Acq: 13 Mar 2023 13:11




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion	Ratio	Lower	Upper
212	100		
106	18.9	15.4	23.2
104	11.0	8.6	13.0



#28
Fluoranthene
Concen: 0.091 ng
RT: 19.480 min Scan# 2180
Delta R.T. 0.000 min
Lab File: BN024243.D
Acq: 13 Mar 2023 13:11

Tgt Ion	Ratio	Lower	Upper
202	100		
101	13.8	11.3	16.9
203	17.1	13.6	20.4









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.610 min Scan# 2485
Delta R.T. -0.009 min
Lab File: BN024243.D
Acq: 13 Mar 2023 13:11




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:240 Resp: 23906
Ion Ratio Lower Upper
240 100
120 16.0 12.8 19.2
236 27.7 22.5 33.7



#30
Pyrene
Concen: 0.098 ng
RT: 19.842 min Scan# 2266
Delta R.T. 0.000 min
Lab File: BN024243.D
Acq: 13 Mar 2023 13:11

Tgt Ion:202 Resp: 9449
Ion Ratio Lower Upper
202 100
200 21.2 17.0 25.4
203 18.0 14.2 21.2









#31
 Terphenyl-d14
 Concen: 0.101 ng
 RT: 20.042 min Scan# 2309
 Delta R.T. 0.000 min
 Lab File: BN024243.D
 Acq: 13 Mar 2023 13:11




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1


Tgt Ion:244 Resp: 4076
 Ion Ratio Lower Upper
 244 100
 212 9.2 7.6 11.4
 122 14.1 15.1 22.7#



#32
 Benzo(a)anthracene
 Concen: 0.093 ng
 RT: 21.592 min Scan# 2482
 Delta R.T. -0.009 min
 Lab File: BN024243.D
 Acq: 13 Mar 2023 13:11

Tgt Ion:228 Resp: 7431
 Ion Ratio Lower Upper
 228 100
 226 29.0 22.2 33.2
 229 20.4 15.9 23.9









#33
Chrysene
Concen: 0.098 ng
RT: 21.646 min Scan# 2488
Delta R.T. -0.009 min
Lab File: BN024243.D
Acq: 13 Mar 2023 13:11




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:228 Resp: 8381
Ion Ratio Lower Upper
228 100
226 31.0 24.5 36.7
229 19.5 15.5 23.3



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.100 ng
RT: 21.511 min Scan# 2473
Delta R.T. -0.009 min
Lab File: BN024243.D
Acq: 13 Mar 2023 13:11

Tgt Ion:149 Resp: 3752
Ion Ratio Lower Upper
149 100
167 25.7 21.2 31.8
279 4.2 1.8 2.8#









#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 24.123 min Scan# 3158
 Delta R.T. 0.006 min
 Lab File: BN024243.D
 Acq: 13 Mar 2023 13:11




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1


Tgt Ion:264 Resp: 19588
 Ion Ratio Lower Upper
 264 100
 260 26.2 21.4 32.2
 265 44.3 35.7 53.5



#36
 Indeno(1,2,3-cd)pyrene
 Concen: 0.086 ng
 RT: 26.754 min Scan# 4060
 Delta R.T. 0.029 min
 Lab File: BN024243.D
 Acq: 13 Mar 2023 13:11

Tgt Ion:276 Resp: 6279
 Ion Ratio Lower Upper
 276 100
 138 33.7 27.4 41.2
 277 24.5 19.8 29.6





#37

Benzo(b)fluoranthene

Concen: 0.092 ng

RT: 23.351 min Scan# 2896

Delta R.T. 0.000 min

Lab File: BN024243.D

Acq: 13 Mar 2023 13:11

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:252 Resp: 6976

Ion Ratio Lower Upper

252 100

253 29.1 18.9 28.3#

125 26.4 13.8 20.6#



#38

Benzo(k)fluoranthene

Concen: 0.087 ng

RT: 23.404 min Scan# 2914

Delta R.T. 0.003 min

Lab File: BN024243.D

Acq: 13 Mar 2023 13:11




Tgt Ion:252 Resp: 6846


Ion Ratio Lower Upper

252 100

253 30.0 19.0 28.6#

125 27.4 13.5 20.3#









#39
Benzo(a)pyrene
Concen: 0.087 ng
RT: 24.009 min Scan# 3119
Delta R.T. 0.006 min
Lab File: BN024243.D
Acq: 13 Mar 2023 13:11




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:252 Resp: 6071
Ion Ratio Lower Upper
252 100
253 31.0 19.5 29.3#
125 32.3 16.2 24.4#



#40
Dibenzo(a,h)anthracene
Concen: 0.083 ng
RT: 26.772 min Scan# 4066
Delta R.T. 0.026 min
Lab File: BN024243.D
Acq: 13 Mar 2023 13:11

Tgt Ion:278 Resp: 4709
Ion Ratio Lower Upper
278 100
139 29.8 19.4 29.0#
279 32.8 20.7 31.1#





#41
 Benzo(g,h,i)perylene
 Concen: 0.089 ng
 RT: 27.564 min Scan# 43
 Delta R.T. 0.035 min
 Lab File: BN024243.D
 Acq: 13 Mar 2023 13:11

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1

Tgt Ion:	276	Resp:	5900
Ion	Ratio	Lower	Upper
276	100		
277	27.4	19.9	29.9
138	33.9	24.5	36.7

